MIL-A-36763A
14 March 1974
SUPERSEDING
MIL-A-36763
6 May 1971

MILITARY SPECIFICATION

ANESTHESIA SET, SPINAL, DISPOSABLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

* 1.1 Scope. This specification covers a sterile, disposable spinal anesthesia set suitable for use in anesthetizing patients for surgical procedures. Potency is 24 months, which is determined by the shortest dated item (procaine hydrochloride injection) in the set.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
GG-N-196	Needles, Hypodermic, for Luer Syringes.
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-Plywood.
PPP-B-621	Box, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-C-00186	Containers, Packaging and Packing for Drugs, Chemicals and Pharmaceuticals.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water-Resistant (for Carton Sealing).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547	Liners, Case and Sheet Overwrap; Water-Vaporproof, or Waterproof, Flexible.
MIL-S-36157	Syringes, Hypodermic, Disposable.
MIL-N-36214	Needle, Hypodermic, Spinal Anesthesia.

STANDARDS

Federal

Fed. Test Method STD-191	Textile Test Methods.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

* 2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals, shall apply.

THE U. S. PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042).

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE
Bureau of Biologics. Food and Drug Administration
Code of Federal Regulation Part 273.
Minimum Requirements (as applicable).

(Application for copies should be addressed to the Food and Drug Administration, 5600 Fishers Lane, Rockville, Maryland 20852.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

AMERICAN NATIONAL STANDARDS INSTITUTE, INC.

A.N.S.I. Z70.1 American Standard Dimensions of Glass and Metal Luer Tapers of Medical Applications.

(Application for copies should be addressed to the American National Standards Institute Inc., 1430 Broadway, New York, N.Y. 10018).

3. REQUIREMENTS

3.1 Material.

* 3.1.1 Plastic components. All plastic components shall be of the best quality and suitable in all respects for the intended purpose. All plastic components shall be sterilizable, free from pyrogens, compatible, and safe (nontoxic) when tested as specified in section 4. Those plastic components coming in contact with solutions being administered, and end protectors, shall contain not more than 5 parts per million (p.p.m.) of lead when tested as specified in 4.4.1. The plastics shall meet the requirements of the Pharmacopeia of the United States (U.S.P.) for Class II plastics and for transfusion and infusion assemblies, when tested as specified in 4.4.2.

3.1.1.1 Needle adapter. The needle adapter shall be fabricated from a suitable rigid plastic.

3.1.1.2 End protectors. End protectors shall be fabricated from plastic tubes or sheaths.

3.1.2 Needles.

3.1.2.1 Cannulae of needles. Cannulae of needles shall be fabricated from 304 AISI stainless steel, and shall meet the corrosion resistance, elasticity, and resistance-to-breakage tests as specified in GG-N-196. The cannulae shall be fabricated from either seamless drawn, or welded and drawn tubing.

3.1.2.2 Hubs. Hubs shall be class 3, type 302, 303, or 304, or class 8, alloy number 260, 270, or 360 of GG-I-526; or hubs shall be plastic. Class 8 (brass) hubs shall be nickel plated.

3.1.3 Other material. All other materials as specified herein shall be of the best quality and suitable in all respects for the purpose intended.

3.2 Style, design and construction. Shall be spinal anesthesia set of sterile items consisting of a compartmented tray containing the following components, overwrapped with a durable plastic material:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

3.2 (continued).

Component	Quantity required
Dextrose Injection, USP, 10%, 5 cc ampul	1
Drape, waterproof, fenestrated	1
Epinephrine Injection, USP, 0.1%, 1 cc ampul	1
Needle, 22 gage, 3-1/2 inch	1
Needle, 25 gage, 3-1/2 inch, with introducer	1
Procaine Hydrochloride Injection, USP, 1%, 2 cc ampul	1
Sponge, 3 by 3 inch	4
Sterilization indicator	1
Swab with handle	2
Syringe, 2-1/2 cc, with 25 Gage, 5/8 inch needle	1
Syringe, 5 or 6 cc with 20 or 21 gage, 1-1/2 to 3 inch needle	1
Tetracaine hydrochloride, USP, 20 mg ampul, sterile	1
Towel	1

3.2.1 Dextrose injection. Shall be Dextrose Injection, USP, 10%, 5cc. The dextrose injection shall be in accordance with the requirements of the current Defense Medical Purchase Description for NSN 6505-00-116-4603, except as follows:

Page 1 of Purchase Description:

In the Item Identification - Delete "3 cc" and substitute "5 cc"

Delete the designation "10s".

Under "S5.2.1 - in line 11, delete "3-cc ampuls" and substitute: "5-cc ampuls."

Page 2 of Purchase Description:

S6.3 Volume in final container. Line 2, delete "3.0 cc" and substitute "5.0 cc".

In subparagraph "(c)" - delete "3 cc" and substitute "5 cc".

The name and address of the manufacturer need only appear on the ampul, providing the contractor's name and address appear on the anesthesia set.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

3.2.1 (continued).

Page 3 of Purchase Description:

The labeling of the unit packages shall not apply.

The circular, brochure, or other printed matter shall be furnished in each anesthesia set.

Page 4 of Purchase Description:

Packaging, Packing and Marking shall not apply.

* 3.2.2 Drape. The drape shall be waterproof and fenestrated, approximately 32 by 36 inches or 18 by 27 inches. The drape shall meet the requirement for waterproof when tested in accordance with 4.4.3 or 21-3/4 by 26 inches.

3.2.3 Epinephrine injection. Shall be Epinephrine Injection, USP, 1:1000, 1 cc. The epinephrine injection shall be in accordance with the requirements of the current Defense Medical Purchase Description for NSN 6505-00-299-8760, except as follows:

The designation "25s" shall not apply.

Page 2 of Purchase Description:

The name and address of the manufacturer need only appear on the ampul, providing the contractor's name and address appear on the anesthesia set.

Page 3 of Purchase Description:

The labeling of the unit packages shall not apply, except for the statement in subparagraph (h) which shall appear on the anesthesia set.

The circular, brochure, or other printed matter shall be furnished in each anesthesia set.

Page 4 of Purchase Description:

Packaging, Packing, and Marking shall not apply.

* 3.2.4 Needle, 22 gage. Shall be in strict accordance with MIL-N-36214 for size 2, except the packaging requirements of that specification shall not apply. As an alternate, the stylus knob shall be fabricated from a suitable plastic.

* 3.2.5 Needle, 25 gage. Shall be in strict accordance with MIL-N-36214 for size 1, except that packaging requirements of that specification shall not apply. As an alternate, the stylus knob shall be fabricated from a suitable plastic.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

3.2.6 Procaine hydrochloride injection. Shall be Procaine Hydrochloride Injection, USP, 1%, 2cc. Shall be in accordance with the requirements of the current Defense Medical Purchase Description for NSN 6505-00-138-0900, except as follows:

The designation "12s" shall not apply.

Page 2 of the Purchase Description:

The name and address of the manufacturer need only appear on the ampul, providing the contractor's name and address appear on the anesthesia set.

Page 3 of the Purchase Description:

The labeling of the unit packages shall not apply, except for the statements 1, 2, and 3 subparagraph (i) which shall appear on the anesthesia set.

The circular, brochure, or other printed matter shall be furnished in each anesthesia set.

Page 4 of the Purchase Description:

Packaging, Packing, and Marking shall not apply.

3.2.7 Sponge. The gauze sponge shall be 12 ply, approximately 3 inches by 3 inches when folded.

3.2.8 Sterilization indicator. A sterilization indicator showing a color change when sterilized shall be provided.

* 3.2.9 Swab with handle. The swab shall be foam sponge, and shall be approximately 1" x 1-5/8" x 1/2". Shall be securely mounted on a durable handle or 1-1/2" x 1-1/2" x 1".

* 3.2.10 Syringe with 25 gage needle. Shall be a 2-1/2 cc, or 3 cc disposable, plastic or glass syringe with needle, or as an alternate, the needle need not be attached to the syringe." At end of paragraph add the following: "Material and construction requirements of MIL-S-36157 shall not apply to syringes fabricated from glass. Syringe shall contain a standard male Luer hub, and needle shall contain a standard female Luer hub. The needle shall be 25 gage with a 5/8 inch length. The material, style, design, dimensions and construction shall be in accordance with MIL-S-36157 for Type II, size 2, except that packaging requirements of that specification shall not apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

* 3.2.11 Syringe with 20 or 21 gage needle, 5 or 6 cc. Shall be a 5 or 6 cc disposable, plastic or glass syringe with needle or as an alternate, the needle need not be attached to the syringe. Syringe shall contain a standard male Luer hub, and needle shall contain a standard female Luer hub. The needle shall be 20 or 21 gage with a 1-1/2 to 3 inch length. The material, style, design, dimensions and construction shall be in accordance with MIL-S-36157 for Type II, size 3 (or 6 cc), except the packaging requirements of that specification shall not apply. Material and construction requirements of MIL-S-36157 shall not apply to syringes fabricated from glass.

* 3.2.12 Tetracaine Hydrochloride. Shall be in accordance with the requirements of the current Defense Medical Purchase Description for NS 6505 for Sterile Tetracaine Hydrochloride, USP, 20 mg. As an alternate, the item may be designated as "Pontocaine Hydrochloride, brand of Tetracaine Hydrochloride, USP", in lieu of "Sterile Tetracaine Hydrochloride, USP", as required. The word "Sterile" shall appear on the set. The name and address of the contractor need not appear on the vial provided the name and address of the manfacturer appears. The statement 'Sterile powder for Spinal Anesthesia' need not appear on the vial. The date of manufacture may be stated in code on the vial label, providing the chemical compound is not older than six (6) months from shipment date of contract items.

* 3.2.13 Towel. Towel shall be a white cellulose wiper, minimum of 4 ply, approximately 14 inches by 18 inches or 17 inches by 19 inches.

3.3 Sterility. All components of the kit shall be sterile. When tested as specified in 4.4.4, the needles, towel, sponges, swabs, shall comply with the sterility test requirements of the U.S.P.

3.4 Toxicity. When tested as specified in 4.4.5, needles, towel, sponges, swabs, and syringes shall comply with the requirements for safety (toxicity) of the U.S.P.

3.5 Freedom from pyrogens. When tested as specified in 4.4.6, the fluid path, end connectors, and needles shall be pyrogen free.

3.6 Certificates of sterility, freedom from pyrogens, safety, and biological tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested and found to be sterile, free from pyrogens, nontoxic (safe), and have passed the biological tests. This certificate shall include the specific tests performed and the tests results. For destination inspection, certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, 2800 South 20th Street, Philadelphia, Pa. 19101, Attention: Quality Assurance Branch. One copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government inspector.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

3.7 Manufacturer's records. The manufacturer shall place a lot control number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made available for inspection by any duly qualified medical officer of the Armed Forces, National Institutes of Health, any official of the Federal Food and Drug Administration, or any duly authorized Government representative. No shipment shall be made until all the tests have been completed, verified, and recorded in the permanent records of the manfuacturer.

3.8 Compatibility. The materials of which the sets are manufactured shall not react with, or be acted upon by the liquids with which they will be in contact during normal use and operation, nor shall there be any materials incorporated in the set which will cause any deleterious effect when the set is used for the purpose intended.

3.9 Instructions. Each set shall be furnished with a list of the contents of the pack and procedural instructions for use.

3.10 Workmanship. Workmanship shall be first class throughout. All materials, treatments, and procedures used in the fabrication and assembly of the sets shall be such that the finished sets shall not show any evidence of decomposition, deterioration, or other conditions which may adversely affect the appearance or serviceability of the sets at time of delivery or during storage.

* "3.11 Delivery. Not less than 21 months of the expiration dating (potency period) shall remain at time of delivery to the Government."

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the plastic, metal, and cotton, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one anesthesia set.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

Inspection lot shall be the same as sterilization lot unless otherwise specified.

4.2.2 For tests.

4.2.2.1 For sterility test. Sampling for the sterility test shall be in accordance with the United States Pharmacopeia (U.S.P.) for parenteral devices.

4.2.2.2 For pyrogen, and safety tests. Sampling procedures for pyrogen, and safety tests shall be conducted as indicated in the U.S.P., for transfusion and infusion assemblies.

4.2.2.3 For lead test. Samples shall be taken from each lot of plastic as indicated in 4.4.

* 4.2.3 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

Table II. Sampling for Other Tests

Component (unit of product)	Characteristic	Req.	Test procedure	Insp. level	AQL (percent defective)
Plastic components	Material - Class II	3.1	4.4.1	**	*
	Plastic, lead (Biological)	3.1	4.4.2	**	*
Drape	Waterproof drape test	3.2.2	4.4.3	****	*
Set	Sterility	3.3	4.4.4	****	*
Set	Safety	3.4	4.4.5	****	*
Set	Freedom from pyrogen	3.5	4.4.6	***	*

* Acceptance number shall be zero for all sample sizes.
** Each lot of product need not be biologically tested provided evidence is furnished that samples of the formulation and identical method of processing have previously been tested and that they are in accordance with USP requirements for class II plastics and that they are the same as the material being furnished.
*** Sampling for tests shall be on component lots.
**** Sterility, Safety, and Freedom from pyrogen shall be tested by sampling sterility lots.

4.2.4 Sample unit usage. Where possible, the sample unit shall be used for determination of two or more test characteristics.

4.3 Examination. The sets shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the classification of defects specified in Table III. Inspection, however, shall not be restricted to the classified possible defects listed in Table III.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Set incomplete, one or more components missing.
102	Plastic or metal components not smoothly finished; not free of fits, blisters, scratches, tears, cracks, nicks, bubbles, roughness, tackiness, bloom or crazing, as applicable.
103	Components not free of foreign matter.
104	Towels not free from defects such as cuts and tears.
105	Labeling of drugs not as specified.
106	Swab not securely mounted on handle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects (Cont'd)

Minor	
201	Packaging, packing and marking not as specified.
202	Instructions missing, incomplete, illegible.
203	Poor workmanship (not resulting in a major defect).

4.4 Tests.

* 4.4.1 Test for lead. A 2 gram composite sample shall be obtained from plastic components. The material shall be cut into small slivers, so that no thickness exceeds 3 mm., and eluted for 72 hours at 50° C., 24 hours at 70° C., or 60 minutes at 121° C. in two menstrua: 0.9% sodium chloride and cottonseed oil. Each gram of the material shall be covered with 5 milliliters (ml.) of the menstruum. The menstrua are collected, agitated after elution, and then submitted to spectographic analysis for lead. The lead content shall be in parts per million. Thus 1.0 ppm in elute is equivalent to extracting 1.0 mg. from 200 g test material dispersed in 1,000 ml of elution menstruum. Alternates to the above manner specified in determining lead content may be considered acceptable, provided the alternate method is formally submitted to the contracting officer for approval, prior to shipment of contracted items.

4.4.2 Biological tests. The plastics shall be tested in accordance with the U.S.P. for Class II plastics.

4.4.3 Waterproof drape test. The drape shall be tested for waterproof requirement in accordance with Test Method 5526 of Fed. Test Method Std. 191.

4.4.4 Sterility test. The kits shall be tested for sterility in accordance with U.S.P. The testing for sterility shall be performed on at least needles, syringes, towel, sponges and swabs as representative of the sterility of all the set components.

4.4.5 Safety (toxicity) test. Samples of the finished kits shall be tested for safety, in accordance with the U.S.P., for transfusion and infusion assemblies.

4.4.6 Freedom from pyrogens. Samples of the finished kits shall be tested for freedom from pyrogens, in accordance with the U.S.P., for transfusion and infusion assemblies.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One anesthesia set, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	20 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

* 5.1.3.1 Unit package. Components comprising each anesthesia set shall be contained in individual cells of a preformed plastic or polystyrene tray of appropriate size and design. Each tray shall be wrapped in such a manner as to present an uncontaminated working surface when opened. Each wrapped tray shall be further packaged in paperboard or plastic container of appropriate size and design except that commercial colors are acceptable. The container shall be wrapped in plastic wrap and wrap secured; or enclosed in a heat sealed polyethylene bag. The unit shall be packaged in such a manner as to maintain sterility of the tray and contents unless the package is damaged or opened. As an alternate, the wrapped tray may be placed in a heat-sealed bag and further over-packaged in a paperboard box of appropriate size and design. Closure of box shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification. In addition, each fiberboard box shall be waterproofed in accordance with 30.4 of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

* 5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636. Fiberboard boxes shall conform to the special requirements as specified in Table I of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

* 5.3.1 Unit package. In addition to any commercial markings, each unit package shall be marked in accordance with MIL-STD-129 and the requirements of the Federal Food, Drug and Cosmetic Act. Lot (control) number, contract or purchase order number and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, S-Steam. D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A-1/70-G." Final sterilization shall be on or after indicated packaging date. In addition, Type I Shelf-Life markings as specified in 5.2.2.1 (a) of MIL-STD-129 shall be shown, except that the date of manufacturer shall not be required. Marking shall include the legends:

"DO NOT PERMIT TO FREEZE" and
"STERILITY GUARANTEED UNLESS
PACKAGE IS DAMAGED OR OPENED"

Each package shall be prominently marked indicating a warning not to remove plastic film in an area where static electrical discharge could be hazardous.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

"NOTE: The labeling (on the immediate containers, the individual carton (box) and unit packages) and other accompanying data, shall not include any statement with respect to refrigerated storage, storage in a cold place, storage at 2° -10°C. or 2° -8°C., or in a freezer, unless specifically required in the procurement document; and such labeling and other accompanying data, shall not bear an expiration date if the procurement document does not include a potency period (expiration dating period) for the item. When a potency period (expiration dating period) is specified in the procurement document, the labels shall not set forth an expiration date which is longer or shorter than the potency period specified in such procurement document, unless specifically authorized in writing by the procuring activity.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number, and method of sterilization as specified in 5.3.1 shall be shown. Type I Shelf Life marking as specified in 5.5.1 (a) of MIL-STD-129 shall be shown, except the date of manufacture shall not be required. Marking shall include the legend:

"DO NOT PERMIT TO FREEZE".

Note. The item identification shall be shown on the unit, intermediate, and exterior container as follows:

"ANESTHESIA SET, SPINAL, DISPOSABLE"

6. NOTES

6.1 Intended use. The anesthesia set is a spinal set suitable for use in anesthetizing patients for surgical procedures.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number, and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 National Stock Number (NSN) coverage. This specification is intended to cover the following item appearing in the Federal Supply Catalog: NSN 6515-00-453-4609 ANESTHESIA SET, SPINAL, Disposable. It does not include all types, sizes, or grades as indicated by the title or all types which are commercially available.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

* 6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1809

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-A-36763A

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.