MIL-A-36763A
AMENDMENT-2
10 October 1974
SUPERSEDING
AMENDMENT-1
4 June 1974

MILITARY SPECIFICATION

ANESTHESIA SET, SPINAL, DISPOSABLE

This Amendment forms a part of Military Specification MIL-A-36763A, dated 14 March 1974, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 5:

3.3 Line 1, delete second sentence entirely and substitute: "When tested with a biological indicator as specified in 4.4.4, the needles, towel, sponges, and swabs shall be sterile."

Add the following new paragraphs:

"3.3.1 Certificates of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and date of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or designated qualified alternate.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

When inspection is made at destination, the certificates of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Assurance Representative for submission to the procuring activity.

"3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s)."

* Page 10:

4.2.3 In Table II under characteristic delete "Plastic, lead (Biological)" and substitute "Biological test." In Table II for Drape delete Insp level "****" and substitute "S-2", and delete AQL of "*" and substitute "2.5".

Page 11:

* 4.4.4 Delete entirely and substitute:

* "4.4.4 Sterility test. The anesthesia kit shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). The testing for sterility shall be performed on at least needles, syringes, towels, sponges and swabs as representative of the sterility of all the set components. Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in Table II, Sampling Criteria for Tests.

* 4.4.4.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65° C
Dry Heat	B. stearothermophilus or	55° - 65° C
	B. subtilis var. niger (or globigii)	30° - 35° C

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

Type of Sterilization	Organism	Incubation Temperature
Gas	B. subtillis var. niger (or globigii)	30° - 35° C
Radiation	B. pumilus	30° - 35° C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

* "4.4.4.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.2 Sterility test and 4.4.2.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.2.1 NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

* "4.4.4.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract."

Page 5:

5.3.1 Unit package. The legend "STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86° F)" shall also be shown.

5.3.2 Exterior container. Line 3, delete "5.5.1a" and substitute: "5.4.1a". After "DO NOT PERMIT TO FREEZE" add: "STORE AT CONTROLLED ROOM TEMPERATURE (59° - 86° F)".

Add the following new paragraph:

"6.7 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue."

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1878

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:16Z
Check the source to verify that this is the current version before use.