MIL-F-36733A
28 July 1986
Superseding
MIL-F-36733
6 January 1971

MILITARY SPECIFICATION

FILE, BONE DENTAL, DOUBLE-ENDED, Seldin, No. 11

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a file of the Seldin design suitable for filing the bone processes of the jaw in oral surgery.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specification and Standards (DoDISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding Paperboard.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-B-676	Boxes, Set-up.

AMSC N/A FSC 6520

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia , PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

DISTRIBUTION STATEMENT A. Approved for public release; distribution

Military

MIL-P-116	- Preservation, Methods of
MIL-L-10547	- Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-151	- Metals, Test Methods.

Military

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

(Copies of specifications, standards and drawings required by contractors in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting activity.)

DRAWING

Defense Personnel Support Center (DPSC)

10072 FILE, BONE, SELDIN, No. 11

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of the documents which are DoD adopted shall be those listed in the issue of the DoDISS specified in the solicitation. Unless otherwise specified, the issues of the documents not listed in the DoDISS shall be the issue of the nongovernment documents which is current on the date of the solicitation.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

E 18	Standard Test Methods for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
E 92	Standard Test Method for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa 19103).

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3.1 General requirements. The file shall be in accordance with the applicable requirements of GG-I-526 with the options and additional requirements specified herein.

3.2 Material. The file shall be fabricated from class 4, type 440A martensitic corrosion-resisting steel, when tested as specified in 4.4.1.

3.3 Design and dimensions. The bone file shall be of Seldin No. 11 design and shall be suitable for filing the bone processes of the jaw in oral surgery. The blades shall be contra-angled from centerline of handle, and shall have deep teeth faced toward the handle for a pull stroke. The design and dimensions of the file shall conform to Defense Personnel Support Center (DPSC) Drawing 10072.

3.4 Construction. The file shall be of one piece, double-ended construction. The handle shall be solid.

3.5 Hardness. Operative ends of the file shall have a Rockwell hardness of C52 to C56 after proper stress relieving, when tested as specified in 4.4.2.

3.6 Corrosion resistance. The instruments shall be properly passivated, and shall pass the boil test and copper sulfate test requirements specified in GG-I-526.

3.7 Finish. The operative ends (except serrated portions) and shanks shall have a mirror finish. Handles shall have a satin or bright polished finish. Teeth of operative ends shall be sharp and clean cut with a fine, honed finish.

3.8 Identification marking. The handle of each file shall be permanently and legibly marked with the number "11".

3.9 Workmanship. The finished file shall be free from cracks, burrs, nicks, pits or other defects which detract from its appearance or may impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilites suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in tables I, II, and III (see 4.2.1, 4.2.2 and 4.3) and 4.3.1.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQLs) shall be as indicated in table I. Unit of product for examination purposes shall be one file.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination.

Type Defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one file.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level
Operative end	Composition	3.2	4.4.1	*
	Hardness	3.5	4.4.2	S-1
	Copper sulfate test	3.6	GG-I-526	S-2
	Boil Test	3.6	GG-I-526	S-2

*Sufficient material from each lot of metal used in the construction of the files shall be selected for composition analysis.

4.2.3 Sampling unit usage. Where feasible, the same sample unit shall be tested for the determination of two or more test characteristics.

4.3 Examination. The files shall be examined for defects including but not limited to, those listed in table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects.

Categories	Defects
Major A:	
101	Design not as specified.
102	Instrument not of one piece construction.
103	Operative end finish not as specified.
104	Operative end not free from cracks or burrs.
105	Working surfaces of instrument not free from nicks, feathers, pits, burrs or corrosion.
106	Teeth not file sharp.
Major B:	
151	Operative end dimensions not as specified.
Minor A:	
201	Handle finish not as specified.
202	Instrument not permanently or legibly marked with name or registered trademark of the manufacturer.
203	Instrument number not present.
Minor B:	
251	Handle dimensions not as specified.

4.4 Tests. Tests shall be conducted to determine compliance with the requirements of this specification.

4.4.1 Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instruments shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of conflict, method 111 shall govern.

4.4.2 Hardness. The hardness shall be determined in accordance with ASTM Desination E18 or E92.

4.5 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Files shall be cleaned and dried in accordance with MIL-P-116.

5.1.1.2 Preservative applications. Preservatives shall not be used.

5.1.1.3 Unit package. Each file shall be packaged in accordance with MIL-P-116, method IC-1.

5.1.1.4 Intermediate package. Twelve unit packages shall be packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the applicable box specification.

5.1.2 Level C. The level C preservation for files shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Twenty-four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, domestic type, PPP-B-621, class 1 or PPP-B-636, class domestic. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. The level C packing for files shall conform to the MIL-STD-794 requirements for this level.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T09:14Z
Check the source to verify that this is the current version before use.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length, and shall conform to Type IV or Type V.

5.2.5.1 Levels A and B. Files packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets conforming to MIL-STD-147.

5.2.5.2 Level C. Files packed as specified in 5.2.3 shall be unitized as specified in 5.2.5.

5.3 Marking.

5.3.1 Levels A, B, and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Ordering data.

6.1.1 Acquisition requirements. Acquisition documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following item listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6520-00-528-5050	FILE, BONE, Seldin, No. 11

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

6.3 Limitations. This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

6.4 Subject term (key word) listing.

Bone File
Dental File
Double-Ended File
Seldin No. 11

MILITARY INTERESTS:

Custodians.

Army - MD
Navy - MS
Air Force - 03

Review Activities:

Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordination Activity:
VA-OSS

47966

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1937

Location: NCR

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use i
bmitting comments and suggestions for improvements. All users of military standardization documents are invited to provid
uggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and
mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, wa
too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the
problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an
acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being
considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of
specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization
to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

7

Fold along this line)

DEPARTMENT OF THE ARMY

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATE

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 12062 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEPARTMENT OF THE ARMY

Commander
Defense Personnel Support Center
Directorate of Medical Materiel
ATTN: DPSC-RST (9-3-F)
2800 South 20th Street
Philadelphia PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-F-36733A	FILE, BONE DENTAL, DOUBLE-ENDED, Seldin, No. 11

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.