MILITARY SPECIFICATION

FUNNEL, COMMON, LABORATORY

Polyethylene, Outside Ribbed

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers two sizes of ribbed polyethylene laboratory funnels.

1.2 Classification. Funnels shall be the following sizes:

Size 1 - 2 quart nominal capacity.
Size 2 - 4 quart nominal capacity.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bid or request for proposals form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-123	Marking for Domestic Shipment (for Civilian Agencies).

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS

ASTM D-1248	Polyethylene Plastics Molding and Extrusion Materials.
ASTM D-898	Plastics, Recommended Practice for Sampling of.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103.)

3. REQUIREMENTS

3.1 Material. Material shall be of the highest quality throughout and shall be as herein specified.

3.1.1 Polyethylene. Polyethylene shall be in accordance with ASTM Designation D 1248, type I, class A, grade 3.

3.2 Construction. The funnel shall consist of a bowl and stem. Top outside perimeter shall have a reinforcing bead, and bottom of stem shall be cut at an angle. The funnel shall be fabricated of polyethylene and shall be constructed in accordance with figure 1. The funnel shall be so constructed that the longitudinal axes of the bowl and stem combine

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

to form a straight longitudinal axis for the funnel. Inner perimeters of the bowl and stem shall be circular and the longitudinal axis of the funnel shall coincide with the centers of all circles. Walls of bowl and stem shall be of uniform thickness. Funnel shall have four longitudinal ribs which are integral with the exterior surfaces of the bowl and stem. Ribs shall be spaced equally 90° apart around the bowl and stem. Height and width of ribs shall be such that when the stem is placed in a round hole (such as the hole in a container the ribs separate the main portion of the stem or bowl and the container sufficiently to prevent a seal. The thickness or height of the ribs shall be greatest at the juncture of the bowl and stem.

3.2.1 Interior ribs and hanger. Interior ribs and hanger are not required but may be furnished. If internal ribs are furnished, they shall be integral with the interior of the bowl, shall be thickest (highest) near the stem and extend and taper toward the top perimeter, and shall be equally spaced around the bowl. If hanger is furnished, it shall consist of an integral projection protruding from the top outside perimeter. Projection shall have a hole in it which shall make the funnel suitable for hanging on a nail or peg.

3.3 Identification marking. The exterior surface of each funnel shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

3.4 Workmanship. Workmanship shall be first class throughout. Funnels shall be free of defects listed in table II and other defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, or auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificate of quality. Certificate of quality, supplied by the manufacturer of the polyethylene, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. Unit of product shall be one funnel.

Table I. Sampling for examination

Type Defect	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For testing. Sample size of a lot of raw material shall be as recommended in ASTM D 1898, Note 6. Lot shall be as defined in MIL-STD-105.

4.3 Examination. The funnel shall be examined for defects including, but not limited to, those listed in table II.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

Table II. Classification of defects

Category	Defect
Major	
101	Funnel not free of cracks.
102	Funnel not free of deformation.
103	Funnel not free of sharp edges.
104	Funnel not free of tears and splits.
105	Funnel not free of blisters.
106	Ribs not continuous or clean cut.
Minor	
201	Funnel not free of protrusions (other than ribs, identification marking and handle).
202	Surfaces not smooth.
203	Funnel not free of deep scratches.
204	Identification marking not complete, correct, legible or permanent.

4.3.1 Dimensional examination. The funnels shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements (see 3.1.1).

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One funnel, of the size specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III. Packaging Quantities

Size	Unit package	Intermediate package	Exterior container
1 or 2	1 unit	Not Required	6 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with paragraph 30.4 of PPP-B-636.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipments shall be in accordance with FED-STD-123.

5.3.2 Military activities.

5.3.2.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. Funnels are intended to be used for pouring liquids into large containers.

6.2 Ordering data. Purchasers should exercise any desired options offered herein and procurement documents should specify the following:

a. Title, number, and date of this specification.
b. Size and capacity of funnels desired (see 1.2.1).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
d. When weather-resistant grade fiberboard shipping containers are required for level B packing (see 5.2.2.1).
e. Federal stock number (FSN).

6.3 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03


Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0687

Review activities:
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.



FUNNEL
FIGURE 1

SIZE	"A" ±⅛	"B" ±⅛	"C" ±.10	"D" ±.10	"E" ±½	"F"	"G" MIN	"H" MIN.	"J" MIN.
1	$7\frac{29}{32}$	$7\frac{3}{16}$	.80	.52	$3\frac{1}{16}$	$9\pm\frac{1}{4}$	$\frac{1}{16}$	2	$2\frac{1}{4}$
2	$11\frac{3}{8}$	$10\frac{3}{8}$	.85	.63	$3\frac{1}{2}$	$11\frac{1}{2}\pm\frac{1}{4}$	$\frac{1}{8}$	3	$2\frac{1}{4}$

DIMENSIONS ARE IN INCHES

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document (s) or serve to amend contractual requirements.

SPECIFICATION

MIL-F-36748(DSA-DM) FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 66 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.