MIL-H-0036734B (DSA-DM)
22 Dec 1976
USED IN LIEU OF
MIL-H-36734A
16 July 1975

MILITARY SPECIFICATION

HANDPIECE, DENTAL, CONTRA-ANGLE, ROUND SLEEVE, LATCH TYPE

This limited coordination military specification has been prepared by the Defense Personnel Support Center, Defense Supply Agency based upon currently available technical information, but it has not been approved for promulgation as a coordinated revision of MIL-H-36734A. It is subject to modification. However, pending its promulgation as a coordinated military specification, it may be used in procurement.

1. SCOPE

1.1 Scope. This specification covers a latch type, round sleeve, contra-angle dental handpiece for attachment and use with a belt or air driven straight handpiece with a round nose.

1.2 Classification. The handpieces covered by this specification shall be of the following styles, as specified (see 6.2):

Style 1 - Handpiece, non-adjustable head, external latch.

Style 2 - Handpiece, adjustable head, external latch.

Style 3 - Handpiece, adjustable head, internal latch.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 S. 20th St., Phila., PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup.

MILITARY

MIL-P-116 - Methods of Preservation-packaging.
MIL-M-7298 - Manual, Technical, Commercial Equipment.
MIL-L-10547 - Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

FEDERAL

FED-STD-151 - Metal, Test Methods.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24252 - HANDPIECE, DENTAL, CONTRA-ANGLE, ROUND SLEEVE, LATCH TYPE.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

B 16 Specification for Free-Cutting Brass Rod, Bar, and Shapes for Use in Screw Machines.

B 584 Specification for Copper Alloy Sand Castings for General Applications.

E 18 Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.

E 92 Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

3. REQUIREMENTS

3.1 Materials, hardnesses, and finishes.

3.1.1 Style 1. The handpiece components materials, hardnesses, and finishes for style 1 shall be as specified in table I.

TABLE I. Materials, hardnesses, and finishes for style 1.

Part	Component	Material	Rockwell hardness	Finish
Head assembly	Head shell	Brass	---	Chrome over nickel
	Cap	Stainless steel, type 440F	C58-62 1/	Passivated
	Bur tube	Stainless steel, type 440F	C58-62 1/	---
	Bearing	Stainless steel, type 440F	C58-62 1/	---
	Latch	Stainless steel, type 420	C45-50 1/	Passivated
	Latch screw	Stainless steel, type 416	---	Passivated
	Gear	Stainless steel, type 440F	C58-62 1/	---
	Slinger	Stainless steel, type 416	---	Passivated
Neck assembly	Neck bearing	Stainless steel, type 416	C37-45 1/	---
	Gear	Stainless steel, type 440F	C58-62 1/	---
	Shaft	Stainless steel, type 440F	C58-62 1/	---
	Rivet	Brass, half hard	---	---

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

TABLE I. Materials, hardnesses, and finishes for style 1. (Continued)

Part	Component	Material	Rockwell hardness	Finish
Knee assembly	Knee	Brass	---	Chrome over nickel
Round sleeve back end assembly	Round sleeve back end	Phosphorus bronze	---	Chrome over nickel
	Shaft	Stainless steel, type 440F	C58-62	---
	Ring	Stainless steel, type 440F	C58-62	---
	Gear	Stainless steel, type 440F	C58-62	---
	Rivet	Brass, half hard	---	---
Wrench	Octagon cap wrench	Steel, AISI B1113	2/	Nickel plate

1/ Hardness after proper tempering.
2/ Hardened and tempered.

Note: Bur tube and gear assembly hard chrome plated 0.002 inch, minimum, before grinding.

3.1.2 Style 2. The handpiece components materials, hardnesses, and finishes for style 2 shall be as specified in table II.

TABLE II. Materials, hardnesses, and finishes for style 2.

Part	Component	Material	Rockwell hardness 1/	Finish
Head assembly	Head shell	Cast brass, ASTM B-584	---	Chrome
	Top bearing	Phosphorus bronze, #544	---	---
	Latch	Spring steel, AISI 1070	---	Chrome or chromium alloy 3/ 5/
	Latch screw	Steel, AISI C1112-13 or 12L14	---	Black oxide dip

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

TABLE II. Materials, hardnesses, and finishes for style 2. (Cont'd)

Part	Component	Material	Rockwell hardness	Finish
Spindle	Spindle	Steel, AISI 01 drill rod	C 53-58	Chrome or chromium alloy 3/ 5/
	Ring gear	Steel, AISI 01 drill rod	C 53-58	Chrome or chromium alloy 3/ 5/
	Lower bearing	Phosphorus bronze, #544	------	----------
	Slinger	Stainless steel, type 303 or 416	------	----------
Head to elbow bearing assembly	Bearing	Steel, AISI 12L14	Case hardened 2/	4/
	Pinion gear	Steel, AISI 01 drill rod	C 53-58	Chrome or chromium alloy 3/ 5/
	Shaft	Steel, AISI W1 or 01 drill rod	------	4/
	Elbow gear	Steel, AISI 01 drill rod	C 53-58	Chrome or chromium alloy 3/ 5/
Elbow and rear end assembly	Elbow	Brass, ASTM B-16	------	Chrome
	Elbow gear	Steel, AISI 01 drill rod	C 53-58	Chrome or chromium alloy 3/ 5/
	Rear end housing	Brass, ASTM B-16	------	Chrome
	Rear end shaft	Steel, AISI W1 or 01 drill rod	------	4/
	E Ring	Stainless or carbon spring steel	C 48 min.	4/
Head locknut	Head locknut	Steel, AISI 12L14	Case hardened 2/	Chrome plate
Wrench	Wrench	Steel, AISI 1010, 1/2 hard	Case hardened 2/	Nickel plate

1/ Hardness after proper tempering.
2/ Minimum depth 0.0015 inch.
3/ Minimum thickness 0.0001 inch.
4/ Carbon steel shall be coated or plated to render the part corrosion-resistant.
5/ See 6.4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

3.1.3 Style 3. The handpiece components materials, hardnesses, and finishes for style 3 shall be as specified in table III.

TABLE III. Materials, hardnesses, and finishes for style 3.

Component	Material	Hardness 1/ or finish
Latch head cap	Stainless steel, type 303	
Threaded bearing	Stainless steel, type 440FSe	C54-58 2/
Bur tube gear	Stainless steel, type 440FSe	C58-60 2/
Bur tube	Stainless steel, type 440FSe	C54-58 2/
Thrust bearing	Stainless steel, type 440FSe	C58-60 2/
Head	Brass, #360	Satin chromium
Swivel bearing drive gear	Stainless steel, type 440FSe	C54-58 2/
Intermediate shaft	Stainless steel, type 440FSe	
Extender bearing	Brass, #360	Satin chromium
Liner bearing	Brass, #360	
Swivel bearing	Stainless steel, type 303	
Swivel bearing pin	Stainless steel, (type 304) wire 3/	Spring type
Swivel nut	Brass, #360	Satin chromium
Swivel bearing driven gear	Stainless steel, type 440FSe	C54-58 2/
Contra-sleeve drive gear	Stainless steel, type 440FSe	C54-58 2/
Roll pin (3 required)	Stainless steel, type 420	C46-55 2/
Contra-sleeve bearing	Phosphorous bronze, #510	
Drive shaft	Stainless steel, type 440FSe	C54-58 2/
Drive shaft collar	Stainless steel, type 303	
Contra-sleeve	Brass, #360	Satin chromium
Pilot bearing	Phosphorous bronze, #510	

1/ Hardness after proper tempering.
2/ Rockwell hardness.
3/ Minimum thickness 0.020 inch.

3.2 Style, design and dimensions. The handpiece shall be of the type known to the dental trade as a latch type, round sleeve, contra-angle handpiece. The handpiece shall be entirely suitable for attachment and use with belt or air driven, round nose straight handpieces (see 6.1). The contra-angle handpiece shall be suitable for use in the performance of operative dentistry. The handpiece shall consist essentially of a head assembly, neck assembly, knee (elbow) assembly and round sleeve back end assembly. The handpiece shall be a durable, precision made handpiece

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

entirely suitable for operation at handpiece speeds of up to at least 12,000 revolutions per minute (rpm) as a maximum speed. The design and dimensions of the handpiece shall be in strict accordance with Defense Personnel Support Center (DPSC) Drawing No. 24252 for style 1, 2 or 3, as specified (see 6.2). The requirements specified herein for one style shall apply to that style only, and shall not be interchangeable with the other style. Style 1 handpiece shall be of a non-adjustable head, external latch design. Style 2 handpiece shall be of an external latch, adjustable head design to provide for adjustment of the angulation of the head with respect to the back end assembly. Style 3 handpiece shall be of an internal latch, adjustable head design.

3.2.1 Head assembly. Head assembly shall have one fixed bearing and one screwed-in bearing with a minimum total bearing surface of 0.152 square inch for style 1 and 2, and 0.090 square inch for style 3. The maximum clearance between the fixed bearing and the bur socket shall not exceed 0.001 inch throughout the entire length of the bearing. The fixed bearing shall be attached to the outer shell of the head assembly in such a manner that the required tolerance is maintained. A grease chamber shall surround the fixed bearing. As an alternate, a grease recess shall be put into the entire inside periphery of the fixed bearing. The gear mechanism between the bur socket and the drive shaft shall have gears that have a line contact. One gear shall be fixed on the bur socket, and the other gear shall be fixed on the drive shaft.

3.2.1.1 Bur holding. Bur socket for retaining bur shall receive the conventional design angle bur shank 0.0919 to 0.0925 inch in diameter. A latch of suitable design shall be provided for locking bur in place. To insure interchangeability with standard angle burs, a standard test gauge (conforming to dimensions shown in DPSC Drawing 24252) must assemble into socket and latch for retaining bur, and must close without binding on gauge or interfering with free rotation of the test gauge by the socket. Total end play of gauge in socket with reference to head of the angle shall not exceed 0.0095 inch. Side play of gauge measured at 3/8 inch from nosepiece shall not exceed 0.010 inch. The shank of test gauge shall clear nosepiece by 3/8 inch plus 1/32, minus 1/64 inch. End play of bur tube shall range between 0.0001 and 0.0025 inch for styles 1 and 2, and 0.001 to 0.004 inch for style 3, when tested as specified in 4.5.3. The internal latching mechanism of style 3 handpieces shall show no failure in proper latching and unlatching of bur, when tested as specified in 4.5.7.

3.2.2 Neck and knee (elbow) assemblies. The neck assembly shall be so designed that it can be removed intact with the bearing and drive shaft gear so that the head assembly can be properly cleaned and lubricated. The knee (elbow) assembly shall be provided with the conventional universal joint gear mechanism.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

3.2.3 Round sleeve back end assembly. The round sleeve back end assembly shall consist of a sheath and a drive shaft. The sheath shall be knurled on the outside to provide a gripping surface, and the interior surface portion of the round sleeve shall fit the round nose portion of the nosepiece on the straight handpiece (see 3.2).

3.3 Runout. The total runout of the angle handpiece when measured with the arbor shall not exceed 0.005 inch at 3/8 inch outside the bur tube point when tested as specified in 4.5.4.

3.4 Vibration and heat rise. The handpieces shall show no excessive vibration and heat rise when tested as specified in 4.5.5.

3.5 Drive train smoothness and backlash. The drive train of the handpiece shall operate smoothly with no backlash when tested as specified in 4.5.6.

3.6 Finish. The external surfaces of the handpiece shall be smooth. The finish shall be bright or satin finish except for the knurled portion.

3.7 Autoclaving and disinfecting. Instructions shall be provided for sterilization and disinfecting, as applicable.

3.8 Lubrication. Instructions shall be provided for proper lubrication of the handpiece components and surfaces requiring lubrication for proper operation and maintenance.

3.9 Tools. Applicable tools such as wrenches shall be supplied for self repairable components and for routine maintenance.

3.10 Identification marking. Each handpiece shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

3.11 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance, and lists of component and replacement parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should be supplier have furnished

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.12 Workmanship. Workmanship shall be first class throughout. Handpieces shall be free from defects which detract from their appearance or which may impair their serviceability.

3.13 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper design, function, fit, assembly, and operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials used in the handpieces, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in tables IV and VI (see 4.2.1, 4.3. and 4.4) and table V (see 4.2.2).

4.2.1 Sampling for examination. Sampling for examination for the handpiece shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table IV. The unit of product for examination purposes shall be one handpiece.

TABLE IV. Examination criteria.

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor A	II	2.5
Minor B	S-1	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table V for the handpiece. The acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one handpiece. Where feasible, the same sample unit may be used for the determination of two or more test characteristics.

TABLE V. Test criteria.

Component	Characteristic	Requirement	Test procedure	Inspection level
All components in tables I, II and III	Chemical analysis	3.1	4.5.1	*
Components indicated in tables I, II and III	Hardness	3.1	4.5.2	S-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

TABLE V. Test criteria. (Cont'd)

Component	Characteristic	Requirement	Test procedure	Inspection level
Assembled handpiece	End play	3.2.1.1	4.5.3	S-1
Assembled handpiece (style 3 only)	Latching	3.2.1.1	4.5.7	**
Assembled handpiece	Runout	3.3	4.5.4	S-1
Assembled handpiece	Vibration and heat rise	3.4	4.5.5	S-1
Assembled handpiece	Drive train smoothness and backlash	3.5	4.5.6	100%

* A sufficient amount from each lot of raw materials used in the construction of the item shall be selected for chemical analysis when certificates of quality are not available as specified in 4.1.3.

** One handpiece from the initial inspection lot shall be tested per contract or purchase order.

4.3 Examination. The handpiece shall be examined to determine compliance with requirements contained in this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in table VI.

TABLE VI. Classification of defects.

Categories	Defects
Critical	None defined.
Major A	
101	Design and construction of style 1, 2, or 3 not as specified.
102	Round nose straight handpiece will not accept and lock in handpiece (see 3.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

TABLE VI. Classification of defects. (Cont'd)

Categories	Defects
Major A	
103	Latch will not lock bur in place (style 1 and 2 only, see 3.2.1.1).
104	Slivers or burs present.
Major B	
151	Dimensions affecting fit of handpiece to round nose straight handpiece not as specified.
152	Dimensions affecting fit of bur shank to bur socket not as specified.
Minor A	
201	Service data, technical manuals missing.
202	Identification marking not present, correct, complete, or legible.
203	Pits, knicks, or discoloration present on outer surfaces.
204	Plating not free of defects.
205	Knurling not as specified.
206	Instructions for sterilization, autoclaving, and lubrication not included.
Minor B	
251	Dimensions (other than 151 and 152) not as specified.

4.4 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.5 Tests.

4.5.1 Chemical analysis. Metal analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151.

4.5.2 Hardness test. Hardness shall be determined in accordance with ASTM Standard E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

4.5.3 End play test. For style 1 and 2 handpieces, end play of socket shall be determined with a dial indicator, sensitive to 0.0002 inch. The socket shall be pushed toward and away from the head of the handpiece, and the end play of the socket alone shall be measured. For style 3 handpieces, the end play shall be determined by measuring dimensions "A" and "B" as shown in Drawing 24252.

4.5.4 Runout. The runout shall be measured by inserting a true-running arbor 0.0925 inch in diameter into the bur tube. At the end of the arbor, an indicator shall be placed measuring in increments of 0.001 inch. The arbor shall be rotated through 360 degrees with maximum and minimum readings recorded. The difference between these readings shall be the total runout.

4.5.5 Vibration and heat rise. The assembled and lubricated handpiece shall be tested by running the handpiece, containing a dental bur, for three minutes at 10,000 rpm under free load. The handpiece shall be tested for excessive vibration when hand held, and for excessive heat rise (above 115° Fahrenheit) when tested with a pyrometer held against the head housing and against the back end housing.

4.5.6 Drive train smoothness and backlash test. The assembled and lubricated handpiece shall be tested by spinning the back end drive shaft through at least two full revolutions, both clockwise and counterclockwise, using finger pressure only as applied by the thumb and index finger. Any detectable binding, slipping, backlash, or irregular mechanical interference shall cause the handpiece to be rejected.

4.5.7 Latching test (style 3 only). The internal latching mechanism shall be loaded and unloaded for 50,000 cycles. A cycle shall consist of one loading and one unloading operation. Loading is defined as engagement of a standard latch type bur. Unloading is defined as disengagement.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. Each handpiece of the style specified shall be thoroughly cleaned and dried in accordance with the applicable method specified in MIL-P-116. Each handpiece shall be wrapped in neutral tissue paper or other suitable anticorrosive material and packaged in accordance with MIL-P-116, method IC-2 except that a preservative shall not be used.

5.1.1.2 Intermediate package. Twelve units of one style only shall be packaged in a box constructed in accordance with PPP-B-566, PPP-B-676 or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

5.1.2 Level C. Each handpiece shall be clean, dry, and individually packaged in a manner that will afford adequate protection against corrosion, deterioration, and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B, or C as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages (288 units) of one style shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraph of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Twenty-four intermediate packages (288 units) of one style shall be packed in an exterior container designed for a type 2 laod and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601 domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged handpieces shall be packaged in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The contra-angle handpiece covered by this specification is intended for use with Teledyne Emesco's "No. 4N" Handpiece, Midwest American Company's "Tru-Torc" Handpiece and Star Dental Manufacturing Company's "Titan-2" Handpiece with "T-10A" Straight Nose Cone.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

(a) Title, number, and date of this specification.
(b) Style required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 National Stock Number (NSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

NSN	Item Identification
6520-00-935-7142	HANDPIECE, DENTAL, Contra-angle, Round Sleeve: Latch type.

6.4 "Electrolizing" and "Duroplate" chromium alloy coatings are acceptable.

6.5 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue due to the extensiveness of the changes.

Preparing activity:
DSA-DM

Project No. 6520-1453

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-H-0036734B (DSA-DM) HANDPIECE, DENTAL, CONTRA-ANGLE, ROUND

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -: Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.