MIL-H-36734A
16 July 1975
SUPERSEDING
MIL-H-36734
6 January 1971

MILITARY SPECIFICATION

HANDPIECE, DENTAL, CONTRA-ANGLE, ROUND SLEEVE, LATCH TYPE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of a conventional latch type, round sleeve, contra-angle dental handpiece for attachment and use with a belt or air driven straight handpiece with a round nose.

1.2 Classification. The handpieces covered by this specification shall be of the following styles, as specified (see 6.1):

Style 1 - Handpiece, non-adjustable head.

Style 2 - Handpiece, adjustable head.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards, and drawing. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-B-676	Boxes, Setup.

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

Military

MIL-P-116	Methods of Preservation.
MIL-M-7298	Manual, Technical Commercial Equipment.
MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Federal

FED-STD-151	Metal, Test Methods.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

24252	Handpiece, Dental, Contra-Angle, Round Sleeve, Latch Type.

(Copies of specifications, standards, drawings and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

B 16-74	Specification for Free-Cutting Brass Rod, Bar, and Shapes for Use in Screw Machines.
E 18-74	Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
E 92-72	Tests for Vickers Hardness of Metallic Materials.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

3.1 Materials, hardnesses and finishes.

3.1.1 Style 1. The handpiece components materials, hardnesses, and finishes for style 1 shall be as specified in Table I.

Table I. Materials, Hardnesses and Finishes for Style 1

Part	Component	Material	Rockwell Hardness	Finish
Head assembly	Head shell	Brass	-------	Chrome over nickel
	Cap	Stainless Steel, Type 440F	C53-58 1/	- - - -
	Bur tube	Stainless Steel, Type 440F	C53-58 1/	- - - -
	Bearing	Stainless Steel, Type 440F	C53-58 1/	- - - -
	Latch	Stainless Steel, Type 420	C56-60 1/	- - - -
	Latch screw	Stainless Steel, Type 416	-------	- - - -
	Gear	Stainless Steel, Type 440F	C53-58 1/	- - - -
	Slinger	Stainless Steel, Type 416	-------	- - - -
Neck assembly	Neck bearing	Stainless Steel, Type 416	C35-40 1/	- - - -
	Gear	Stainless Steel, Type 440F	C55-60 1/	- - - -
	Shaft	Stainless Steel, Type 440F	C53-58 1/	- - - -
	Rivet	Brass, half hard	--------	- - - -
Knee assembly	Knee	Brass	--------	Chrome over nickel
Round sleeve back end assembly	Round sleeve back end	Phosphorus bronze	-------	Chrome over nickel
	Shaft	Stainless Steel, Type 440F	C55-60	- - - -
	Ring	Stainless Steel, Type 440F	C55-60	- - - -
	Gear	Stainless Steel, Type 440F	C55-60	- - - -
	Rivet	Brass, half hard	--------	- - - -
Wrench	Octagon cap wrench	Steel, AISI B1113	2/	Dull mechanical nickel

1/ Hardness after proper tempering.
2/ Hardened and tempered.

Note: Bur tube and gear assembly hard chrome plated 0.0004 to 0.0006 inch before grinding.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

3.1.2 Style 2. The handpiece components materials, hardnesses, and finishes for style 2 shall be as specified in Table II.

Table II. Materials, Hardnesses and Finishes for Style 2

Part	Component	Material	Rockwell Hardness	Finish
Head assembly	Head shell	Brass, ASTM B-16	------	Chrome
	Top bearing	Phosphorus bronze, #544	------	------
	Latch	Spring steel, AISI 1070	------	Electrolize 3/
	Latch screw	Steel, AISI C1112-13 or 12L14	------	Black oxide dip
Spindle	Spindle	Steel, AISI 01 drill rod	C 40-50	Electrolize 3/
	Ring gear	Steel, AISI 01 drill rod	C 40-50	Electrolize 3/
	Lower bearing	Phosphorus bronze, #544	-------	-------
	Slinger	Stainless Steel, type 303 or 416	-------	-------
Head to elbow bearing assembly	Bearing	Steel, AISI 12L14	Case hardened 2/	4/
	Pinion gear	Steel, AISI 01 drill rod	C 40-50	Electrolize 3/
	Shaft	Steel, AISI W1 or 01 drill rod	-------	4/
	Elbow gear	Steel, AISI 01 drill rod	C 40-50	Electrolize 3/
Elbow and rear end assembly	Elbow	Brass, ASTM B-16	-------	Chrome
	Elbow gear	Steel, AISI 01 drill rod	C 40-50	Electrolize 3/
	Rear end housing	Brass, ASTM B-16	-------	Chrome
	Rear end shaft	Steel, AISI W1 or 01 drill rod	-------	4/
Head locknut	Head Locknut	Steel, AISI 12L14	Case hardened 2/	Chrome
Wrench	Wrench	Steel, AISI, 1010, 1/2 hard	Case hardened 2/	Nickel

1/ All hardnesses specified apply prior to plating.
2/ Minimum depth 0.003 inch.
3/ Minimum thickness 0.0001 inch.
4/ Material, not stainless steel, shall be coated or plated to render the part corrosion-resistant.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

3.2 Style, design and dimensions. The handpiece shall be of the type known to the dental trade as a conventional latch type, round sleeve, contra-angle handpiece. The handpiece shall be entirely suitable for attachment and use with belt or air driven, round nose straight handpieces (see 6.3). The contra-angle handpiece shall be suitable for use in the performance of operative dentistry. The handpiece shall consist essentially of a head assembly, neck assembly, knee (elbow) assembly and round sleeve back end assembly. The handpiece shall be a durable, precision made handpiece entirely suitable for operation at handpiece speeds of up to at least 12,000 revolutions per minute (rpm). The design and dimensions of the handpiece shall be in strict accordance with Defense Personnel Support Center (DPSC) Drawing No. 24252 for style 1 or 2, as specified (see 6.1). The requirements specified herein for one style shall apply to that style only, and shall not be interchangeable with the other style. Style 1 handpiece shall be a non-adjustable head. Style 2 handpiece shall have an adjustable head to provide for adjustment of the angulation of the head with respect to the back end assembly.

3.2.1 Head assembly. Head assembly shall have one fixed bearing and one screwed in bearing with a minimum total bearing surface of 0.152 square inch. The maximum clearance between the fixed bearing and the bur socket shall not exceed 0.001 inch throughout the entire length of the bearing. The fixed bearing shall be attached to the outer shell of the head assembly in such a manner that the required tolerance is maintained. A grease chamber shall surround the fixed bearing. The gear mechanism between the bur socket and the drive shaft shall have gears that have a line contact. One gear shall be fixed on the bur socket, and the other gear shall be fixed on the drive shaft.

3.2.1.1 Bur holding. Bur socket for retaining bur shall receive the conventional design angle bur shank 0.0919 to 0.0929 inch in diameter. A latch of suitable design shall be provided for locking bur in place. To insure interchangeability with standard angle burs, a standard test gauge (conforming to dimensions shown in DPSC Drawing 24252) must assemble into socket and latch for retaining bur, and must close without binding on gauge or interfering with free rotation of the test gauge by the socket. Total end play of gauge in socket with reference to head of the angle shall not exceed 0.0095 inch. Side play of gauge measured at 3/8 inch from nosepiece shall not exceed 0.010 inch. An inspection plug 0.0936 inch in diameter shall not enter hole in socket. The shank of test gauge shall clear nosepiece by 3/8 inch plus or minus 1/64 inch. End play of bur tube shall range between 0.0001 and 0.0025 inch when tested as specified in 4.4.3.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

3.2.2 Neck and knee (elbow) assemblies. The neck assembly shall be so designed that it can be removed intact with the bearing and drive shaft gear so that the head assembly can be properly cleaned and lubricated. The knee (elbow) assembly shall be provided with the conventional universal joint gear mechanism.

3.2.3 Round sleeve back end assembly. The round sleeve back end assembly shall consist of a sheath and a drive shaft. The sheath shall be knurled on the outside to provide a gripping surface, and the interior surface portion of the round sleeve shall fit the round nose portion of the nosepiece on the straight handpiece (see 3.2).

3.3 Runout. The total runout of the angle handpiece when measured with the arbor shall not exceed 0.005 inch at 3/8 inch outside the bur tube point when tested as specified in 4.4.4.

3.4 Vibration and heat rise. The handpieces shall pass the test for vibration and heat rise specified in 4.4.5.

3.5 Drive train smoothness and backlash. The handpiece shall pass the test for drive train smoothness and backlash when tested as specified in 4.4.6.

3.6 Finish. The external surfaces of the handpiece shall be smooth. The finish shall be bright or satin finish except for the knurled portion.

3.7 Autoclaving and disinfecting. Instructions shall be provided for sterilization and disinfecting, as applicable.

3.8 Lubrication. Instructions shall be provided for proper lubrication of the handpiece components and surfaces requiring lubrication for proper operation and maintenance.

3.9 Tools. Applicable tools such as wrenches shall be supplied for self repairable components and for routine maintenance.

3.10 Identification marking. Each handpiece shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

3.11 Service Data. The contractor shall furnish with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance, and lists of component and replacement parts. The manual shall comply with the requirements of MIL-L-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

3.12 Workmanship. Workmanship shall be first class throughout. Handpieces shall be free from defects which detract from their appearance or which may impair their serviceability.

3.13 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper design, function, fit, assembly and operation of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the contracting officer.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials used in the handpieces, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified in Tables III, V and VI (see 4.2.1, 4.3. and 4.3.1) and Table IV (see 4.2.2).

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

4.2.1 Sampling for examination. Sampling for examination for the handpiece shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in Table III. The unit of product for examination purposes shall be one handpiece.

Table III. Inspection and Acceptable Quality Levels

Type defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor A	II	2.5
Minor B	S-1	4.0

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table IV for the handpiece. The acceptance number shall be zero for all sample sizes. The unit of product for test purposes shall be one handpiece. Where feasible, the same sample unit may be used for the determination of two or more test characteristics.

Table IV. Sampling criteria for Testing

Component	Characteristic	Require-ment	Test procedure	Inspection level
All components in Tables I and II	Chemical Analysis	3.1	4.4.1	*
Components indicated in Tables I and II	Hardness	3.1	4.4.2	S-1
Assembled handpiece	End play	3.2.1.1	4.4.3	S-1
Assembled handpiece	Runout	3.3	4.4.4	S-1
Assembled handpiece	Vibration and heat rise	3.4	4.4.5	S-1
Assembled handpiece	Drive train smoothness and backlash	3.5	4.4.6	100%

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:14Z
Check the source to verify that this is the currentversion before use.

* A sufficient amount from each lot of raw materials used in the construction of the item shall be selected for chemical analysis when certificates of quality are not available as specified in 4.1.3.

4.3 Examination. The handpiece shall be examined to determine compliance with requirements contained in this specification. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in Table V.

Table V. Classification of Defects

Categories	Defects
Critical	None defined.
Major A	
101	Design and construction of style 1 and 2 not as specified.
102	Round nose straight handpiece will not accept and lock-in handpiece (see 3.2).
103	Latch will not lock bur in place (see 3.2.1.1).
104	Slivers and burs present.
105	Shaft does not turn freely and smoothly.
Major B	
151	Dimensions affecting fit of handpiece to round nose straight handpiece not as specified.
152	Dimensions affecting fit of bur shank to bur socket not as specified.
Minor A	
201	Service data, technical manuals missing.
202	Identification marking not present, correct, complete, or legible.
203	Pits, knicks and discoloration present on outer surfaces.
204	Plating not free of defects.
205	Knurling not as specified.
Minor B	
251	Dimensions (other than 151 and 152) not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

4.3.1 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table VI. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table VI. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper location, sequence, or method of application.

4.4 Tests.

4.4.1 Chemical analysis. Metal analysis shall be conducted in accordance with method 111 or 112 of FED-STD-151.

4.4.2 Hardness test. Hardness shall be determined in accordance with ASTM Standard E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

4.4.3 End play test. End play of socket shall be determined with a dial indicator, sensitive to 0.0002 inch. The socket shall be pushed toward and away from the head of the handpiece, and the end play of the socket alone shall be measured.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

4.4.4 Runout. The runout is measured by inserting a true-running arbor .0929 inch in diameter into the bur tube. At the end of the arbor an indicator is placed measuring in increments of .001 inch. Arbor shall be rotated through 360 degrees with maximum and minimum readings recorded. The difference between these readings is the total runout.

4.4.5 Vibration and heat rise. The assembled and lubricated handpiece shall be tested by running the handpiece, containing a dental bur, for three minutes at 10,000 rpm under free load. The handpiece shall be tested for excessive vibration when hand held, and for excessive heat rise (above 115° Fahrenheit (F)) when tested with a pyrometer held against the head housing and against the back end housing.

4.4.6 Drive train smoothness and backlash test. The assembled and lubricated handpiece shall be tested by spinning the back end drive shaft through at least two full revolutions, both clockwise and counterclockwise, using finger pressure only as applied by the thumb and index finger. Any detectable binding, slipping, backlash, or irregular mechanical interference shall cause the handpiece to be rejected.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. Each handpiece shall be thoroughly cleaned and dried in accordance with the applicable method specified in MIL-P-116. Each handpiece shall be wrapped in neutral tissue paper or other suitable anticorrosive material and packaged in accordance with MIL-P-116, method IC-2 except that a preservative shall not be used.

5.1.1.2 Intermediate package. Twelve units shall be packaged in a snug fitting box constructed in accordance with PPP-B-566, PPP-B-676 or PPP-B-636, class domestic. Closure shall be adequate to prevent accidental opening.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraph of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Twenty-four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601 domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in the appendix of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or others carriers, at the lowest rate, to point of delivery called in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:14Z
Check the source to verify that this is the currentversion before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Style required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following item appearing in the Federal Supply Catalog:

NSN	Item Identification	Specification Designation
6520-00-935-7142	HANDPIECE, DENTAL, Contra-angle, Round Sleeve: Latch type.	Style 1 or 2

6.3 Intended use. The contra-angle handpiece covered by this specification is intended for use with Teledyne Emesco's "No. 4N" Handpiece, S. S. White Company's "No. 14" Handpiece, Midwest American Company's "Tru-Torc" Handpiece and Star Dental Manufacturing Company's "Titan-2" Handpiece with "T-10A" Straight Nose Cone.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6520-1395
is

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-H-36734A HANDPIECE, DENTAL, CONTRA-ANGLE, ROUND

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.