MIL-O-36788
Amendment-3
7 December 1973
SUPERSEDING
Amendment-2
11 May 1973 and
Amendment-1
13 February 1973

MILITARY SPECIFICATION

OXYGEN, USP, WITH TURE AND FACE MASK

24 GAL. (90 liters)

This amendment forms a part of Military Specification MIL-O-36788, dated 23 August 1971, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2

3.1.1 Oxygen. Line 2, delete "of the current edition".

* 3.1.3 Valve. Line 2, delete "nickel or chrome plated,".

Page 3

3.2 Style, design and dimensions. Line 4, delete "an interstate Commerce Commission and remainder of paragraph." and substitute "Department of Transportation Regulations."

3.2.1 Cylinder. Delete first sentence and substitute. "The cylinder for Type I shall be not less than 1-15/16 inches in diameter nor more than 18 inches in overall length and for Type II shall not be more than 4 inches in diameter nor more than 18 inches in overall length, and shall have a minimum wall thickness of 0.045 inch and a minimum volume of 93 cubic inches (1570 cubic centimeters)." Line 8, delete "Interstate Commerce Commission" and substitute "Department of Transportation Regulations". Line 13, delete "4.4.3" and substitute "4.4.2" and add "The unit shall be capable of being stored for a period of fourteen (14) days at 150°F without exhibiting any loss of oxygen content on the contents indicator."

3.2.2 Regulator. Delete entirely and substitute "The regulator shall be so constructed that it can be attached to the unit without the use of tools and shall permit the flow of gas to be turned off without disconnecting the regulator from the unit. The operation of the regulator shall be controlled by rotating the regulator in a clockwise direction to the maximum extent which will permit oxygen to flow at a fixed rate of 6 liters per minute minimum across the entire range of tank pressure.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:19Z
Check the source to verify that this is the current version before use.

3.2.3 Oxygen. Line 1 delete "90 to 100 liters" and substitute "a minimum of 90 usable liters." Line 2, delete "0°" and substitute "21°C". Line 4: Delete "may" and substitute "shall".

Page 4:

3.2.4 Mask, reservoir bag, and tubing. At end of paragraph add· "Tubing shall be readily attached to mask and regulator."

3.4 Performance characteristics. Delete entirely and substitute· "For Type I the unit shall deliver oxygen at a constant rate of 6 liters per minute minimum at 1000 pounds nominal at 70° F for a minimum 15 minute duration. For Type II the unit shall deliver oxygen at a constant rate of 6 liters per minute minimum at 1800 pounds nominal at 70° F for a minimum 15 minute duration.

3.5.1 Tubing. Line 2, delete "4.4.2.1" and substitute "4.4.3.1".

3.5.2 Mask and reservoir bag. Line 1, delete "4.4.2.2" and substitute "4.4.3.2".

3.6 Certificates of safety (nontoxicity). Line 4, delete "3.6 and 4.4.2" and substitute "3.5 and 4.4.3".

3.7 Manufacturer's record. Lines 3 and 4 delete "of all tests made on nontoxic masks, reservoir bags, and tubes," and substitute· "so that it is possible to determine all the facts as to source, material and as to the manufacture of such lot."

3.8 Finish. Line 1, after "enamel" insert "or lacquer".

Page 5

3.11 Workmanship. Line 1, delete "workmamship" and substitute "workmanship".

4.1.2 Inspection. Line 2, delete "of" and substitute "or".

4.1.3 Certificates of quality. Line 2 after brass delete "or" insert comma and after "aluminum" insert "or fuse metal,".

Page 6·

* 4.2.2.2 For other tests. Delete third sentence. In Table III add the following as indicated below:

Component	Characteristic	Requirement	Test procedure	Insp. level	AQL (percent defective)
"Cylinder	Performance	3.4	---	S-1	* "

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:19Z
Check the source to verify that this is the current version before use.

4.3 Examination. At end of third sentence delete period (.) and add: "and shall be removed from the offered lot."

Page 7:

Table IV. Delete "Major 105 entirely".

4.4.3.1 Delete Tubes in title and line 1 and substitute "Tubing".

4.4.4 Oxygen. Second sentence delete "method" and substitute "methods".

Page 8:

5.1.3.1 Unit package. Line 4, following "reservoir bag" insert. "and regulator". Line 6, delete the word "also". Line 7, following "plastic", delete remainder of sentence and substitute "sleeve".

Page 9

5.2.1.1 Exterior container and 5.2.2.1 Exterior container. At the end of each paragraph add "Fiberboard boxes shall conform to the special requirements as specified in the applicable table referenced in PPP-B-636."

5.3.2 Exterior container. At end of paragraph add "Marking shall include the legend:

'DO NOT EXPOSE TO TEMPERATURES ABOVE 130°F.'"

Add the following new paragraph

"6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity
Defense Supply Agency - DM

Project No. 6505-1304

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:19Z
Check the source to verify that this is the current version before use.