MIL-S-36750
22 April 1971

MILITARY SPECIFICATION

SCISSORS-DAGGER, TRACHEOTOMY, UNIVERSITY OF KANSAS DESIGN, CURVED BLADES, 5-3/16 INCH LONG

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a dagger-scissors, suitable for use in cricothyroidotomies concerning infants and adults.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Specifications, Dental and Surgical, General Specifications.
MMM-A-250	Adhesive, Water-Resistant, For Sealing Fiber
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Military

MIL-P-116	Preservations, Methods of.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

23172	SCISSORS-DAGGER, TRACHEOTOMY, University of Kansas Design, Curved Blades, Angulated Pointed Tips, 5-3/16 inches Long.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General Requirements. The dagger scissors shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Material. The dagger-scissors shall be fabricated from Class 4, Type 420 or 420F material.

3.3 Hardness. After proper stress relieving, the dagger-scissors shall have a Rockwell hardness of C50 to C55.

3.4 Style, design and dimensions. Shall be a dagger-scissors, University of Kansas design, suitable for use in cricothyroidotomies concerning pediatric and adult surgical procedures. Shall be designed to cut a nick in the skin and to incise the cartilage of the trachea. Shall have specially curved cutting blades with angulated pointed tips. Tips shall be suitable for piercing into the trachea without causing enlarged openings, particular with reference to pediatric patients. Tips shall possess a special bevel on both blades, with the last third of the bevel forming a non-sharp edge blending into its sharp portion. Cutting action shall be such that skin is grasped and cut cleanly without causing the skin to be pushed away from the cutting blades. The dagger-scissors shall be in strict accordance with DPSC Drawing 23172.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

3.5 Finish. Finish shall be mirror or mirror and satin.

3.5.1 Passivation. The entire item shall be properly passivated.

3.6 Cutting edges. Scissors shall cut a continuous length of at least 1-1/16 inches of the test material, when tested as specified in 4.4.1.

3.7 Tip protectors. Tips of cutting blades shall possess a protective covering in each unit package.

3.8 Identification marking. In addition to GG-I-526 requirements, instruments of foreign manufacture shall also be marked "stainless".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Shall be in accordance with GG-I-526.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resistant steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. The unit of product for examination purposes shall be one dagger-scissors.

Table I. Sampling Criteria for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one dagger-scissors.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQl
Dagger-scissors	Composition analysis	3.2 and GG-I-526	GG-I-526	*	-
	Hardness	3.3 and GG-I-526	GG-I-526	S-2	1.0
	Dimensions	3.4	- - - -	S-2	1.5
	Copper sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Boil test	GG-I-526	GG-I-526	S-2	1.0
	Scissors performance test	3.6 and GG-I-526	4.4.1 and GG-I-526	II	1.0

* Four ounces of each lot of metal used in the construction of the dagger-scissors shall be selected for chemical analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

4.3 Examination. The dagger-scissors shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in Table I and III. Inspection, however, shall not be restricted to defects listed in Table III. Any dagger-scissors in the sample containing one or more defects shall be rejected and if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Tips of blades not sharp at distal end.
102	Bevel of blades sharp throughout its length.
103	Dagger-scissors not free from cracks, pits, dents, nicks, or burrs.
104	Finish not uniform.
105	Scissors not free from excessive play or excessive stiffness during cutting action.
106	Scissors not of the proper design specified.
Minor	
201	Identification marking not present, correct, complete or legible.
202	Tips of cutting blades not protected in unit package.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Scissors-performance test. The test material shall be rubber dam meeting the requirements of the Defense Medical Purchase Description for Federal Stock No. 6520-754-2769 Rubber Dam, Medium, 6 inches square. Three separate, consecutive tests shall be performed by each dagger-scissors as described and specified in GG-I-526.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One scissors, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	12 units	144 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be over-packed in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordnace with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Not required.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive material.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method IA-8.

5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flaps with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard boxes shall be closed by method II of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate pac-age shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The dagger-scissors are to be used in cricothyroidotomies in conjunction with FSNs 6515-935-7098 and 6515-935-7090 Cannula, Tracheotomy, University of Kansas Design, Metal, Pediatric and Adult sizes respectively; and FSN 6515-935-7091 Dilator-Dagger, Tracheotomy, University of Kansas Design, Curved Blades, 4-1/4 inch Long, Forceps Type, with box lock.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Applicable Federal Stock Number (FSN).
(c) Selection of applicable level of packaging and packing required (see 5.1 and 5.2).

6.3 This specification covers FSN 6515-935-7092 SCISSORS-DAGGER, TRACHEOTOMY, University of Kansas Design, Curved Blades, Angulated Pointed Tips, 5-3/16 inch Long, listed in the Federal Supply Catalog. It does not include all types as indicated by the title nor all types commercially available.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1562

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:15Z
Check the source to verify that this is the current version before use.