MIL-S-36771A
19 August 1975
SUPERSEDING
MIL-S-36771
9 July 1971

MILITARY SPECIFICATION

SNARE, TONSIL, TYDING

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specificaion covers a tonsil snare of the ydding design, suitable for use in tonsillectomy procedures.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of nvitation for bids or request for proposals, form a part of this ecification to the extent specified herein:

ECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Shipping, Fiberboard.

Military

MIL-P-116	Preservation, Methods of.

STANDARDS

Federal

FED-STD-123	Marking for Domestic Shipment (Civilian Agencies).
FED-STD-151	Metals, Test Methods.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

Drawing No. 1905 Snare, Tonsil, Tyding.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Components of the tonsil snare shall be fabricated from class 3, type 302, 303 or 304 material, or class 4, type 420, or type 420F material, all in accordance with GG-I-526. All material shall be passivated.

3.2 Hardness. Rockwell hardness for class 3 material shall be B75 minimum. Rockwell hardness for class 4 material shall be C40 to C45, after proper stress relieving. Hardness shall be as stated when tested as specified in 4.4.1.

3.3 Style, design and dimensions. The tonsil snare shall be fabricated from materials specified in 3.1. The tonsil snare shall be of the Tyding design suitable for the surgical removal of tonsils. The snare shall consist of a movable member, a stationary member, a stylet, a cannula and a pivot screw. The distal end of the cannula shall possess a pronounced internal chamfer to allow easy action of the snare when a formed tonsil snare wire is assembled to the stylet. The tonsil snare shall be in accordance with Defense Personnel Support Center Drawing No. 1905.

3.4 Finish. Finish of all the snare components shall be bright. Finish on all surfaces and edges shall be uniform, free from burrs, sharp edges, crevices, grind marks, rough areas, cracks, scale and corrosion.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

3.5 Operation requirements. When tested as specified in 4.4.2, the stylet shall not stick or bind in the cannula lumen. In addition, when a formed tonsil snare wire having a diameter of 0.022 inch (see 6.3) is assembled to the stylet, and the movable member set in the closed position (i.e., when the moveable member is retracted against the stationary member), the stylet shall not bind in the cannula lumen.

3.6 Identification marking. Tonsil snares shall be permanently marked with the contractor's name or registered trademark and the term "stainless". In addition to the above, snares manufactured in countries other than the United States shall bear the country of origin.

3.7 Workmanship. Workmanship shall be first class throughout. All components of the tonsil snare shall be free from defects which detract from their appearance or impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

.3 Certificates of quality. Certificates of quality, supplied
manufacturer of the corrosion-resistant steel alloys, may
nished in lieu of actual performance of such testing by the
ctor, provided lot identity has been maintained and can be
trated to the Government. The certificate shall include the
f the contractor, the contract number, the name of the
cturer or supplier, the National Stock Number (NSN), the
dentification, the name of the component/material, the lot
, the lot size, the sample size, the date of testing, the
ethod, individual test results, and the specification
ements.

Sampling.

.1 For examination. Sampling for examination shall be
ted in accordance with MIL-STD-105 and as indicated in
I. Unit of product for examination purposes shall be one
snare.

Table I. Sampling Criteria for Examination

efect	Inspection level	AQL (percent defective)
or	II	1.0
or	II	2.5

.2 For tests. Sampling for tests shall be conducted in
ance with MIL-STD-105, and as indicated in Table II. Unit
duct for test purposes shall be one tonsil snare. Where
le, the same sample unit shall be used for the determination
or more test characteristics.

Table II. Sampling Criteria for Tests

ent	Characteristic	Require-ment	Test procedure	Insp level	AQL
	Hardness (for all material)	3.2	4.4.1	S-2	1.5
	Dimensions	3.3	-----	S-2	4.0
	Boil test (for all material)	GG-I-526	GG-I-526	S-2	1.0
	Copper Sulfate test (for class 4 material)	GG-I-526	GG-I-526	S-2	1.0
	Operation test	3.5	4.4.2	S-2	1.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in construction of the snare shall be selected for chemical analysis.

4.3 Examination.

4.3.1 Examination of tonsil snares. The tonsil snares shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any snares in the sample containing one or more defects shall be rejected and if the number of defective snares in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Snare not of the design specified.
102	One or more components missing.
103	Cannula not readily assembled to, or removed from, the stationary member.
104	External threads of cannula and pivot screw not complete or well defined.
105	Internal threads of stationary member not complete or well defined.
106	Snare components not of finish specified.
107	Snare components not free from sharp edges, crevices, cracks, burrs, grind marks, rough areas, scale, and corrosion.
108	Pivot screw not easily assembled to, or removed from, the stationary member.
Minor	
201	Snare not permanently and legibly identified as specified in 3.6.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

2 Examination for preparation for delivery. An examination
e made to determine whether the packaging, packing, and
; comply with the requirements of Section 5. Defects shall
ed as specified in Table IV. Sampling shall be in accordance
IL-STD-105. The sample unit shall be one container fully
d for delivery. The lot shall be the number of containers
for inspection at one time. The inspection level shall
and the acceptable quality level shall be 4.0 defects per
units.

Table IV. Examination for Preparation for Delivery

ation	Defect
ers	Not as specified.
y	Quantity in interior packages and shipping containers not as specified.
ing and packing tals	Any component missing or damaged.
ship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
gs	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

Tests. Tests shall be conducted to determine compliance with
ication requirements.

.1 Hardness test. Rockwell hardness of classes 3 and 4
al shall be determined in accordance with method 243 of
l Test Method Standard No. 151.

.2 Operation test. The operation of the tonsil snare shall be
. Place the moveable member into the slot of the stylet and
the moveable member to the stationary member with the pivot
The moveable member shall be retracted back and forth through
ll travel. Evidence of sticking or binding action shall be

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

4.4.2 Operation test (cont'd)

cause for rejection immediately. The cannula shall be secured to the barrel of the stationary member. The moveable member shall then be placed as far from the stationary member as its travel will permit and a formed snare wire (see 6.3) shall be assembled to the stylet to form an appropriate loop. The moveable member shall then be retracted back and forth a total of 10 times so that the wire assembled to the stylet shall be allowed to travel through the cannula lumen, with the stylet going through its full travel.

5. PREPARATION FOR DELIVERY

5.1 Preservation and packaging. Preservation and packaging shall be level A or C as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning, drying and preservative. Each snare shall be thoroughly cleaned and dried in accordance with applicable method as outlined in MIL-P-116. Preservative shall not be applied.

5.1.1.2 Unit package. Each snare shall be wrapped in neutral tissue or other suitable anticorrosive material. Each snare shall be further packaged in accordance with MIL-P-116, method IC-2.

5.1.2 Level C. Snares shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Forty eight (48) snares shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B. Forty eight (48) snares shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3, PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. Snares shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit package and exterior container. Each unit package and exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.

(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

(c) National Stock Number.

6.2 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

National Stock No.	Item Identification
6515-01-149-1103	SNARE, TONSIL, Tyding

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

6.3 The formed snare wire specified in 4.4.2 shall meet the material and dimensional requirements specified for National Stock Number 6515-00-149-1102.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Proejct No. 6515-1962
is

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-S-36771A SNARE, TONSIL, TYDING

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A GIVE PARAGRAPH NUMBER AND WORDING.

   B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:17Z
Check the source to verify that this is the current version before use.