MILITARY SPECIFICATION

TRACTION DEVICE

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1 SCOPE

1.1 Scope. This specification covers three types of spring-loaded traction devices

1.2 Classification.

Type I	3 pounds
Type II	5 pounds
Type III	10 pounds.

2 APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

Federal

QQ-A-225/6	Aluminum Alloy, Bar, Rod, Wire and Special Shapes, Rolled, Drawn or Cold Finished, 2024.
PPP-B-566	Boxes, Folding, Paperboard
PPP-B-585	Boxes, Wood, Wirebound
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard
PPP-B-676	Box, Setup.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing)

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-66	Steel Chemical Composition and Hardenability

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage

DRAWING

Defense Personnel Support Center (DPSC)

25045	Spring Traction Device, Exerciser, 3, 5 & 10 Lb Force

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Shall be as specified in the Bill of Materials and as listed below.

3.1.1 Corrosion-resisting steel. Corrosion-resisting steel shall conform to AISI 302 of FED-STD-66.

3.1.2 Aluminum bars and rod. Aluminum bars and rod shall conform to QQ-A-225/6 for alloy 2024, temper T3.

3.2 Construction. Each traction device shall consist of a constant force spring housed in a plastic case to which is attached a wire cable with an eye bolt for attaching to the individual to be put in traction. The traction spring tension shall be within 1/4 pound of its rating when tested as specified in 4.4.1. The spring shall have an active excursion length of 57 inches, minimum. The traction devices shall be in accordance with DPSC Drawing 25045 and Bill of Material thereto. There shall be no loosening of the cable within the swaged sleeve shown on DPSC Drawing 25045 when tested as specified in 4.4.2.

3.2.1 Type I. The Type I traction device shall utilize a 3 pound constant force spring.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

3.2.2 Type II The Type II traction device shall utilize a 5 pound constant force spring.

3.2.3 Type III The Type III traction device shall utilize a 10 pound constant force spring.

3.3 Identification marking. Each plastic case shall be impressed on one side with the name or registered trademark of the manufacturer, the contract number, the Federal Stock Number, item identification and the letters U.S The other half of the case shall bear markings appropriate to the spring it encloses, in accordance with DPSC Drawing 25045

3.4 Workmanship Workmanship shall be first class throughout. Traction devices shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate

4.1.2 Inspection Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality for the materials (3.1), may be supplied by the manufacturer of the traction devices in lieu of actual performance of material testing, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the Item Identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table I.

Table I

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	I	1.0
Minor defects	I	2.5
For dimensional examination	S-2	2.5

4.2.2 For tests All units shall be tested as specified herein. The AQL shall be 1.

4.3 Examination The units shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects in Table II

Table II

Categories	Defects*
Major	
101	Traction device not complete.
102	Parts loose, broken or damaged.
103	Constant force spring rating does not match case marking
104	Traction spring active excursion less than 57 inches
Minor	
201	Case or clip not free of sharp edges.
202	Traction device not clean.
203	Identification marking missing, incomplete or illegible.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The traction device shall be examined for defects in dimension. Any dimension not within the tolerance specified on the drawing shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

4.4.1 Spring force. The traction devices shall be tested to determine the force exerted by their springs. A calibrated spring scale, attached to the end of the traction cable, shall be used for measuring the force. The force measured from the end of the cable when it is extended 18 inches from the case shall be recorded. The test shall be repeated for extensions of 36 inches and 54 inches.

4 4.2 Cable sleeve strength. The cable shall be attached to the swaged sleeve as shown on DPSC Drawing 25045. With the sleeve clamped in place a 50 ± 1 pound pull shall be applied through the cable.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1 1 Unit of issue One traction device, as specified, constitutes one unit of issue.

5 1.2 Packaging quantities The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	10 units

5.1.2.1 Packing variation permitted If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3 1 Unit package Unit package shall be a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling

5.1 4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its itended purpose.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3, PPP-B-601 domestic type, PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129 When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container Exterior container shall be marked in accordance with MIL-STD-129.

6 NOTES

6.1 Intended use. The traction devices shall be suitable for traction and therapeutic exercise programs in military medical facilities.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

6.2 Ordering data Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packing (see 5.1 and 5.2).

6.3 This specification covers the following items appearing in the Federal Supply Catalog

Federal Stock Number	Item
6530-403-3116	TRACTION DEVICE, 3 lb.
6530-168-6815	TRACTION DEVICE, 5 lb.
6530-403-3117	TRACTION DEVICE, 10 lb.

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements

Custodians.
Army - MD
Navy - MS
Air Force - 03

Preparing activity
Defense Supply Agency - DM

Project No 6530-1198

Review activities.
Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

INSTRUCTIONS: This sheet is to be filled out by personnel using documents cited in the solicitation in the use of the specification in procurement. It provides a formal means for submitting comments to the Department of Defense. This sheet is provided for the purpose of the use of this specification and will insure that suitable [illegible] be given the [illegible] a minimum [illegible] least cost [illegible] comments and recommendations submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

SPECIFICATION: MIL-T-36736 PSN:

ORGANIZATION:

CITY AND STATE: CONTRACT NUMBER:

MATERIAL PROCURED UNDER A: ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE

A. GIVE PARAGRAPH NUMBER & WORDING

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE

☐ YES ☐ NO (If "yes" in what way?)

4. REMARKS (Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)

SUBMITTED BY (Printed or typed name and activity - Optional): DATE:

DD FORM 1 JAN 86 1426 REPLACES EDITION OF 1 OCT 84 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS OFFERORS

DO NOT CONDITION OR BASE YOUR BID OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA PA 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS DEFENSE PERSONNEL SUPPORT CENTER
ATTN DIRECTORATE OF MEDICAL MATERIEL CODE DMT
2800 SOUTH 20TH STREET
PHILADELPHIA PA 19101

Fold

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:14Z
Check the source to verify that this is the current version before use.