MIL-T-36754A
12 January 1977
SUPERSEDING
MIL-T-36754
16 April 1971

MILITARY SPECIFICATION

TRAY, INSERT, DENTAL INSTRUMENT AND SUPPLIES CABINET, PLASTIC

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers plastic trays used as inserts in dental cabinets for containing various dental instruments and supplies.

1.2 Classification. Insert trays shall be of the following styles as specified (see 6.1):

Style 1 - Tray No. 16A.
Style 2 - Tray No. 17.
Style 3 - Tray No. 18.
Style 4 - Tray No. 19.
Style 5 - Tray No. 20.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Boxes, Shipping, Fiberboard.
PPP-B-676 - Boxes, Setup.

MILITARY

MIL-P-116 - Preservation-packaging, Method of.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24201 - Tray, Insert, Dental Instrument and Supplies Cabinet, Plastic.

(Copies of specifications, standards, drawings and publications required by the contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE
FOOD AND DRUG ADMINISTRATION (FDA)

FDA Regulation

121.2501 - Food Additives Resulting from Contact with Containers or Equipment and Food Additives Otherwise Affecting Foods, Olefin Polymers.

Federal Food, Drug, and Cosmetic Act and applicable regulations thereto.

(Application for copies should be addressed to the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.)

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:15Z
Check the source to verify that this is the current version before use.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM D256	- Tests for Impact Resistance of Plastics and Electrical Insulating Materials.
ASTM D638	- Test for Tensile Properties of Plastics.
ASTM D648	- Test for Deflection Temperature of Plastics Under Flexural Load.
ASTM D785	- Test for Rockwell Hardness of Plastics and Electrical Insulating Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

3. REQUIREMENTS

3.1 Materials. Insert trays shall be white and shall be fabricated of polypropylene or copolymer polypropylene conforming to the physical properties specified in table I. The material shall be odor-free, and shall conform to the specifications and requirements of Food and Drug Administration (FDA) Regulation 121.2501.

TABLE I. Physical properties of raw material.

Physical property	Unit	Value Required	Test Method
Tensile yield strength, minimum	p.s.i. at 0.2 in./min.	3000*	4.3.1
Izod impact strength, minimum	ft. lb./in-notch at 73°F.	0.4*	4.3.2
Deflection temperature, minimum	°F at 66 p.s.i. fiber stress	190**	4.3.3
Rockwell hardness, minimum	"R" Scale	65*	4.3.4

*A minimum of 5 determinations shall be made (5 specimens). Failure of the average of these determinations to comply with the minimum value required shall be cause for rejection.

**Two determinations shall be made (2 specimens). Failure of either of these determinations to comply with the minimum value required shall be cause for rejection.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:15Z
Check the source to verify that this is the currentversion before use.

3.2 Style, design and dimensions. Insert trays shall be of the styles, as specified in table II, and shall be suitable as inserts in dental cabinets to contain the various instruments and supplies, as specified for the applicable style.

TABLE II. Styles.

Style	Tray number	Use
1	16A	For long-handle instruments.
2	17	For clamps and matrix bands.
3	18	For polishing strips.
4	19	For scissors and pliers.
5	20	For disks, wheels and rotary stones.

Insert trays shall conform to the requirements specified herein, and shall be in accordance with Defense Personnel Support Center Drawing 24201 for the style, design and dimensions for the applicable style number specified.

3.3 Identification marking. Each tray shall be legibly and permanently marked with the name or registered trademark of the manufacturer or contractor and with the tray number, as specified in paragraph 3.2 (table II) for the applicable style.

3.4 Finish. Insert trays shall have a smooth and glossy finish on all surfaces.

3.5 Workmanship. Workmanship shall be first class throughout. Trays shall be free from defects which detract from their appearance or may impair their serviceability. Inward bowing of the sides and ends of insert trays shall be limited to 1/16 inch.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order,

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the polypropylene material, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the requirements of this specification.

4.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties). Sampling for inspection shall be performed in accordance with MIL-STD-105, except where otherwise indicated hereinafter.

4.2.1 Examination. Insert trays shall be examined for defects including, but not limited to those listed in table III.

TABLE III. Classification of defects.

Categories	Defects
Critical	None Defined.
Major A	
101	Style and design not as specified for style of tray.
102	Tray not marked with applicable tray number.
103	Cracks and chips present in tray.
104	Objectionable odor present in tray.
105	Flash present on tray surfaces.
106	Finish not smooth and glossy.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects.(Cont'd)

Categories	Defects
Major B	
151	Overall dimensions of tray not as specified.
152	Dimensions affecting fit of hold-down strip (when provided) not as specified.
153	Inward bowing of tray sides or end greater than 1/16 inch.
Minor A	
201	Marking (other than 102) not as specified or illegible.
202	Color not as specified.
Minor B	
251	Dimensions (other than 151 or 152) not as specified.

The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table IV. The unit of product for examination purposes shall be one tray of each style.

TABLE IV. Examination criteria.

Type defect	Inspection level	Acceptable Quality Level (percent defective)
Major A	II	1.0
Major B	S-2	1.5
Minor A	II	2.5
Minor B	S-2	4.0

4.2.2 Testing. Trays shall be tested as specified in 4.3 and in the ASTM Test Methods. A sufficient quantity of material from each lot of polypropylene used in the fabrication of the insert trays shall be selected for the physical property tests specified in table I.

4.2.3 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.

4.3 Tests. Tests shall be conducted to determine compliance with the requirements of this specification.

4.3.1 Tensile yield strength. Tensile yield strength shall be determined in accordance with ASTM Method D638.

4.3.2 Impact resistance. Impact resistance shall be determined in accordance with the notched Izod test of ASTM Method D256.

4.3.3 Deflection temperature. Deflection temperature shall be determined in accordance with ASTM Method D648.

4.3.4 Rockwell hardness. Rockwell hardness shall be determined in accordance with ASTM Method D785.

5. PACKAGING.

5.1 Preservation-packaging. Preservation-packaging shall be level A or C, as specified (see 6.1).

5.1.1 Unit of issue. Each tray of the style specified constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Cleaning. Trays shall be cleaned in accordance with MIL-P-116, process C-1.

5.1.2.2 Drying. Trays shall be dried in accordance with MIL-P-116.

5.1.2.3 Preservative application. Preservatives shall not be used.

5.1.2.4 Unit packaging. Each tray shall be packaged in accordance with MIL-P-116, method III in a box conforming to PPP-B-566 or PPP-B-676.

5.1.2.5 Intermediate package. Six unit packages shall be packaged in a box conforming to PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure of the box shall be adequate to prevent accidental opening.

5.1.3 Level C. Trays shall be clean, dry and individually packaged in a manner that will afford adequate protection against physical damage during shipment from the supply source to the first receiving activity.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T09:15Z
Check the source to verify that this is the current version before use.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Twelve intermediate packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Twelve intermediate packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged trays shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier rules and regulations.

5.3 Marking.

5.3.1 Interior packages. Each unit and intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES.

6.1 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

a. Title, number and date of this specification.
b. Style required (see 1.2).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 National Stock Number (NSN) coverage. This specification covers the following items appearing in the Federal Supply Catalog:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-00-435-3076	TRAY, INSERT, DENTAL INSTRUMENT AND SUPPLIES CABINET, Plastic, Long-Handled Instruments, No. 16A.
6520-00-435-3077	TRAY, INSERT, DENTAL INSTRUMENT AND SUPPLIES CABINET, Plastic, Clamps and Matrix Bands, No. 17.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:15Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6520-00-435-3078	TRAY, INSERT, DENTAL INSTRUMENT AND SUPPLIES CABINET, Plastic, Polishing Strips, No. 18.
6520-00-435-3079	TRAY, INSERT, DENTAL INSTRUMENT AND SUPPLIES CABINET, Plastic, Foil Carriers, Pliers, and Shears, No. 19.
6520-00-435-3080	TRAY, INSERT, DENTAL INSTRUMENT AND SUPPLIES CABINET, Plastic, Disks, Wheels, and Stones, No. 20.

6.3 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Military custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Project No. 6520-1436

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:15Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:15Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-T-36754A TRAY, INSERT, DENTAL INSTRUMENT AND SUPPLIES CABINET, PLASTIC

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:15Z
Check the source to verify that this is the current version before use.