INCH-POUND

A-A-50806C
15 August 1993
SUPERSEDING
A-A-50806B
19 January 1988

COMMERCIAL ITEM DESCRIPTION

BOTTLE, URINE SPECIMEN, SHIPPING

The General Services Administration has authorized the use of this Commercial Item Description

This Commercial Item Description covers a compartmented shipping box containing plastic screw cap bottles for mailing urine specimens to biomedical testing laboratories.

Salient Characteristics:

Shall be a compartmented shipping box containing twelve (12) 3-ounce capacity screw cap bottles with leakproof closures. Shall be suitable for mailing urine specimens to biomedical testing laboratories. The compartmented box with urine specimen filled bottles shall be suitable to be shipped via ground transportation, and via airplanes within unheated and unpressured cargo holds.

Bottle. Shall be a wide mouth cylindrical 3-ounce (90 milliliter) capacity bottle. The bottle shall be fabricated from a natural high density polyethylene plastic that is translucent. The translucent plastic shall have sufficient clarity for checking the level of a liquid filled bottle. The top portion of the bottle shall have a threaded neck finish. A rounded shoulder shall be formed below the bottom level of the neck. The wall thickness at the center of the bottle shall be at least 0.75 millimeters (mm). The neck finish shall have a continuous type exterior thread to accommodate a screw cap closure. The diameter of the mouth opening shall be at least 1-1/4 inches. The top surface at the mouth opening shall be level and free of ridges. The fill capacity of the bottle shall be 90 milliliters(ml) +/- 15 ml. The fill capacity in the bottle shall be located at a liquid level that does not extend above the bottom level of the neck or below the bottom level of the shoulder. The 75 ml liquid fill level shall be located under the shoulder

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

round-off on the bottle. The bottle shall include permanent graduation markings with Arabic numerals for the 45, 60 and 75 milliliter liquid levels. Additional graduation markings shall be at the option of the manufacturer. The graduation markings shall also indicate that the graduations are in terms of "milliliters" or its appropriate abbreviation. The 75 ml graduation shall include markings that indicate that it represents the level for a recommended maximum fill (e.g. "maximum fill" or appropriate abbreviation). As an alternate, the marking for the 75 ml graduation may indicate that the bottle should not be filled above that graduation. The overall height of the bottle with the appropriate screw cap closure shall not exceed 3-1/4 inches.

Screw cap. Shall be a screw cap that includes a continuous type thread along the inner edge. The screw cap shall be constructed to provide a leakproof closure on the mating bottle. The screw cap shall be either a rigid one-piece metal or plastic cap with an inside liner, or a rigid one-piece polypropylene linerless screw cap that includes an integral inner seal ring. A screw cap with liner, if supplied, shall include a liner of sufficient thickness to form an effective leakproof seal after closure on the mating bottle. The liner shall be secured to the inside of the screw cap in such a manner that it does not fall out, and does not go off-center when the screw cap is tightened on the mating bottle. The exposed surface of the liner shall be impervious to liquids such as water and urine. The linerless screw cap, if supplied, shall include an inner seal ring that makes contact with the inner surface of the neck on the mating bottle. The tightening of the linerless screw cap on the mating bottle shall result in a tight fit of the seal ring against the inner surface of the mating bottle. The screw caps shall provide leakproof closures on the mating bottles by hand tightening. The effective applied torque for proper hand tightening shall not exceed 20 inch-pounds.

Compartmented shipping box. The compartmented shipping box shall be fabricated of suitable corrugated fiberboard, having a bursting strength of 250 psi. Box design shall include full height, half-slotted partitions. Box shall be suitable for holding and mailing 12 filled specimen bottles in an upright position. Box shall have nominal dimensions of 8-5/8 inches long by 6-1/4 inches wide by 3-1/2 inches high. Box shall be marked on 2 sides with the word "UP" and an arrow toward the top of the box.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Unit package. The unit package shall be a box fabricated of suitable corrugated fiberboard, having a bursting strength of 250 psi. The box shall be marked on 2 sides with the word "UP" and an arrow toward the top of the box. Dimensions of the box shall be sufficient to permit placement of 10 compartmented shipping boxes, each containing 12 capped bottles, into two layers of 5 boxes each. The box shall be suitable for reuse for holding and mailing 10 compartmented shipping boxes, each containing 12 filled specimen bottles.

Leakproof characteristics. The screw cap shall form a leaktight fit after proper closure on the bottle. Leakage tests shall be conducted on the screw capped bottle filled with synthetic urine. The leakage tests shall be conducted under conditions that include the following:

Room temperature
120 degrees Fahrenheit (F)
After being frozen solid and then subjected to a vacuum air pressure
After being thawed and then subjected to a vacuum air pressure

The testing procedures and sampling plan shall be as indicated in the Quality Assurance Provisions of this document.

Workmanship. The bottle, screw cap, shipping box and shipping container shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 10 compartmented shipping boxes (120 capped bottles), as specified, constitutes one unit. Each compartmented shipping box shall include 12 capped bottles.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the bottle and screw cap in accordance with commercial practice. Sampling for inspections and testing shall be in accordance with commercial practice. In addition, leakage tests shall be performed on the bottle and closure screw cap as hereinafter indicated.

Bottle leakage testing (room temperature). The bottle with screw cap closure shall be tested for leakage at room temperature. The bottle shall be filled with synthetic urine to the level indicated by the 75 ml graduation mark. Screw cap closure shall be applied hand tight. The synthetic urine filled bottle shall be vertically suspended in an inverted position for two (2) hours, and then placed in a horizontal position for two (2) hours. Any evidence of leakage from the bottle shall be cause for rejection of the bottle. Sampling for the test shall be in accordance with MIL-STD-105 and an S-1 inspection level. The unit of product for the leakage testing shall be an individual bottle with screw cap. The acceptance number shall be zero.

Bottle leakage testing (120 degrees (F)). The bottle with screw cap closure shall be tested for leakage at 120 degrees (F). The bottle shall be filled with synthetic urine to the level indicated by the 75 ml graduation mark. Screw cap closure shall be applied hand tight. The synthetic urine filled bottle shall be placed in an environment having a temperature of 120 degrees (F) +/- 5 (F). The filled bottle shall be vertically suspended in an inverted position for two (2) hours, and then placed in a horizontal position for two (2) hours. Any evidence of leakage from the bottle shall be cause for rejection of the bottle. Sampling for the test shall be in accordance with MIL-STD-105 and an S-1 inspection level. The unit of product for the leakage testing shall be an individual bottle with screw cap. The acceptance number shall be zero.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Leakage testing (frozen urine under vacuum conditions). The bottle shall be filled with synthetic urine to the level indicated by the 75 ml graduation mark. A screw cap closure shall be applied hand tight. The synthetic urine filled bottle shall be placed in a freezer chamber with a temperature of at least -40 degrees (F) for the appropriate time period for the liquid synthetic urine content to become completely frozen solid. The bottle with its frozen content shall be removed from the freezer and transferred to a vacuum chamber that operates at room temperature. The vacuum chamber shall be set for a 151 mm Hg vacuumatic pressure. Under the 151 mm Hg vacuumatic pressure condition, the bottle shall be vertically suspended in an inverted position for 1 hour and then positioned in a horizontal position for 1 hour. Any evidence of leakage from the bottle shall be cause for rejection of the bottle. If the bottle has shown no evidence of leakage in both vertical and horizontal positions, the frozen liquid shall be allowed to thaw back to a complete liquid form (with ambient conditions at option of the contractor). The bottle with the thawed liquid contents shall be subjected to an additional leakage testing procedure (see "Leakage testing (frozen/thawed urine under vacuum testing)" hereinafter specified. Sampling for the tests shall be in accordance with MIL-STD-105 and an S-1 inspection level. The unit of product for the leakage tests shall be an individual bottle with screw cap. The acceptance number shall be zero.

Leakage testing (frozen/thawed urine under vacuum testing). The bottle with synthetic urine contents shall be initially subjected to the testing procedure specified in the "Leakage testing (frozen urine under vacuum conditions)" paragraph. After completion of that testing procedure the frozen urine content shall be allowed to become completely thawed and returned to its liquid state. The liquid urine filled bottle shall be placed in the vacuum chamber that is at room temperature. The vacuum chamber shall be set for a 151 mm Hg vacuumatic pressure. Under the vacuumatic pressure condition, the bottle shall be suspended in the inverted position for 1 hour and then positioned in the horizontal position for 1 hour. Any evidence of leakage from the bottle shall be cause for rejection of the bottle. The sampling plan and acceptance number shall be as indicated in the previous paragraph.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimension will be considered on an equal basis with those manufactured using inch-pound unit, providing they fall within the tolerances specified using conversion tablets contained in the latest revision of the Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

Ordering data. Unit/intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6640-00-165-5778	BOTTLE, URINE SPECIMEN, SHIPPING 120's

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6640-1813

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-50806C	2. DOCUMENT DATE (YYMMDD): 15 AUGUST 1993

3. DOCUMENT TITLE: BOTTLE, URINE SPECIMEN, SHIPPING

4. NATURE OF CHANGE (*Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.*)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (*Last, First, Middle Initial*)	b. ORGANIZATION	
c. ADDRESS (*Include Zip Code*)	d. TELEPHONE (*Include Area Code*) (1) Commercial: (2) AUTOVON (*If applicable*)	7. DATE SUBMITTED (*YYMMDD*)

8. PREPARING ACTIFITY

a. NAME efense Personnel Support Center ATTN MSE	b. TELEPHONE (*Include Area Code*) (1) Commercial: (215)737-2870	(2) AUTOVON: -444-2870
c. ADDRESS (*Include Zip Code*) 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:47Z
Check the source to verify that this is the current version before use.