A-A-51405A
15 July 1988

COMMERCIAL ITEM DESCRIPTION

SPONGE, SURGICAL (NONWOVEN, RAYON, PATTIE, RADIOPAQUE)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a Nonwoven, Rayon, Radiopaque, Surgical Sponge Pattie, suitable for use with forceps as a miniature sponge in neurological surgical procedures.

Salient Characteristics:

The sponge shall be a sterile, nonwoven sponge, x-ray detectable, and suitable for use with forceps as a miniature sponge in neurological surgical procedures.

The sponge shall be fabricated of random oriented rayon fiber fabric, double-needled and lightly bonded and shall retain its shape when wet.

The sponge shall resist shredding, stretching or shedding, wet or dry. Radiopaque stitching or bonding shall run the full length of the sponge and shall be readily detected by x-ray.

The sponge shall be single strung and the string shall be firmly attached. The sponge string shall be at least 6 inches long and shall be dyed green, blue or black for visibility and easy retrival. The dyes shall meet the requirements of the Federal Food, Drug and Cosmetic Act for dyes.

The sponge shall be capable of absorbing more than five times its own weight of water in less than a second.

AMSC/NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

Sponges shall have the following dimensions for the applicable sizes given below in Table I.

TABLE I. DIMENSION

ITEM DIMENSIONS (IN INCHES - NOMINALS)	SIZE
1/4 by 1/4 inch	1
1/2 by 1/2 inch	2
1/2 by 1-1/2 inches	3
1/2 by 2 inches	4
1 by 1 inch	5
1 by 3 inches	6
3 by 3 inches	7

Ten sponges of one size only, shall be precounted and mounted on a count card.

When testing for sterility, shall be in accordance with the Sterility Testing Method specified in the U.S.P.

Workmanship. The Sponge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). Ten sponges of one size only shall be mounted on a count card and packaged in a peel-open type immediate container capable of maintaining sterility of contents unless container is damaged or opened. One package containing 20 peel-open type containers (200 sponges) of one size only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one immediate container of 10 sponges. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:55Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-01-139-7547	SPONGE, SURGICAL, Non-woven Rayon, Pattie, Radiopaque, 1/4 by 1/4 inch, (.64 by .64 cm), 200s	1
6510-00-261-9015	SPONGE, SURGICAL, Non-woven Rayon, Pattie, Radiopaque, 1/2 by 1/2 inch (1.3 by 1.3 cm), 200s	2
6510-01-190-4401	SPONGE, SURGICAL, Non-woven Rayon, Pattie, Radiopaque, 1/2 by 1-1/2 inch, (1.3 by 3.8 cm), 200s	3
6510-01-190-0884	SPONGE, SURGICAL, Non-woven Rayon, Pattie, Radiopaque, 1/2 by 2 inches, (1.3 by 5.1 cm), 200s	4

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION	SIZE
6510-01-113-9209	SPONGE, SURGICAL, Non-woven Rayon, Pattie, Radiopaque, 1 by 1 inch, (2.5 by 2.5 cm), 200s	5
6510-01-139-7542	SPONGE, SURGICAL, Non-woven Rayon, Pattie, Radiopaque, 1 by 3 inches, (2.5 by 7.6 cm), 200s	6
6510-01-110-9191	SPONGE, SURGICAL, Non-woven Rayon, Pattie, Radiopaque, 3 by 3 inches, (7.6 by 7.6 cm), 200s	7

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PCOs 06608
06774
96374
59524
60858

Project No. 6510-0721

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51405A	SPONGE, SURGICAL (Nonwoven, Rayon, Pattie, Radiopaque)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:55Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (*DO NOT STAPLE*), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-[redacted]
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:55Z
Check the source to verify that this is the current version before use.