A-A-53781
14 December 1988

COMMERCIAL ITEM DESCRIPTION

MASK, ORONASAL, RESUSCITATOR AND ANESTHESIA, PEDIATRIC

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers pediatric oronasal masks for administering oxygen and anesthetic gases to pediatric patients.

Salient characteristics:

General. The mask shall be a sized face mask which is lightweight for administering oxygen and anesthetic gases to infants and pediatric patients.

Materials. The mask shall be fabricated from flexible, transparent plastic material which is non-conductive.

Style and design. The mask shall be lightweight and provide a tight seal to limit air leaks around the edges. The masks shall be designed to minimize dead air space in the internal cavity. They shall be molded to fit the contour of anatomically "average" infant faces. The dome in the center of the mask shall contain a tapered hole for securing a 22 mm adapter. The tapered hole shall maintain a gas-tight seal when tested as specified herein. The masks shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22342.

Physical Properties.

Elongation. The elongation of the mask shall be a minimum of 200% when tested in accordance with ASTM Designation D638.

Hardness. The mask shall have a Shore "A" durometer hardness reading of 65 + 5 when tested in accordance with ASTM Designation D785.

Sterilization. The masks shall be capable of withstanding repeated sterilization in ethylene oxide or soaking in disinfectant solutions.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Performance. The tapered hole of the mask shall maintain a gas tight seal when tested as specified herein.

Workmanship. The masks shall be free from defects which detract from their appearance or impair their serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One mask of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more characteristics.

Source: https://assist.dla.mil - Downloaded: 2026-01-31T04:24Z
Check the source to verify that this is the current version before use.

Performance. The tapered hole of the mask shall maintain a gas tight seal when tested as specified herein.

Workmanship. The masks shall be free from defects which detract from their appearance or impair their serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One mask of the size specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as follows. At an Inspection Level S-2, dimensional testing shall be at a 2.5% AQL, elongation and hardness testing shall be a 1.5% AQL, and pressure testing shall be a 1.0% AQL. The unit of product for test purposes shall be one mask.

Pressure test. The mask shall be subjected to the pressure test for non-rigid female connectors described in Section 6.1.2 of the Compressed Gas Association Pamphlet M-1 Standard for 22 mm anesthesia breathing circuit connectors. The mask shall be capable of maintaining a gas-tight seal when an internal pressure of 150 mm Hg above ambient is applied.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-7793	MASK, ORONASAL, RESUSCITATION AND ANESTHESIA, Pediatric, Size 1
6515-01-139-7794	MASK, ORONASAL, RESUSCITATION AND ANESTHESIA, Pediatric, Size 2
6515-01-139-7795	MASK, ORONASAL, RESUSCITATION AND ANESTHESIA, Pediatric, Size 3

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3558

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53781	MASK, ORONASAL, RESUSCITATOR AND ANESTHESIA, PEDIATRIC

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.