A-A-53783
20 February 1989

COMMERCIAL ITEM DESCRIPTION

MIRRORS AND MIRROR HANDLE, MOUTH EXAMINING

The General Services Administration has authorized the use of this Commercial Item Description.

Salient characteristics:

This commercial item description covers dental mirrors and a mirror handle suitable for use in oral examinations.

The mirrors and mirror handle shall be of the following types and sizes, as specified:

Type I - Front reflective surface mirror.

Size 1 - 1 inch in diameter.

Size 2 - 3/4 inch in diameter.

Type II - Mirror handle.

General. The mirrors and mirror handle shall be in accordance with the applicable requirements as specified herein.

Materials.

Mirror glass. Glass for mirrors shall conform to Type II, class 1, quality q4-AA or q5-A, photo glass in accordance with Federal Specification DD-G-451, except as specified herein. Composition of glass shall be of such quality as to withstand prolonged use, frequent heat, and chemical sterilization without dulling the surface.

Mirror frame and shank. Mirror frame and shank shall be fabricated of either chromium plated brass or austenitic stainless steel. The brass shall conform to alloy number 260 of ASTM B36 for chemical composition only. As an alternate, the mirror shank only shall conform to alloy numbers 342 or 353, as specified in ASTM B121 or conform to alloy number 360, as specified in ASTM B16 for chemical composition only. The corrosion-resisting steel shall conform to type 302, 303 or 304, as specified in ASTM F899 for chemical composition only. The corrosion tests, as specified herein shall apply.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Handle. The handle shall be fabricated of chromium plated brass, austenitic or martensitic stainless steel. The brass shall conform to alloy number 360 of ASTM B16 or brass alloy C 36000 of ASTM B36 for chemical composition only. The austenitic or martensitic corrosion-resisting steel shall conform to ASTM F899 for chemical composition only. The corrosion tests, as specified herein shall apply.

Style, design and dimensions. Mirrors shall be of the short threaded shank type, with tapered shoulder, suitable for use in oral examinations. Mirrors shall consist of mirror glass, cushioning material, mirror frame and shank. Mirror and mirror handle shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 20773 for the type and size specified.

Type I mirror. Type I mirror shall be a plane front reflective surface mirror.

Type II handle. Type II handle shall be suitable for use with Type I mirror.

Construction.

Mirror glass. The front surface of the Type I mirror glass shall be coated with rhodium. The rhodium coating shall be no less than 700 Angstrom units thick (as determined by the optical density requirements), and shall be obtained by means of the vacuum deposition process. The coating shall be uniform with an optical density of no less than 3.0 and a luminous reflectance of 70.0 percent maximum. Reflectance and optical density shall be tested as specified herein. The rhodium coating shall be firmly adherent and shall show no visual evidence of flaking, peeling, pitting, or any other defect; nor shall the mirror dull, fog or loosen in the frame, when subjected to the boil test and sterilization tests, as specified herein.

Mirror, mirror frame, and shank. The shank shall be securely welded or brazed to the frame. The joint formed shall be uniform, with both members blending, and it shall be free from cracks, crevices or any other defects. The mirror frame shall not show any evidence of distortion as a result of welding or brazing. The edge of the mirror frame shall be formed over the beveled edge of the glass to hold the mirror securely within the frame. The formed edge of the frame shall be below the front surface of the mirror and shall be free from irregularities, sharp edges, or burrs.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:24Z
Check the source to verify that this is the currentversion before use.

Spring cushioning shims, layers of water-resistant paper (impregnated or coated with waterproofing material) or a shim of an elastomer material shall be provided between the mirror frame and glass to protect the surface and to prevent the glass from moving within the frame.

The spring cushioning shim shall be fabricated of brass. The water-resistant paper or the elastomer material shall not separate or melt in a way which would render the mirror unserviceable or contaminate the reflecting surface. The metal frame and shank shall not show any evidence of corrosion, scumming, or dulling of the surface, when tested for the boil and sterilization tests, as specified herein.

Mirror handle. The mirror handle shall be of one-piece construction and shall be suitable for use with type I mirrors. The threaded shanks of type I mirrors shall securely fit in the cone-socket of type II handle.

Finish. All exposed metal parts of the mirror shall be smooth and shall have a bright or satin finish. The handle shall be smooth and shall have a bright or satin finish, except for the knurled portion of the round style handle.

Identification marking. The back of each mirror shall be permanently and legibly marked with the commercial size, and name or registered trademark of the manufacturer. Mirrors shall be marked with the commercial designation for the applicable sizes. In addition, mirrors of both sizes may be, but are not required to be marked "Front Surface" or "FS". Each handle shall be permanently and legibly marked with the name or registered trademark of the manufacturer. The back of the mirror and handle shall have the word "Stainless" or "SS". For brass composition of the mirrors and handle, the word "Stainless" or "SS" shall be omitted.

Workmanship. The mouth examining mirrors and mirror handle shall be free from pits, cracks, burrs, sharp edges or any other defects which may affect proper functioning, appearance, or may impair their serviceability..

Unit. Each (EA). One mirror and mirror handle, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the steel may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible; the same sample shall be used for the determination of two or more inspection characteristics. The unit or product for inspection purposes shall be one mirror, or one mirror handle, as applicable. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for the dimensions of the mirror and handle which shall be conducted at inspection level S-1, AQL 2.5 and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Examination. Mirrors and mirror handle shall be examined for defects, including but not limited to, those listed below.

Defects.

Mirrors

Size not as specified.
Style and design not as specified.
Construction not as specified.
Mirror glass not free from cracks, chips, scratches, seeds, bubbles, or breaks in the reflective coating.
Mirror glass insert not firmly secured within mirror frame.
Shank not properly and securely welded or brazed to mirror frame.
Shank not free from burrs and sharpness.
Welded or brazed joint between shank and frame not smoothly finished.
Metal parts not free from cracks and discoloration.
Threads of shank do not comply with thread size.
Mirror shank not threaded.
Dimensions of mirror (glass, frame, and shank) not as specified.
Plating on brass parts not adherent (not free from blistered, peeled or bare (unplated) surface areas).
Finish not brightly plated.
Finish not free from coarse grind marks; is excessively scratched or marred, or has abraded surface areas.
Identification marking not present or not permanent, legible, and complete.
Mirror not free from dirt, oil, polishing compound, and foreign matter.

Handle.

Handle style and design not as specified.
Handle not threaded.
Threads of handle do not comply with thread.
Handle finish not as specified.
Handle finish not free from coarse grind marks.
Handle not straight.
Identification marking not present or incomplete.
Dimensions of handle not as specified.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:24Z
Check the source to verify that this is the current version before use.

Tests.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instrument shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Corrosion tests.

Boil test. The mirrors and all handles shall be tested for corrosion resistance in accordance with the requirements of ASTM F1089.

Copper sulfate test. The handles fabricated of martensitic stainless steel shall be tested for corrosion resistance in accordance with the requirements of ASTM F1089.

Sterilization. The finished mirror shall be autoclaved under 27 pounds of pressure and at 270° Fahrenheit (F) for 20 minutes. The assembled mirror (finished unit) shall be examined for compliance with the requirements for construction.

Reflectance test. Luminous reflectance shall be tested in accordance with MIL-M-13508.

Optical density. Optical density of each coated mirror shall be determined by means of visual comparison of the rhodium plated mirror glass with a sample having an optical density of 3.0 as determined by means of a suitable densitometer. In case of dispute as to compliance with requirements by visual comparison, a suitable densitometer of standard make and model shall be used as the referee method.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Numbers	Item Identification
6520-00-541-9350	HANDLE, MOUTH EXAMINING MIRROR.
6520-00-782-2648	MIRROR, MOUTH EXAMINING, Plane Glass, Front Reflective Surface, 1 inch diameter.
6520-01-056-7378	MIRROR, MOUTH EXAMINING, Plane Glass Front Reflective Surface, 3/4 inch diameter.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the currentversion before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

49056
05857

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

PROJECT NO. 6520-2008

LOCATION: NCR DISK #5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53783	Mirrors and Mirror Handle, Mouth Examining

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.