A-A-53787
15 December 1988

COMMERCIAL ITEM DESCRIPTION

COLLECTING TUBING SET, UTERINE, PLASTIC, DISPOSABLE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile disposable collecting tubing set suitable for vacuum aspiration of uterine contents.

Salient characteristics:

General. Shall be a plastic disposable collecting tubing set specifically designed for vacuum aspiration of uterine contents. The collecting tubing set serves to control the suction curettes attached to the distal end and to connect them to the aspiration collection bottle and to the vacuum source.

Characteristics. The set shall have a handle which is rotatable and a slip ring for finger tip vacuum control. It shall be 6 feet long and have a tapered fitting on the proximal end of the set for connecting to the aspiration collection bottle. On the distal end shall be a female taper fitting to accommodate male taper fitting on the proximal end of the suction curettes. The set shall accommodate suction curettes ranging in size from 4 mm OD to 12 mm OD.

The set shall be composed of a handle assembly, flexible tubing and a fitting (to the aspiration collection bottle). The handle assembly shall contain a mechanism to control vacuum, allow the handle to swivel and have a female taper fitting to connect with the suction curettes. The set shall be capable of operating as indicated at a vacuum of 510 mm Hg. The set shall be in accordance with DPSC Drawing No. 24507.

Material. The set shall be fabricated of clear, non-conductive plastic.

Sterility. The collecting tubing sets shall be sterile when tested in accordance with applicable portions of the USP. The sterility assurance level shall be 0.0001 as a minimum.

AMSC N/A FSC-6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Workmanship. The sets shall be processed in such a manner as to be uniform in quality and free from defects that would affect the life, serviceability, or appearance of the end item.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing ten collecting tubing sets, each set individually packaged sterile in a sealed peel-open type container capable of maintaining sterility of contents, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Operational test. When the set is connected to a suitable vacuum source until a vacuum of 510 mm Hg has been attained the handle shall be able to rotate and control the degree of vacuum and the tubing shall not collapse.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. At an Inspection Level of S-2, dimensional testing shall be at a 2.5% AQL and operational testing shall be at a 1.0 percent AQL. The unit of product for test purposes shall be one collecting tubing set.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-156-2452	COLLECTING TUBING SET, UTERINE, PLASTIC, Disposable

NOTE 2: The following suction curettes and their corresponding National Stock Numbers (NSNs) fit the tubing collection set described in this specification:

NSN	ITEM IDENTIFICATION
6515-01-153-4933	Curette, Suction, Uterine, Semi-Rigid, 8 mm Outer Diameter, 10s
6515-01-185-8836	Curette, Suction, Uterine, Semi-Rigid, Curved, 10 mm Outer Diameter,10s

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

NSN	ITEM IDENTIFICATION
6515-01-185-8837	Curette, Suction, Uterine, Semi-Rigid, 12 mm Outer Diameter, 10s

Ordering data (intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PPS
FDA-MFQAS

PGC 53323

Project No. 6515-3387

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53787	COLLECTING TUBING SET, UTERINE, PLASTIC, DISPOSABLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 1426 82 MAR PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil – Downloaded: 2026-10-26T09:242
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.