INCH-POUND

A-A-58074
31 JULY 1996

COMMERCIAL ITEM DESCRIPTION

TEST SET ,ULTRASONIC, BOND

The General Services Administration has authorized the use of this commercial item description for all federal agencies.

1. SCOPE. This commercial item description covers the minimum performance requirements for a portable ultrasonic test set used for ultrasonic bond testing equipment of metallic and nonmetallic composite material.

2. Salient characteristics:

2.1 The test set shall provide reliable means for performing ultrasonic inspections to detect disbonds in metal to metal, metal to nonmetal and nonmetal to nonmetal substances.

2.2 Frequency Range. The frequency range of the test set shall be from 500 hertz to 1 megahertz.

2.3 The test set shall operate on 115 to 230 volts alternating current, 50-60 hertz automatic switched or battery power with provision to charge the battery power pack from the AC line when connected.

2.3.1 Battery Packs. The battery pack shall be able to power the test set for six hours while in use on battery power.

2.3.2 The test set shall be provided with two battery packs.

Beneficial comments(recommendations, additions, deletions) and any pertinent data which may be of use in improving the document should be addressed to San Antonio ALC/TILDD, 306 Tinker Drive, Bldg. 207, Kelly AFB TX 78241-5916(DD Form 1426) appearing at the end of this document or by letter.

AMSC/NA FSC 6635

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.

2.4 Gain. The gain shall be 0.0 to 50.0 decibels.

2.5 Connector. The test set shall have an analog output to connect to with an external video display.

2.6 The test set shall provide visible and audible indications of disbond occurrence either in positive or negative directions . The alarm shall be selected and adjusted for momentary or continuous audible alarm.

2.7 The automatic signal type detection of synchronous, phase and amplitude shall be selected by the type of probe to be used .

2.8 The display shall provide dual x-y screen with side by side presentation of active and recalled stored information. The display presentation shall be a minimum of 11 inches in size measured diagonally.

2.9 The test set shall be able to store a minimum of 50 parameters and users shall able to easily retrieve from memory any parameter stored at any time.

2.10 Cable. The test set shall have a RS-232 cable and handshake software.

2.11 Software. The test set shall have the latest commercially available version of the operating system software.

2.12 The test set shall provide an electronic null that will center the signal or return to home position on the active screen in all functional modes.

2.13 Probes.

2.13.1 The test set shall be provided with one each 30 kHz, 60 kHz, 110 kHz, 165 kHz, 250 kHz and 330 kHz resonance probes.

2.13.2 The test set shall be provided with one pitch/catch impulse probe, one pitch/catch sweep probe and one mechanical impedance probe.

2.13.3 The test set shall be provided with a method of probe recognition for the test mode selected.

2.13.4 All probes shall provided with a quick release type connector.

2.14 The maximum weight of the test set with battery pack shall be 20 pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.

2.15 The test set shall be equipped with high impact supports on the bottom surface.

2.16 Protective case. A protective carrying case for the instrument shall be designed to exclude harmful or undesirable elements, and should include space for the probes and accessories.

2.17 The test set shall be able to operate in temperature range of -10 ºF to + 130 ºF with a relative humidity of 95%.

2.18 The test set shall be capable of operating in hazardous explosive area which shall have the been approved by the Underwriter's Laboratories (UL), Factory Mutual (FM) or reputable independent testing laboratory.

3. REGULATORY REQUIREMENTS.

3.1 Recovered Materials. The contractor is encouraged to use recovered materials to the maximum extent practical, in a accordance with paragraph 23.403 of the Federal Acquisition Regulation(FAR).

3.2 Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pounds test set, provided they fall within specified tolerances using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this Commercial Item Description are met. If a product is manufactured to metric dimensions and those dimensions exceed tolerances specified in the inch-pound test sets, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

4. Quality assurance provisions

4.1. Product conformance. The products provided shall meet the salient characteristics of this commercial item description, conform to the producer's own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance.

4.2 Market acceptance criteria. The contractor shall have had available in the commercial market place for use for not less than one year bond ultrasonic test set that are the similar to the one to be provided on this Commercial Item Description. The contractor must have sold at least 100 test sets of the model provided.

5. Preservation, packaging, packing, labeling, and marking. Preservation, packaging, packing, labeling, and marking shall be specified in the contract or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.

6.0 NOTES

6.2 Procurement data. The procuring activity should specify the preferred options permitted herein and include the following information in procurement documents:

a) Title, number, and date of this commercial item description.
b) Specify special packaging requirements.
c) certification of commercial item prior to first contract delivery.

6.3 National Stock Number(s). The applicable NSNs for this test set are:

NSN	REFERENCE NO.	CAGEC
6635-01-290-2927	93601	51294
6635-01-161-4551	CID95LDEA6635-0177	98750

6.4 Sources of documents.

6.4.1 Federal Standards may be obtained from the General Services Administration, Federal Supply Service, Office of Commodity Management, Washington, DC, 20406.

6.4.2 Federal Acquisition Regulations may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC, 20402.

6.4.3 Electronic Industries Association(EIA) document may be obtained from EIA, Administration Dept, 2001 I ST N W, Washington, DC, 20006-1807.

MILITARY INTEREST:

CIVIL AGENCY COORDINATING ACTIVITY: GSA-FSS

CUSTODIANS:
Air Force-99

DLA-GS

PREPARING ACTIVITY:
AIR FORCE-82

Review Activities:
DLA - GS

AGENT:
AIR FORCE-99

PROJECT NUMBER: 6635-0170

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waiver any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-58074	96/07/31

3. DOCUMENT TITLE
TEST SET, ULTRASONIC, BOND

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME *(Last, First, Middle Initial)*	b. ORGANIZATION	
c. ADDRESS *(Include Zip Code)*	d. TELEPHONE *(Include Area Code)* (1) Commercial (2) AUTOVON *(If applicable)*	e. DATE SUBMITTED *(YYMMDD)*

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE *(Include Area Code)*	
SA-ALC/TILDD	(1) Commercial (210) 925-6314	(2) AUTOVON 945-6314
c. ADDRESS *(Include Zip Code)* 306 TINKER DR, BLD 207 KELLY AFB TX 78241-5916	**IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:** Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD FORM 1426, OCT 89 ***(EF-V1)*** *(PerFORM PRO)* PREVIOUS EDITIONS ARE OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:00Z
Check the source to verify that this is the current version before use.