
FED. SUP CLASS
6640



PART NUMBER	CAPACITY mL	TOL. ± mL	GRADUATED INTERVAL mL	NUMBER INTERVAL EACH mL
MS35982-1	0.1	0.005	0.01	0.01
MS35982-2	0.2	0.008	0.01	0.02
MS35982-3	1.0	0.02	0.01	0.10
MS35982-4	2.0	0.02	0.10	0.20
MS35982-5	5.0	0.04	0.10	1.0
MS35982-6	10.0	0.06	0.10	1.0
MS35982-7		CANCELED		
MS35982-8		CANCELED		
MS35982-9		CANCELED		
DISPOSABLE PIPET				
MS35982-10	1.0	0.02	0.01	0.10
MS35982-11	2.0	0.04	0.1	0.20
MS35982-12	5.0	0.15	0.1	1.0
MS35982-13	10.0	0.30	0.1	1.0

MATERIAL: Pipets shall be made of borosilicate glass in accordance with ASTM E438, Type I, Class A or B.

DESIGN: Pipets shall be one piece. Any cross section of the pipet taken in a plane perpendicular to the longitudinal axis shall be circular. Delivery tips shall be made with a gradual taper. The end of the tip shall be perpendicular to the longitudinal axis, and the outside edge of the tip shall be slightly beveled or fire polished.

All pipets shall be designed for use with cotton plugs. Sizes 0.1 through 2.0 mL shall have plain fire polished tops that will permit insertion of plugs; while sizes 5.0 and 10.0 mL shall have a constriction to allow for cotton plugging. The outside edge of the mouthpiece shall be beveled slightly and fire polished.

All pipets shall have permanently fused or etched-on graduation marks. All pipets shall be calibrated "to deliver" (symbol TD) at 20°C, when the last drop is blown out.

NOTES:
1. Color coding bands on pipets are optional.
2. The Kahn type pipet has the same design as MS35982-1 and -2 pipets.
3. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

Ⓑ ENTIRE STANDARD REVISED

APPROVED 23 Dec 1959 REVISED Ⓐ 14 July 1969 Ⓑ 4 OCT 1983

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS, YD
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		PIPET, SEROLOGICAL, GENERAL PURPOSE AND KAHN TYPES	MS35982
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1205

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.

[redacted]

FED. SUP CLASS
6640

4. For design feature purposes, this standard takes precedence over procurement documents referenced herein.

5. Referenced documents shall be of the issue in effect on date of invitations for bid.

6. FSCM: 14674, 64484, and 75278.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

is approved for use by all
ies of the Department of
all new engineering and
d for repetitive use shall
ument.

APPROVED REVISED

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:32Z
Check the source to verify that this is the current version before use.