A-A-53584
29 July 1988

COMMERCIAL ITEM DESCRIPTION

RESERVOIR ASSEMBLY, SIMULATED BLOOD

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a reservoir and pump assembly intended to emit simulated blood to a bleeding moulage.

Salient characteristics:

The unit shall consist of a bag (reservoir), upper section tubing with a sliding occluding device, a pump bulb with valving, and lower section tubing with moulage connector at bottom end.

The bag shall be fabricated from 12 mil vinyl sheet conforming to Federal Standard L-P-375, Type II, Class II. The bag shall have a volume capacity of, at least, 16 fluid ounces. There shall be a flap at the top of the bag with 2 elongated eyelets for use to secure the bag by means of a one inch wide belt, strapping, or a 5/16 inch diameter hook. The eyelets shall be appropriately positioned to facilitate suspension of the bag by belting/strapping which is secured to a mock victim. The bag shall have a water tight valve port subassembly near the top for filling purpose. The filling valve cap shall not be removeable from the bag.

The tubing shall be clear commercial quality vinyl, approximately 1/8 inch I.D. Both sections shall be 30 inches (± 4 inches) long. The combined length of upper tubing section, pump bulb, and lower tubing section shall be 5 feet(± 4 inches) long.

The sliding occluding device shall be made from either aluminum or plastic material and shall not be removable from the upper section tubing. The device shall be capable of being set to occlude the tubing passage way, and set to allow full passage of fluid through the tubing.

The pump bulb shall be fabricated from clear vinyl. The pump bulb shall be appropriately shaped for hand squeezing/pumping and shall be capable of pumping one fluid ounce per squeeze. The pump bulb shall have a valving arrangement to suck simulated blood from the bag and expel it from the moulage set connector. The valving shall allow gravity feed of fluid from the bag out of the moulage connector.

AMSC N/A FSC 6910

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

The moulage connector shall be approximately 3/16 inch OD by approximately 3/32 inch ID by approximately 1/2 inch long (beyond tubing end). The connector shall be a straight shank which may have ring depressions to assist securing a sealed connection to a moulage. The connector shall be fabricated of brass or anodized aluminum.

The reservoir and pump assembly shall be unitized. When the bag is secured to a hook by a single eyelet, the bag shall be capable of supporting 5 pounds, attached to the bottom of the bag without causing damage. The connection points of tubing to bag, pump bulb, and moulage connector shall be capable of withstanding a tension force of 5 pounds without causing evidence of damage/joint failure.

The filling port salient characteristics shall be: The uncapped port opening shall be approximately 1 inch across to facilitate refilling without spilling. The cap shall be capable of being dislodged from a closed condition when subjected to an opening force not to exceed 3 pounds. The cap shall maintain a sealed closed condition when a full (of water) bag is dropped (freefall) from a height of 4 feet onto a steel or concrete surface. When the valve cap is open, its connection to the bag shall be capable of withstanding a tension force of 5 pounds between the valve cap and either one of the eyelets (see note 1).

Workmanship. The reservoir assembly shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package(PG). One package containing eight reservoir assemblies, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for Inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Tests shall include testing reservoir capacity, pump volume per squeeze, unitized strength of all joints, and sealing capability of refill cap.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 and an inspection level of II. The unit of product for examination purposes shall be one reservoir assembly.

Defects.

Unit not complete with all components.

Any component broken.

Manuals, as specified, not included.

Unit not clean.

Any rivet not properly installed; screw not tight.

Sampling for tests. Sampling for tests, (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-4. The unit of product for test purposes shall be one reservoir assembly.

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a first article sample has successfully passed all tests, performance test shall not be required at acceptance.

Examination. The reservoir assembly shall be examined to determine compliance with the requirements of this specification in accordance with the classification of defects in table II.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:21Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The Halkey Roberts filling port Part No. 330 AC Valve is reported as meeting the specified characteristics.

NOTE 2: The following National Stock Number is covered by this document:

6910-01-129-3194

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPQAS

Project No. 6910-0073

Location: CID/RESERVOIR/H12-3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53584	RESERVIOR ASSEMBLY, SIMULATED BLOOD

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:21Z
Check the source to verify that this is the current version before use.