REVIEW ACTIVITIES: OTHER - NSA
USER ACTIVITIES: ARMY - ME

FED. SUP CLASS
6640



PART NUMBER		
MS360[illegible]		

CANCELLED
CANCELLED AFTER 1DEC 1987, NO SUPER-
SEDING STANDARD

		8
		10.5
	50 ± 5	10.5
	100 ± 5	16
	50 ± 5	150 ± 5	16
	250 ± 10	150 ± 5	16

..all be in accordance with Federal Specification DD-G-541, Type I, Class A.

..s shall be smooth overall.

..r wall of the bowl shall diverge from the stem at an angle of 30° ± 1.2°.

The stem shall have oblique termination, and the tip shall be heavily glazed.

Top of funnels shall have a glazed reinforcing bead.

NOTES:
1. For design feature purposes this standard takes precedence over procurement document referenced herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 28 Jun 1961 REVISED 4 JUN 69 (B) 1 DEC 1987

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

(B) CANCELLATION

P.A. MU Other Cust MS 82	TITLE FUNNEL, COMMON, LABORATORY smooth	MILITARY STANDARD MS36084	
PROCUREMENT SPECIFICATION DD-F-806	SUPERSEDES:	SHEET 1 OF 1	

DD FORM 1 SEP 67 672-1 (Coordinated)

PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1969 - 341-520/A-7274

Project Number 6640-1463-1

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:34Z
Check the source to verify that this is the current version before use.