USER ACTIVITIES: ARMY - EL, ME

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

FED. SUP CLASS
6640



PART NUMBER	NOMINAL OVERFLOW CAPACITY ml	OUTSIDE DIAMETER mm	LENGTH ± 2.5 mm	WALL THICKNESS ± 0.2 mm
MS36180-1	0.7	6 ± 0.5	50	0.8
MS36180-2	6	12 ± 0.5	75	0.9
MS36180-3	10	13 ± 0.5	100	1.0
MS36180-8	16	15 ± 0.5	125	1.0
MS36180-4	24	16 ± 0.5	150	1.0
MS36180-5	55	25 ± 0.8	150	1.2
MS36180-6	75	25 ± 0.8	200	1.2
MS36180-7	130	32 ± 1.5	200	1.2

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A.

DESIGN: Test tubes shall be in accordance with Federal Specification NNN-T-189, Type I, Style I.

Test tubes shall be cylindrical with one end having a rounded bottom and top or open end having a lipless, fire-polished edge.

NOTES:
1. For design feature purposes this standard takes precedence over procurement document referenced herein.
2. Referenced documents shall be of the issue in effect on date of invitations for bid.

(A) ENTIRE STANDARD REVISED

APPROVED 30 DEC 1960 REVISED (A) 20 June 1969

P.A. MU Other Cust MS	TITLE TEST TUBE culture and general purpose type		MILITARY STANDARD MS36180
Procurement Specification NNN-T-189	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 SEP 63 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:35Z
Check the source to verify that this is the current version before use.