FED. SUP CLASS
6640



MATERIAL: Sampler shall be made of copper and lead.

DESIGN: The sampler shall have a funnel-shaped top and lead-weighted bottom. Trunnions riveted to sampler hold flat brass wire bail handle. Cork stopper has brass disc on top and brass loop for easy removal. Unit is aluminum finished. Dimensions shall be in accordance with ASTM D270.

USE: Sampler is used for ASTM D270, Method of Sampling Petroleum and Petroleum Products.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 7 JAN 1983 REVISED

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		SAMPLER, LIQUID	MS36499
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1133

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:42Z
Check the source to verify that this is the current version before use.