FED. SUP CLASS
6640



PART NUMBER	CAPACITY ML	DIVISION ML	BOTTLE CAP ML	ȝ NO.
MS36501-1	2	0.02	1000	24/40
MS36501-2	5	0.05	1000	24/40
MS36501-3	10	0.05	1000	24/40
MS36501-4	25	0.10	1000	24/40
MS36501-5	50	0.10	2000	29/42
MS36501-6	100	0.20	4000	29/42

MATERIAL: Buret shall be made of glass in accordance with Federal Specification, DD-G-541, Type I, Class A. Bulb shall be rubber and the stopcock plug shall be polytetrafluoroethylene.

DESIGN: Buret shall have automatic zero, all-glass check valve, a rubber bulb and glass reservoir bottle.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

APPROVED 7 JAN 1983 REVISED

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		BURET, AUTOMATIC	MS36501
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1133

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:42Z
Check the source to verify that this is the current version before use.