FED. SUP CLASS
6630



CANCELLED
CANCELLED AFTER 25 JUL 1986. NO SUPERSEDING STANDARD

PART NUMBER	O/A DIMENSIONS LG
MS36520-1	233.[illegible]

0 to 25

MATERIAL:

[illegible]ials suitable in all respects

DESIGN

[illegible]ced, portable instrument which [illegible]l conductivity. Four platinum [illegible] Wheatstone Bridge shall be used to [illegible]e of a sample gas. When the Bridge [illegible]o the lower resistance of hydrogen in [illegible]esults in a deflection on a galvanometer [illegible]e hydrogen concentration.

[illegible]dicate in direct terms the percent by volume [illegible]ir.

[illegible]ll contain two flame arrestors, eighteen inches of [illegible]ing, a rubber hand aspirater bulb, and two 1-1/2 volt batteries.

[illegible]ems differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.

B CANCELLATION

APPROVED 29 Jun 1967 REVISED A 9 DEC 1977 B 25 JUL 1986

This standard has been approved by the Department of the Navy and is mandatory for use by that activity. All other military activities are required to employ this standard where suitable.

P.A.	EA	TITLE	MILITARY STANDARD
Other Cust	YD 99	ANALYZER, HYDROGEN portable	MS36520
Procurement Specification NONE		SUPERSEDES:	SHEET 1 OF 1

DD FORM 1 SEP 66 672 (Coordinated)

EDITION OF 1 SEP 57 MAY BE USED.

NOTES TO PRINTER: REDUCE BORDER TO 6-27/32 x 9-3/4. PRINT LEFT BORDER-LINE ONE INCH FROM LEFT EDGE OF 8 x 10½ SHEET. PRINT BOTTOM BORDER LINE 3/8 INCH FROM BOTTOM EDGE OF 8 x 10½ SHEET. SIGNATURE FORM NOT TO BE INCLUDED ON 8 x 10½ SHEET.

DRAWN BY _____ _____
TYPED BY _____ _____
CHECKED BY _____ _____

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:42Z
Check the source to verify that this is the current version before use.