A-A-50059
September 22, 1983

COMMERCIAL ITEM DESCRIPTION

SCALES, PERSON WEIGHING

The General Services Administration has authorized the use of this commercial item description in preference to types I, and II of Federal Specification AAA-S-84.

Salient characteristics:

The scales shall be of the following types:

Type I - High beam, person weighing, platform scale with a measuring rod.

Type II - Automatic-indicating platform scale with a minimum capacity of 250 pounds.

The type I weighing range shall be from 0 to no less than 300 avoirdupois pounds. The scale shall consist of a platform, a notched and fractional beam, base, column and height measuring rod. The platform shall be metal and shall have a mat secured to the top surface. The mat shall be of rubber or similar water resistant material. The platform shall be at least 10 inches wide by 13 inches long. The beam shall be graduated from 0 to at least 250 pounds with 50 pound notched graduations. Fractional bar shall be smooth and have a 50 pound capacity with 1/4 pound graduations so that the total weight is the sum of the beam bar and the fractional bar. Beam bars shall be graduated on both sides so that the markings shall remain visible for the weight in question. Center of beam shall be at least 52 inches above the platform. The base shall be metal and shall not tip or rock on a flat surface. The column shall be metal and shall enclose the linkage mechanism connecting the beam to the platform. The measuring rod shall be telescoping and be designed to measure the height of a person standing on the platform. The measuring rod shall measure heights from 30 to 78 inches and shall have a tolerance of $\pm$ 1/8 of an inch. The sensitivity requirement shall be 4 ounces. The scale shall meet the requirements stated in the National Bureau of Standards Handbook 44.

The type II shall have an indicating means consisting of a reading face and an indicator. The indicator shall be parallel to the gradations on the reading face. Weight gradations shall be readable by a person standing on the platform. Reading face shall be covered by a window of transparent material. The scale capacity shall be at least 250 pounds and graduated in 1 pound divisions. Each 10 pound graduation shall be marked with its numerical

FSC 6670

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:38Z
Check the source to verify that this is the current version before use.

value. Each 5 pound graduation shall be longer than its adjacent 1 pound graduations. The platform shall be constructed of metal and shall be covered with a mat of either rubber or other water-resistant material. Mat shall be secured in position. The scale tolerance for 0 to 200 pounds shall be ± 2 pounds. The scale tolerance above 200 pounds shall be ± 1 percent. There shall be a zero adjust mechanism operable from outside the scale to allow the indicator to be reset to zero. Except as noted above, the scale shall meet the requirements stated in the National Bureau of Standards Handbook 44.

Certification. The contractor shall certify that the product offered meets the salient characteristics of the description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices, and is the same as the product sold in the commercial marketplace. The Government reserves the right to require proof of such conformance prior to the first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Packaging.

A. Commercial packaging. Each scale shall be packaged in accordance with commercial practice.

B. Commercial packing. Scales packaged as specified shall be packed in accordance with commercial packing complying with pertinent Uniform Freight Classification or National Motor Freight Classification, as applicable.

C. Export packaging. Each scale shall have the measuring rod detached from the platform (if applicable) and both packaged separately. The scales and parts shall be cleaned, preserved, cushioned and immobilized to afford protection against corrosion, deterioration and physical damage during shipment and storage. Interior fiberboard boxes used for packaging disassembled scales shall conform to class weather-resistant conforming to PPP-B-636. Each fiberboard box shall be closed in accordance with method IV as specified in the appendix of PPP-B-636.

D. Export packing.

Type I. One scale shall be packaged as specified above and packed together in a fiberboard box conforming to type CF, class weather-resistant, variety DW, grades V11c or V13C, style FOL of PPP-B-636. The contents of the box shall be cushioned blocked and braced to prevent movement while the box is in transit. Each shipping container shall be closed and strapped in accordance with the appendix of PPP-B-621.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:38Z
Check the source to verify that this is the current version before use.

Type II. Twenty scales shall be packaged as specified above and packed together according to the same requirement specified for a type I export pack.

E. Marking. Commercial markings are acceptable, provided the following information is included:

National stock number	Date of pack
Item description	Contract number
Quantity	Name, address and zip code of contractor

Source of information for nongovernment association documents.

National Motor Freight Traffic Association, Inc., Agent
National Motor Freight Classification
American Trucking Associations, Inc., Traffic Department
1616 P Street, N.W., Washington, DC 20036

Uniform Classification Committee, Agent
Uniform Freight Classification
Uniform Classification Committee, Suite 1106
222 South Riverside Plaza, Chicago, IL 60606

MILITARY INTERESTS:

Custodians

Army - GL
Navy - MS

Review Activities

Army - MD
DLA - GS

CIVIL AGENCY COORDINATING ACTIVITIES:

GSA - FSS
HHS-NIH
Commerce - NBS

PREPARING ACTIVITY:

Army - GL

Project No. 6670-0153

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:38Z
Check the source to verify that this is the current version before use.