INCH-POUND

A-A-50060A
October 26, 1992
SUPERSEDING
A-A-50060
September 23, 1983

COMMERCIAL ITEM DESCRIPTION

BALANCES, TRIP, LABORATORY

The General Services Administration has authorized the use of this Commercial Item Description in preference to MIL-B-43714.

1. CLASSIFICATION

1.1 Classification. This Commercial Item Description covers two types of laboratory trip balances, as specified (see 5.2):

Type I - Single beam, dual scale, metric and avoirdupois, capacity of 2000 grams/5.1 pounds.

Type II - Double beam metric, capacity of 2000 grams.

2. SALIENT CHARACTERISTICS

2.1 Design and construction. The balance shall be of the beam type provided with a durable, friction free and self adjusting damping system. It shall consist of a base, main beam, one or more weightbeams with a poise on each, weighing surfaces, suitable knife edges, bearings, and an easily read pointer. The weighing surfaces shall be approximately 6 inch diameter stainless steel plates, polished or brushed, flat, and free of obstructions. The platforms shall be readily interchangeable without the necessity of replacing any other part. The poises shall be of the center reading, sliding type. Bearings shall be made of agate, and the knife edges shall be made of hardened steel, accurately ground. Bearing blocks shall be self-aligning to

Beneficial comments, recommendations, additions, deletions, clarifications, etc., and any data which may improve this document should be sent to: Commander, U.S. Army Natick Research, Development and Engineering Center, ATTN: SATNC-UXT, Natick, MA 01760-5017.

AMSC N/A

FSC 6670

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:38Z
Check the source to verify that this is the current version before use.

ensure continuous contact along the knife edges. The balance shall be provided with a suitable dust cover of sufficient size to completely cover it. Each balance shall be furnished with two copies of instructions for installation, operation, and maintenance and a list of component parts.

2.2 Sensitivity and beam graduations. The type I balance shall have a zero load sensitivity of 0.1 grams/0.0035 ounces (avdp.) per graduation, and a full load sensitivity of 0.2 grams/0.0070 ounces (avdp.) per graduation, and a capacity of 2000 grams/5.1 pounds. The beam shall be graduated in 0.2 grams and .01 ounce divisions. The type II balance shall have a full load sensitivity of 0.2 grams, and a capacity of 2000 grams. The upper beam shall be graduated from 0 to 10 grams in 0.1 gram divisions while the lower beam shall be graduated from 0 to 200 grams in 10 gram divisions.

2.3 Identification plate. The balance shall have an identification plate permanently and legibly marked with the manufacturer's name, trademark, model number, contract number, and national stock number.

3. QUALITY ASSURANCE

3.1 Certification. The contractor shall certify, and maintain substantiating evidence, that the product offered meets the salient characteristics and requirements of this Commercial Item Description, and that the product conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

3.2 Metric product. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using the inch-pound units, provided they fall within specified tolerances using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this Commercial Item Description are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

3.3 Regulatory requirements. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practicable.

4. PACKAGING

4.1 Overseas packaging. Movable components of the balance shall be secured in a manner to prevent movement or damage during shipment and storage. The balance shall be covered with a plastic dust cover. Each balance shall be

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:38Z
Check the source to verify that this is the current version before use.

packaged in a fiberboard box conforming to style RSC-L, grade 275 of PPP-B-636. The balance shall be cushioned and blocked as needed to prevent damage during shipment and storage. The box shall be closed in accordance with method II as specified in the appendix of PPP-B-636.

4.2 Overseas packing. Four balances shall be packed in a fiberboard box conforming to style RSC, grade V3c of PPP-B-636. The box shall be closed, waterproofed and reinforced in accordance with the appendix of PPP-B-636.

4.3 Overseas palletization. Boxes shall be palletized on a 4-way entry pallet in accordance with load type Ia of MIL-STD-147. Each prepared load shall be bonded with primary and secondary straps in accordance with bonding means C and D, or film bonding means F or G. Pallet pattern shall be in accordance with the appendix of MIL-STD-147.

4.4 Overseas marking. Marking shall be in accordance with MIL-STD-129 or FED-STD-123, as applicable.

4.5 Domestic packaging, packing, palletization and marking. Packaging, packing, palletization and marking shall be in accordance with ASTM D 3951.

5. NOTES

5.1 Intended use. The balances are intended to be accurate weighing instruments in laboratories and pharmacies.

5.2 Acquisition requirements. Acquisition documents must specify the following:

a. Title, number, and date of this document.
b. Type required (see 1.1).

5.3 Sources of documents.

5.3.1 Sources of Government documents. Copies of military and federal documents are available from:

Standardization Documents Order Desk
Building 4D
700 Robbins Avenue
Philadelphia, PA 19111-5094

5.3.2 Sources for non-Government association documents.

ASTM D 3951 - Standard Practice for Commercial Packaging, is available from:

The American Society for Testing and Materials (ASTM)
1916 Race Street
Philadelphia, PA 19103-1187

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:38Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians

Army - GL
Air Force - 03

Review Activities

Army - MD, ME
Air Force - 82, 99
DLA - GS

User Activities

Navy - MS, YD

CIVIL AGENCY COORDINATING ACTIVITY:

GSA - FSS

PREPARING ACTIVITY:

Army - GL

Project 6670-0192

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:38Z
Check the source to verify that this is the current version before use.