INCH-POUND

MIL-D-16153G
22 February 1991
SUPERSEDING
MIL-D-16153F
28 June 1961
(See 6 8)

MILITARY SPECIFICATION

DETECTOR, RADIAC, DT-60( )/PD

This specification is for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of nonself-indicating dosimeter designated as Radiac Detector DT-60( )/PD which measures X- and gamma radiation over the range from 0 to 1000 roentgens (R) It is a personnel dosimeter which is worn about the neck like a pendant. The detecting element is a radio-photoluminescent glass. The amount of dosage received by the detector is indicated in roentgens from 0 to 1000 by the Radiac Computer-Indicator CP-95( )/PD.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications, standards, and handbooks. The following specifications, standards, and handbooks form a part of this document to the extent specified herein. Unless otherwise specified, the issues of these documents are those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation (see 6.2)

SPECIFICATIONS

MILITARY

MIL-E-16400 - Electronic, Interior Communication and Navigation Equipment, Naval Ship and Shore. General Specification For.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commander, Naval Sea Systems Command, SEA 55Z3, Department of the Navy, Washington, DC 20362-5101 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6665

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

MILITARY - Continued

MIL-E-17555 - Electronic and Electrical Equipment, Accessories, and Provisioned Items (Repair Parts): Packaging Of.

STANDARDS

MILITARY

MIL-STD-130 - Identification Marking of U.S. Military Property.
MIL-STD-242 - Electronic Equipment Parts Selected Standards.

(Unless otherwise indicated, copies of federal and military specifications, standards, and handbooks are available from the Standardization Document Order Desk, Bldg. 4D, 700 Robbins Ave., Philadelphia, PA 19111-5094.)

2.1.2 Other Government documents, drawings, and publications. The following other Government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

PUBLICATION

National Institute of Standards and Technology (NIST)

NBS SP250-16 -

(Copies are available from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402-0001.)

2.2 Order of precedence. In the event of a conflict between the text of this document and the references cited herein (except for related associated detail specifications, specification sheets or MS standards), the text of this document takes precedence. Nothing in this document, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

3. REQUIREMENTS

3 1 First article. When specified (see 6.2), a sample shall be subjected to first article inspection (see 6.4) in accordance with 4.3.

3.2 Design objective. Reliability of operation shall be stressed throughout the design and manufacture of the equipment.

3.2.1 Reliability. The equipment shall be subjected to continuous use for long periods under the varied and severe conditions of Military service without overhaul and with little maintenance. If it is to meet these conditions it is imperative that reliability of operation be considered of prime importance in the design and manufacture of the equipment.

3.3 Selection of parts. Parts shall be selected in accordance with MIL-E-16400. MIL-STD-242 shall be the basic source for selection of parts.

3.3.1 Interchangeability. All component parts of each dosimeter shall be interchangeable with every other dosimeter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

3 4 Material. Material used shall be of a high quality commercial grade available and shall be subject to the approval of the bureau or agency concerned (see 3 5.1 through 3 5.6.5).

3 5 Equipment composition The detector shall consist of a case, cover, gasket, sensitive glass element, glass retainer, cushioning pad, and lead shields

3.5 1 Plastic. Plastic used in the DT-60( )/PD dosimeter housing, except for the gasket, shall be black ethyl cellulose. Hercules flakes, formula 427-J-S2 or substitute approved by the bureau or agency concerned. The glass retainer assembly shall be black linear polyethylene In the molding of plastic parts, silicone mold release only shall be employed and only virgin material used.

3.5.1.1 Conditioning. Prior to assembly of plastic parts they shall be stored in a ventilated chamber at 60 degrees Celsius (°C) temperature for a period of not less than 48 hours.

3 5.2 Metal. Metal parts in the glass retainer assembly shall be beryllium copper black nickel plate.

3 5 3 Housing. The housing shall have no deleterious effect when in contact with the skin of tne wearer.

3.5.4 Gasket material. Gasket material shall be irradiated polyethylene, Union Carbide type DND-4400 with 2 percent carbon black or equal, and shall be mechanically secured without the use of adhesives.

3.5.5 Shield. The shields shall be of such a configuration so as to allow the Radiac Detector DT-60( )/PD to meet the requirements as specified in 3.9.2 and 3.9.3.1

3.5.6 Radio-photoluminescent glass. The sensitive element shall be radio-photoluminescent glass. This glass emits fluorescent light in the 4,200 to 7,000 angstrom region under near ultraviolet (uv) irradiation after prior exposure to X or gamma radiation. The intensity of the fluorescent light is proportional to the amount of X- or gamma radiation and the intensity of near ultraviolet light to which the glass has been exposed. The intensity of the fluorescent light is detected and indicated as a roentgen (R) dose by the Radiac Computer-Indicator CP-95( )/PD. The chemical composition of the glass shall be eight parts by weight silver metaphosphate, 25 parts by weight barium metaphosphate, 25 parts by weight potassium metaphosphate, and 50 parts by weight aluminum metaphosphate. It is required that choice of raw materials and processing be so that the glass will fulfill all requirements of this specification.

3.5.6.1 Dimensional tolerances. The sensitive element shall be 0.750 ± 0.025 inch by 0.750 ± 0.25 inch by 0.180 ± 0.020 inch. The entrance face of the sensitive element shall be flat within 0.005 inch. The opposite major surface shall be flat within 0.025 inch and parallel to the entrance face within 0.010 inch. The glass sensitive element shall cover a square of 0.715 by 0.715 inch, using the exit edge as one side of the square. The entrance face shall be

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

perpendicular to the exit face within plus or minus 2 5 degrees The glass sensitive element shall cover a square of 0 715 by 0.160 inch using the entrance edge as one side of the square. The entrance face shall be specified as the major noncoated surface of the parallelpiped The exit face is the minor noncoated surface of the parallelpiped

3 5 6.2 Finish The entrance face shall have a smooth vitreous or polished surface The exit face shall be ground and fine polished All other surfaces shall be vitreous or diamond cut and broken Ground and saw cut surfaces are not acceptable.

3 5 6.3 Coating All surfaces of the sensitive element except the entrance face and exit edge shall be covered by an inorganic, nonreflecting, continuous fitted glass coating, opaque to visible and UV light. The coating shall be composed of lead oxide, silicate type glass, containing as colorant agents the ions of copper, iron, cobalt and chromium only. Components with a luminescence decay constant materially different from the neutral silver shall be excluded. The expansion coefficient shall be compatible with the base glass over the temperature range from minus 55 to plus 85 °C

3 5 6 4 Placement of glass. The glass rectifier assembly shall secure the glass so that any movement of the glass in the assembly will still permit the dosimeter to meet the requirements of 3.9.

3.5.6.5 Outer coating. All surfaces of the glass shall be covered by an optically clear water proof coating. After a storage period as specified in 4.3.2.10, the prestorage reading shall not have changed by more than 10 R.

3.6 Design The design shall be so that the DT-60( )/PD will meet all requirements of this specification when read by the CP-95( )/PD

3 6.1 Size. The DT-60( )/PD size shall be held to a minimum so that it may be conveniently worn about the neck. The overall thickness of the dosimeter without the housing shall be not less than 0.260 inch nor greater than 0.340 inch. The area of the sensitive glass element exposed to UV irradiation shall be not less than a square of 0.650 by 0.650 inch with a center at 0.062 inch from a line through the center of the locating holes toward the exit face. The exit edge shall be paralled in the centerline of the locating holes

3.6.2 Weight. The weight of the DT-60( )/PD shall be not greater than 3 ounces.

3.6.3 Enclosure. The glass shall be completely enclosed within the housing.

3.6.3.1 Accessibility. The detector shall be tamperproof insofar as is consistent with design; that is, it shall not be easily opened by the wearer. A special tool may be used for opening the case. This tool shall be subject to approval by the bureau or agency concerned.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

3 6 3 2 Immersion The detector shall meet the requirements of this specification after being subjected to a hydrostatic pressure of 2 pounds per square inch (lb/in^2) for a period of 1 hour in sea water, with tightly closed cover There shall be no leakage of water into the case as a result of immersion.

3 6.4 Finish. The exterior finish of the detector shall be smooth and void of flaws and crevices insofar as is consistent with design. The finish shall not flake, or peel off, thereby exposing material which might be subject to corrosion.

3.6.5 Adhesives. The use of adhesives is not permitted.

3 7 Temperature.

3.7.1 Storage. The detector shall meet the requirements of this specification after having been stored at temperatures over the range from minus 55 to plus 85 °C.

3 7.2 Exposure temperature. The detector shall be capable of detecting X- and gamma radiation within the accuracy specified in 3.9.3 when exposed at 25 plus or minus 1 °C and over a temperature range from minus 7 to plus 60 °C upon application of a linear temperature-coefficient correction.

3.7.3 Reading temperature. The detector shall indicate dosage received within the accuracy specified in 3.9.3 when read 16 hours after exposure at a temperature of 25 plus or minus 1 °C and over a temperature range of minus 7 to plus 60 °C upon application of a linear temperature-coefficient correction.

3.8 Impact. The detector shall not be damaged by being dropped on a steel floor six times from a height of 6 feet on any side or face.

3.9 Operation. The following shall be determined, using CP-95( )/PD.

3.9.1 Rings. The detector shall cover the range from 0 to 1000 R.

3.9.2 Minimum dose. The minimum detectable dose shall be not greater than 10 R.

3 9.3 Accuracy. The accuracy of the dosage indication over the entire range of detection shall meet the following:

3.9.3.1 For doses up to 1000 R the detector shall indicate the dose received within plus or minus 20 percent of the true roentgen dose, or 10 R whichever is larger, when exposed to X- or gamma radiation over the energy range 80 thousand electron volts (KeV) to 1.3 million electron volts (MeV). Over the range 80 to 46 KeV, the dosimeter response shall decrease smoothly; at 46 KeV and at all X- and gamma radiation energies below 46 KeV the response shall be not greater than 50 percent of true dose. All energies refer to effective energies of X-ray beams filtered as described in NIST NBS Special Publication 250-16. Table 1-1 of NIST NBS Special Publication 250-16 shows the filtering used for the appropriate

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

heavily filtered beam codes. The radiation shall fall onto either major face of the detector at normal incidence The units shall be read in a CP-95( )/PD, calibrated to within plus or minus 5 percent of standards approved by the bureau or agency concerned.

3 9 3 2 Directionality. For radiation incident on the dosimeter at any angle of incidence other than the normal the reading due to directionality shall not deviate by greater than plus or minus 10 percent from that of the normal.

3.9.3.3 Persistence. The fluorescent response to ultra-violet light shall be persistent when measured, starting 16 hours after exposure It shall be possible to read the detector for a period up to 45 days after exposure and the reading shall not change by more than plus or minus 8 percent.

3.9.3.4 Additivity. The accuracy within the limits of 3.9.3.1 and 3.9.3.2, of any reading shall be the same whether the dose is received in a single continuous exposure or in several discrete exposure-obtained at intervals varying from a few hours up to 45 days.

3.9.3.5 Integration rate. The detector shall integrate radiation intensities up to 10,000 R per hour and still indicate dosage within the accuracy specified in 3.9.3.1.

3.9.3.6 Predose. The DT-60( )/PD shall have, when exposed to near ultra-violet irradiation in the CP-95( )/PD, a luminescence that shall be equivalent to 37 $\pm$ 5 R. This final value of predose luminescence must be achieved, by irradiating with Cobalt 60 or Cesium 137 gamma radiation, dosimeters which have a maximum predose luminescence of 25 R. This latter value of predose shall be achieved by adequate purity of raw materials and not adjusted by the addition of substances which when added, produce fluorescence After a 30-day storage in accordance with 4 3 2.7, the predose value shall not have changed by more than plus or minus 3 R.

3.9 3.7 All fluorescent components with decay constant and spectral response materially different from the neutral silver shall be excluded. The decay time of the fluorescent component of an undosed dosimeter shall be not greater than 1000 microseconds, measured as specified in 4.3.2.8.

3.10 Nomenclature and identification plates. Nomenclature identification shall appear on each detector housing and glass retainer assembly. Assemblies and parts shall be marked for identification in accordance with MIL-STD-130.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements (examinations and tests) as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to ensure supplies and services conform to prescribed requirements

4 1 1 Responsibility for compliance All items shall meet all requirements of sections 3 and 5 The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of ensuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract Sampling inspection, as part of the manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however, this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to accept defective material.

4.2 Classification of inspections. The inspection requirements specified herein are classified as follows:

(a) First article inspection (see 4.3).
(b) Quality conformance inspection (see 4.4).

4.3 First article inspection.

4.3.1 Preinspection examination. Prior to submission of the first article models the contractor shall ascertain that the detector is free from defects of material and workmanship and is in a satisfactory operating condition.

4.3.2 Test methods. The contractor shall provide 100 units for the first article testing. First article inspection shall be made at a Government laboratory designated by the bureau or agency concerned on the first article models. First article inspection shall consist of any examination and tests necessary to determine compliance with this specification and shall include the following:

4.3.2.1 General examination. The detector shall be examined to determine compliance with this specification including the following:

(a) Assembly, dimensions, and fit in CP-95( )/PD.
(b) Type and character of materials, parts, and finish, including treatment for prevention of corrosion.

4.3.2.2 Operating tests. Detectors shall be exposed to X-ray and gamma ray radiation at 25 ± 1 °C and read at 25 ± 1 °C no sooner than 16 hours after completion of the exposure to determine conformance with 3.9.

4.3.2.2.1 Testing temperature. The detector shall operate in accordance with this specification providing dosage temperature is within the range 25 ± 1 °C and reading temperature is within the range 25 ± 1 °C. Upon application of a linear temperature-coefficient correction the detector shall operate in accordance with this specification with the dosage temperature and reading temperature varied over a range from minus 7 to plus 60 °C.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

4.3.2.3 Water pressure. The case shall be opened and closed five times after which the detector shall be subjected to a water pressure test to determine conformance with 3 6.3.2.

4 3.2.4 Impact Detectors shall be dropped on a steel floor six times from a height of 6 feet. They shall then be examined visually to determine whether they have been damaged and they shall also be subjected to the operating tests of 4.3.2 2.

4 3.2.5 Altitude Detectors shall be subjected to an external air pressure of 3.44 inches of mercury or less at a temperature of minus 55 °C for a 24-hour period. They shall then be subjected to the operating tests specified in 4.3.2.2.

4.3.2.6 Persistence test. Detectors which have been subjected in the operating tests of 4.3.2.2 shall be stored for a period of 45 days and then measured in a CP-95( )/PD. The measurements taken at the end of the 45-day period shall be compared with the measurements taken at the beginning to determine compliance with 3.9.3 3. The cycling shall be accomplished as described in figure 1. The fifteen day cycle shall be repeated three times to meet the 45 day requirement. Temperature changes shall be at a rate of not less than 5 degrees Celsius/minute and humidity changes shall be at a rate of not less than 5 percent RH/minute.

4.3.2.7 Predose test. Detectors shall be stored for a period of 30 days at a temperature of 50 ± 5 °C and a relative humidity of 25 ± 5 percent. They shall then be tested to determine conformance to 3.9.3.6.

4.3.2.8 Decay time, fluorescent components. A DT-60( )/PD shall be inserted into a CP-95( )/PD modified for square wave 60-cycle pulsed excitation. The output of the multiplier shall be observed on the oscilloscope. The decay time measured from 90 percent to 10 percent of the trailing edge of the waveform shall be less than 1000 microseconds, when measured across the photomultiplier output.

4.3.2.9 After acceptance of first article equipments the manufacturer shall not change the composition of the glass or any other component nor shall he change the method of shaping, finishing or otherwise preparing the glass.

4.3.2.10 Humidity test. Detectors that have been subjected to the operating tests of 4.3.2.2 shall be read and subjected to a relative humidity of 80 percent and a temperature of 90 degrees Fahrenheit (°F) for one hour with the cases open. They shall then be stored for 7 days at 77 °F with the cases closed. At the end of the 7 days, they shall be read. The readings taken before and after this storage test shall conform to 3.5.6.5.

4.4 Quality conformance inspection. Quality conformance inspection shall consist of the group A and group B examination and tests and any other tests necessary to determine compliance with this specification.

4.5 Quality conformance tests. The following paragraphs outline the general procedures to be used in quality conformance testing (see 6.3 and appendix). The procedure in performing these tests may be changed upon mutual agreement by the

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

contractor and the bureau or agency concerned. Group B and C tests are principally radiation tests which will make the detector of no value for reuse. These tests will be made on DT-60( )/PD in excess of the quantity ordered by the Government.

4 5 1 Sampling for group A tests. A random sample of detectors shall be selected from each days production, and subjected to each of the group A tests listed in 4.5.4.1 with acceptance based on the sampling requirements of table I

TABLE I. Sampling for group A tests.

Number of detectors in daily production	Number of detectors in sample
15 and under	7
16 to 40	10
41 to 110	15
111 to 300	25
301 to 500	35
501 to 800	50
801 to 1300	75
1301 and over	110

4.5.2 Sampling for group B test. A random sample of detectors shall be selected from each group of detectors presented for shipment in accordance with table II. All detectors selected shall be submitted to each of the group B tests listed in 4.5.4.2.

TABLE II. Sampling for group B tests.

Number of detectors for shipment	Number of detectors in sample
15 and under	3
16 to 40	5
41 to 110	7
111 to 300	10
301 to 500	15
501 to 800	25
801 to 1300	35
1301 and over	50

4.5.3 Sampling for group C tests. Eighty detectors shall be selected from each 10,000 units of production or from each months production, whichever is larger and shall be subjected to the group C tests specified in 4.5.4.3, five to each test, except that thirty each shall be used for the dependence and temperature operation tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

4.5.4 Test methods.

4 5 4 1 Group A tests. Each of the sample detectors selected in accordance with table I shall be subjected to each of the tests specified in 4 3 2.1. The results of each test shall be compared with specification requirements Failure to conform to the specification requirements for any group A test shall be counted as a defect and the detector shall be rejected

4 5 4.2 Group B tests. Group B tests shall be performed or witnessed by the inspector, and the results of each test compared with the specification requirements. Failure to conform to any requirement shall be counted as a defect and the detector shall be rejected

Group B tests	Reference
Water pressure	4.3.2.3
Impact test	4.3.2.4
Operating tests:	4.3.2.2
Minimum dose	3.9.2
Accuracy (Co^{60} test only)	3.9.3.1

4 5.4 3 Group C tests. Failure to conform to the specification requirements for any group C test shall cause the contractor to correct the cause of failure on future production units and to repair the deficiency in any detectors not yet shipped.

Group C tests	Reference
Temperature operation tests:	4.3 2.2.1
Exposure temperature	3.7.2
Reading temperature	3 7 3
Operation tests:	4.3.2.2
Additivity	3.9.3.4
Persistence	3.9.3.3
	4.3.2.6
Energy dependence	3.9 3.1
Directionality	3 9.3.2
Altitude	4.3.2.5

4 6 Inspection of packaging. Sample packages and packs, and the inspection of the preservation, packing and marking for shipment, storage and stowage shall be in accordance with the requirements of section 5 and the documents specified therein.

5. PACKAGING

5.1 Preservation, packing, and marking. The dosimeters shall be preserved by level A or C; packed by level A, B, or C, and marked as specified (see 6.2) in accordance with MIL-E-17555.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

6 1 Intended use. The dosimeter covered by this specification is intended for personal use in detecting cumulative dosage of X- and gamma radiation

6 2 Acquisition requirements. Acquisition documents must specify the following.

(a) Title, number, and date of this specification.
(b) Issue of DODISS to be cited in the solicitation, and if required, the specific issue of individual documents referenced (see 2.1.1).
(c) When first article inspection is required (see 3.1).
(d) Preservation, packing, and marking is required (see 5.1).

6.3 Consideration of data requirements. The following data requirements should be considered when this specification is applied on a contract. The applicable Data Item Descriptions (DIDs) should be reviewed in conjunction with the specific acquisition to ensure that only essential data are requested/provided and that the DIDs are tailored to reflect the requirements of the specific acquisition. To ensure correct contractual application of the data requirements, a Contract Data Requirements List (DD Form 1423) must be prepared to obtain the data, except where DOD FAR Supplement 27 475-1 exempts the requirement for a DD Form 1423.

Reference paragraph	DID number	DID title	Suggested tailoring
4 5 and appendix	DI-MISC-80678	Certification/ data report	--------

The above DIDs were those cleared as of the date of this specification The curent issue of DOD 5010.12-L, Acquisition Management Systems and Data Requirements Control List (AMSDL), must be researched to ensure that only current, cleared DIDs are cited on the DD Form 1423.

6.4 First article. When first article inspection is required, the contracting officer should provide specific guidance to offerors whether the items should be a preproduction sample, a first article sample, a first production item, a sample selected from the first lot production items, a standard production item from the contractor's current inventory (see 3.1), and the number of items to be tested as specified in 4.3. The contracting officer should also include specific instructions in acquisition documents regarding arrangements for examinations, approval of first article test results, and disposition of first articles. Invitations for bids should provide that the Government reserves the right to waive the requirement for samples for first article inspection to those bidders offering a product which has been previously acquired or tested by the Government,

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

and that bidders offering such products, who wish to rely on such production or test, must furnish evidence with the bid that prior Government approval is presently appropriate for the pending contract Bidders should not submit alternate bids unless specifically requested to do so in the solicitation.

6 5 Lot acceptance and rejection criteria Lot acceptance and rejection for group A and group B tests should be as specified in tables III and IV If the number of nonconforming detectors in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE III. Group A tests - AQL (approx) = 1 5 percent defective

Number of detectors in daily production	Number of detectors in sample	Number of nonconformance to group A tests	
		Acceptance number	Rejection number
15 and under	7	0	1
16 to 40	10	0	1
41 to 110	15	0	1
111 to 300	25	1	2
301 to 500	35	1	2
501 to 800	50	2	3
801 to 1300	75	3	4
1301 and over	110	4	5

TABLE IV. Group B tests.

Number of detectors for shipment	Number of detectors in sample	Number of nonconformance to group B tests	
		Acceptance number	Rejection number
15 and under	3	0	1
16 to 40	5	0	1
41 to 110	7	0	1
111 to 300	10	0	1
301 to 500	15	1	2
501 to 800	25	1	2
801 to 1300	35	2	3
1301 and over	50	3	4

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

6.6 Nomenclature identification. Requests for nomenclature (see 3.10) should include descriptions in accordance with FED-STD-5 plus 2 copies of dimensional outline drawings.

6 7 Subject term (key word) listing

Dosimeter
Gamma radiation
Radio-photoluminescence
X-radiation

6.8 Changes from previous issue. Marginal notations are not used in this revision to identify changes with respect to the previous issue due to the extensiveness of the changes.

Custodians:
Army - EA
Navy - SH
Air Force - 99

Review activity:
DLA - DS

Preparing activity:
Navy - SH
(Project 6665-N585)

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

APPENDIX

CERTIFICATION/DATA REPORT TECHNICAL CONTENT REQUIREMENTS

10. SCOPE

10 1 Scope This appendix covers information that shall be included in the certification/data report when specified in the contract or order This appendix is mandatory only when data item description DI-MISC-80678 is cited on the DD Form 1423.

20. APPLICABLE DOCUMENTS

This section is not applicable to this appendix.

30. CERTIFICATION CONTENT

30 1 Certified test data sheets. The contractor shall furnish certified test data sheets on samples selected from each batch of glass on the basis of one sample per 7-1/2 pounds, showing predose readings and response to a dose of 100 R from a Co^{60} source for each batch of glass. No equipment shall be accepted which shows a predose reading of more than 42 RT1 or less than 32 RT1 or indicates a dose of more than 120 R or less than 80 R when subjected to dose of 100 R from a Co^{60} source. A batch is defined as one mixture of dry ingredients.

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER MIL-D-16153G	2. DOCUMENT DATE (YYMMDD) 91/02/22

3. DOCUMENT TITLE
DETECTOR,RADIAC, DT-60( )/PD

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON

7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Technical Point of Contact (TOPC): Mr. Kevin Predmore (SEA 04R5) PLEASE ADDRESS ALL CORRESPONDENCE AS FOLLOWS:	b. TELEPHONE (Include Area Code) (1) Commercial TOPC: 703-602-2753 (2) AUTOVON 8-332-2753
c. ADDRESS (Include Zip Code) Commander, Naval Sea Systems Command Department of the Navy (SEA 5523)	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T08:21Z
Check the source to verify that this is the current version before use.