MIL-E-1/973B
AMENDMENT 2
15 April 1976
SUPERSEDING
AMENDMENT 1
13 August 1971

MILITARY SPECIFICATION SHEET

ELECTRON TUBE, RADIATION COUNTER
TYPE 5980

This amendment forms a part of Military Specification Sheet MIL-E-1/973B, dated 2 December 1970, and is approved for use by all Departments and Agencies of the Department of Defense.

PAGE 2

*Quality conformance inspection, part 3, Periodic-check tests: Delete "Periodic-check tests" title heading; and for all tests listed (four tests), under conditions column, add reference to new note 14.

PAGE 4

*Add new note 14 as follows:

"14. This test shall be performed during the initial production and once each succeeding 12-calendar months in which there is production. A regular double sampling plan shall be used, with the first sample of three tubes with an acceptance number of zero, and a second sample of three tubes with a combined acceptance number of one. In the event of failure, the test will be made as a part of quality conformance inspection, part 2, code level D, with an AQL of 6.5. The regular "12-calendar month" double sampling plan shall be reinstated after three consecutive samples have been accepted."

PAGE 5

*FIGURE 1, Dimensions table, under Quality conformance inspection, part 3: Delete "(periodic-check)" and substitute "(See note b)."

*NOTE: Delete "NOTE" and substitute "NOTES"; add "a" to present note and add new note "b" as follows:

"b. Dimensions shall be checked during the initial production and once each succeeding 12-calendar months in which there is production. A regular double sampling plan shall be used, with the first sample of three tubes with an acceptance number of zero, and a second sample of three tubes with a combined acceptance number of one. In the event of failure, the test will be made as a part of quality conformance inspection, part 2, code level D, with an AQL of 6.5. The regular "12-calendar month" double sampling plan shall be reinstated after three consecutive samples have been accepted."

PAGE 7

FIGURE 3. Delete "(Hz)" following TUBE RESPONSE and substitute "(Counts per second)."

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.

*NOTE: The margins of this amendment are marked with asterisk to indicate where changes from the previous amendment were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous amendment.

Custodians:
Army - EL
Navy - EC
Air Force - 85

Review activities:
Air Force - 17
DSA - ES

User activities:
Army - ME, WC
Navy - AS, OS, MC, CG
Air Force - 11

Preparing activity:
Navy - EC

Agent:
DSA - ES

(Project 5960-3082)

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.