INCH-POUND

MIL-PRF-1/973C
27 September 1999
SUPERSEDING
MIL-E-1/973B
2 December 1970

PERFORMANCE SPECIFICATION SHEET

ELECTRON TUBE, RADIATION COUNTER
TYPES 5980 AND 5980A

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The requirements for acquiring the electron tube described herein shall consist of this document and the latest issue of MIL-PRF-1.

DESCRIPTION: Halogen-filled, low sensitivity, self-quenching for detection of gamma radiation.

See figures 1 and 1A.

Mounting position: Any.

Weight: .22 ounce, nominal.

ABSOLUTE RATINGS:

Parameter: Unit:	Ebb V dc	TA °C	Rp Meg
Maximum: Minimum:	725 675	75 -40	--- ---
Test conditions:	700	---	1.0

See footnotes at the end of table I.

GENERAL: Qualification - Required. 1/

Marking. 2/

Response characteristics. 3/

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.

TABLE I. Testing and inspection.

Inspection	Method	Notes	Conditions	Acceptance Level 15/	Symbol	Limits		Unit
						Min	Max	
Conformance inspection, part 1								
Background, contamination, and photosensitivity	6201	5/ 6/ 10/	t = 2 minutes	0.65	N/t	---	10	Npm
Starting voltage	6211	10/	Pulse amplitude = 1 volt; Nps = 200 (max); R2 = 1 Meg; R1 = C1 = C2 = 0; C3 = 0.01 µF	0.65	Es	---	640	V dc
Relative plateau slope	6216	12/	Voltage range = 660 to 760 V dc	0.65	Ps	---	0.2	%/Vdc
End-of-plateau voltage	6216	12/	Ps = 0.3 %/V dc	0.65	Ee	780	---	V dc
Response count rate and current (gamma)	6221	7/ 10/	t = 2 minutes	0.65	N/t	75	125	Nps
Conformance inspection, part 2								
Response count rate and current	6221	10/	Gamma (radium); rate = 500 mr/hr; Ebb = 710 V dc	---	Ib	---	10.0	µA dc
Pulse amplitude (1)	6226	10/	Ebb = 675 V dc; Nps = 200 (max); R1 = 0.9 Meg; R2 = 0.1 Meg; C1 = 330 pF; C2 = 3,000 pF; C3 = 0.01 µF; multiply oscillo-scope reading by 10	---	eo	2.5	---	v
Pulse amplitude (2)	6226	---	Ebb = 725 V dc; pulse amplitude(1)	---	eo	---	65	v

See footnotes at end of table.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.

TABLE I. Testing and inspection - Continued.

Inspection	Method	Notes	Conditions	Acceptance Level 15/	Symbol	Limits		Unit
Conformance inspection, part 3								
Life-test provisions	---	---	Group A; counting rate = 1,000 Nps (min); t = 500 hrs	---	---	---	---	---
Life-test end points	---	9/		---	---	---	---	---
Variable-frequency vibration	1031	4/ 11/ 14/		---	---	---	---	---
Temperature cycling	---	8/ 14/		---	---	---	---	---
Leakage current	6205	14/	Ebb = 500 V dc	---	Lib	---	0.5	µA dc
Shock	1041	4/ 13/ 14/	Hammer angle = 40°	---	---	---	---	---

1/ A minimum of 28 samples shall be submitted for qualification testing. The samples to be submitted for qualification testing will be broken down into the following subgroups:

a. Shelf life only 6

b. Electrical, including life test 6

c. Shock test only 10

d. Variable-frequency vibration 6

2/ Each tube shall have an individual serial number permanently marked on the tube.

3/ The response characteristics, current, and count rate versus field intensity of a typical type 5980 or 5980A radiation counter tube shall be shown on figure 3. In determining tube response, Ebb = 700 V dc and Rp = 1 Meg. Tube current shown as measured by means of a microammeter in series with the tube cathode, and the count rates as determined using a scaler having a resolving time of 0.8 microseconds and a discrimination level of 0.25 volts.

4/ Criterion for passing this test shall be compliance after test of at least 80 percent of the sample tubes with the requirements for:

a. Starting voltage

b. Response count rate (gamma)

c. Pulse amplitude (1) and (2)

d. End-of-plateau voltage

e. Relative plateau slope

5/ In determining tube response (count rate) Ebb = 700 V dc and Rp = 1 Meg. Tube count rates shall be determined using a scaler having a resolving time of 5 microseconds and a discrimination level of 0.25 volts.

6/ The tube shall be exposed to radiation from a General Electric 15-watt germicidal lamp and a General Electric 15-watt fluorescent lamp, or equivalents, with relative positions as shown on figure 2.

7/ Use Government approved standard gamma-test unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.

TABLE I. Testing and inspection - Continued.

8/ With the tube in a field giving 100 ±10 Nps at 700 V dc, the tube response (count rate) at 675, 700, and 725 V dc shall be determined at each of the following temperatures and in the order shown:

a. Room temperature.

b. -40°C.

c. Room temperature.

d. 75°C.

e. Room temperature.

A minimum stabilization time of 30 minutes shall be allowed at each temperature. The absolute count rate at 700 V dc, and the average relative plateau slope as determined by the readings at the 3 voltages (best average position of straight line), shall not differ from the initial readings at room temperature by more than 10 percent and ±0.1 percent/volt, respectively, at any of the four subsequent temperatures shown.

9/ The limits for acceptability shall be as follows:

Background, contamination, and photosensitivity: 12 Npm, maximum.

Starting voltage: Initial value ±10 volts.

Plateau length: 125 volts, minimum.

Relative plateau slope: 0.25 percent/V dc, maximum.

End-of-plateau voltage: 750 volts, minimum.

Response count rate (gamma): Initial value ±10 percent.

Pulse amplitude (1) and (2): Initial limits.

10/ This test to be performed at the conclusion of the holding period.

11/ The tube shall be mounted with its longitudinal axis horizontal and perpendicular to the direction of motion by means of a rigid lamp at the center of the cathode. The time for gradually covering the frequency range shall be 15 minutes.


12/ Method B (test method 6216); the scaler of note 5 may be utilized for this test.

13/ The tube shall be rigidly mounted in positions X and Z by means of a clamp at the center of the cathode. The tube shall be given 10 blows at the X position and 5 at the Z position.

14/ This test shall be performed during the initial production and once each succeeding 12-calendar months in which there is production. A regular double sampling plan shall be used, with the first sample of three tubes with an acceptance number of zero, and a second sample of three tubes with a combined acceptance number of one. In the event of failure, the test will be made as a part of conformance inspection, part 2, code level D, with an acceptance level of 6.5. The regular "12-calendar month" double sampling plan shall be reinstated after three consecutive samples have been accepted.

15/ This specification sheet utilizes an accept on zero (c = 0) sampling plan in accordance with MIL-PRF-1, Table III.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.




Ltr	Dimensions (see note 2)			
	Inches		Millimeters	
	Min	Max	Min	Max
A	---	3.500	---	88.90
B	.825	.835	20.96	21.21
C	.058	.068	1.47	1.73
D	.151	.161	3.84	4.09
E	1.344	1.468	34.14	37.29
F	.450	.460	11.43	11.68
G	.442	.452	11.23	11.48
H	.558	.568	14.17	14.43
J	.816	.826	20.73	20.98
K	.151	.161	3.84	4.09
L	.296	.306	7.52	7.77

NOTES
1. Spray with semi-gloss black enamel after pretest.
2. This test shall be performed during the initial production and once each succeeding 12-calendar months in which there is production. A regular double sampling plan shall be used, with the first sample of three tubes with an acceptance number of zero, and a second sample of three tubes with a combined acceptance number of one. In the event of failure, the test will be made as a part of conformance inspection, part 2, code level D, with an acceptance level of 6.5. The regular "12-calendar month" double sampling plan shall be reinstated after three consecutive samples have been accepted.

FIGURE 1. Counter tube assembly (type 5980).


Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.



Ltr	Dimensions (see figure 1, note 2)			
	Inches		Millimeters	
	Min	Max	Min	Max
A	1.964	1.974	49.88	50.14
B	.339	.349	8.61	8.86
C	.288	.290	7.32	7.37
D	.245	.250	6.22	6.35
E	1.344	1.468	34.14	37.29
F	.390	.410	9.91	10.41
G	.018	.O22	0.46	0.56
H	.250	.260	6.35	6.60

FIGURE 1A. Halogen quenched GM counter tube (type 5980A).

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.



Ltr	Dimensions			
	Inches		Millimeters	
	Min	Max	Min	Max
A	.45	.55	11.4	14.0
B	---	1.00	---	25.4
C	---	1.00	---	25.4

NOTE: Center of sensitive volume of tube and centers of lamps lie in plane of paper.

FIGURE 2. Position of tube for photosensitivity test.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.




FIGURE 3. Response characteristics of typical type 5980 or 5980A counter tube.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.

Custodians:
Army - CR
Navy - EC
Air Force - 11
DLA - CC

Review activities:
Army – CR4, MI
Navy - AS, CG, MC, OS
Air Force – 17, 99

Preparing activity:
DLA - CC

(Project 5960-3559)

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T13:55Z
Check the source to verify that this is the current version before use.