

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

REVIEW ACTIVITIES: OTHER - NSA
USER ACTIVITIES: ARMY - ME

FED. SUP CLASS
6640



HIGH FORM



LOW FORM

PART NUMBER	CAPACITY ml	PORE SIZE microns	TYPE OF POROSITY	NOMINAL DIMENSIONS mm	
				DIA OF DISK	OVERALL HEIGHT
*MS36089	30	4-5.5	fine	30	60
*MS36089-2	30	10-15	medium	30	60
*MS36089-3	30	40-60	coarse	30	60
*MS36089-4	50	4-5.5	fine	40	60
*MS36089-5	50	10-15	medium	40	60
*MS36089-6	50	40-60	coarse	40	60
*MS36089-7	30	3.5-5.5	fine	30	50

*Nominal dimensions

MATERIAL: Borosilicate glass in accordance with Federal Specification DD-G-541, Type I, Class A.

DESIGN: Crucibles shall have a fused-in fritted disk. Top rim of the crucible shall be beaded and the bottom rim shall be fire polished.

MS36089-1, through MS36089-6 shall be of the high form design.

MS36089-7 shall be of the low form design.

NOTES:

1. Procurement document for MS36089-7 is Military Specification MIL-C-3659.
2. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
3. For design feature purposes this standard takes precedence over procurement document referenced herein.
4. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 28 June 1961 REVISED (A) 4 June 1969 REVISED (B) 7 June 1973

P.A. MU Other Cust MS 82	TITLE CRUCIBLE, FILTERING, LABORATORY glass, Gooch	MILITARY STANDARD MS36089
Procurement Specification See Note 1	SUPERSEDES:	SHEET 1 OF 1

DD FORM 1 SEP 63 672 (Coordinated)

EDITION OF 1 SEP 57 MAY BE USED.

U. S. GOVERNMENT PRINTING OFFICE: 1969-341-520/A-7328

DATE

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:34Z
Check the source to verify that this is the current version before use.