
REVIEW ACTIVITIES: ARMY - EL, ME
AIR FORCE - 82

USER ACTIVITIES: ARMY - CE

FED. SUP CLASS
6630



PART NUMBER	DROPPER BOTTLE CAPACITY
MS36559-1	

CANCELLED

CANCELLED AFTER 27 MAY 1988, NO SUPER-SEDING STANDARD

...inders,
...ss in
... I, Class A.

...n steel rod.

...

...atural or synthetic rubber.

...ild steel (20 gage) with brazed or welded corners.

...sist of three reagent bottles, three automatic-zero ...ree aspirator-bulb assemblies, one square test bottle, ...asserole, one 10 milliliter graduated cylinder, one 100 milliliter ...raduated cylinder, three dropper bottles, two pipets, six stirring rods, one 1 ounce collapsible tube of stopcock lubricant, and one cabinet.

PURPOSE: For determining the alkalinity, chloride content, and hardness of boiler water.

NOTES:
1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.
2. For design feature purposes, this standard takes precedence over procurement document referenced herein.
3. Referenced documents shall be of the issue in effect on date of invitations for bid.

APPROVED 16 FEB 1984 REVISED (A) 27 MAY 1980

The military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for repetitive use shall be made from this document

P.A. MU	TITLE	MILITARY STANDARD
Other Cust MS 82	WATER TESTING OUTFIT, BOILER	MS 36559
Procurement Specification MIL-W-16642	SUPERSEDES	SHEET 1 OF 1

DD FORM 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED

☆ U. S. GOVERNMENT PRINTING OFFICE: 1988-301-517 44439

DATE

AMSC N/A

Project Number 6630-0471-01

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:43Z
Check the source to verify that this is the current version before use.