A-A-53469
27 April 1988

COMMERCIAL ITEM DESCRIPTION

INTRAVENOUS INJECTION SET

The General Services Administration has authorized the use of this commercial item description as a replacement for Military Specification MIL-I-37990(DLA-DM), dated 25 July 1986, which is cancelled.

This commercial item description covers a secondary intravenous injection set suitable for delivery of IV solutions during patient care.

Salient characteristics:

General. The set shall be a non-vented, sterile, disposable, secondary intravenous injection set. The set shall be compatible with all plastic and vented glass IV solution containers, and be capable of attachment to a primary intravenous injection set.

Material. Tubing and drip chamber shall be fabricated of clear polyvinychloride which meets the USP requirements for Class II plastics. Needle shall be fabricated of type 304 stainless steel in accordance with American Society for Testing and Materials (ASTM) Designation F899.

Style and design. The set shall have an insertion spike capable of piercing an intravenous fluid container (see Note 2), a 15-20 ml flexible drip chamber with a drop-forming drip cannula, flexible tubing, a roller flow-clamp, an intravenous needle and male Luer adapter, removable end protectors, and a suspension hook.

Intravenous needle shall be 18 or 19 gage with Luer hub cavity in accordance with ANSI/HIMA MD70.1. Tubing shall nominally measure 36 inches long and 0.1 inch inner diameter. The suspension hook shall be suitable for lowering the primary intravenous injection set below the level of the secondary set.

The fluid flow to the patient shall be a unique constant rate for each position of the roller within the clamp. The position of the roller in the clamp shall not change when released by the user, nor shall the roller separate from the clamp during typical use.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sterility. The intravenous injection set shall be sterile in accordance with USP or AAMI guidelines. Biological indicator incubation time shall be in accordance with the USP or shall be at least 5 days using procedures in accordance with the FDA Guide For Validation of Biological Indicator Incubation Time.

Pyrogenicity. The intravenous injection set shall be non-pyrogenic in accordance with the USP. In lieu of animal pyrogen testing, USP LAL-Bacterial Endotoxins test is acceptable.

Drop test. The intravenous injection set shall be tested at 72 ± 5° Fahrenheit (F), using water at a rate of 100 ± 5 drops per minute. The number of drops required to deliver 5 milliliters (ml) shall be counted, and the drops per ml calculated. The number of drops per ml shall be 15 ± 3.

Piercing device (spike) durability test. The piercing device shall be immersed for 4 hours in a solution of USP dextrose and sodium chloride. The piercing device shall be removed from the solution and immediately inserted into a standard solution container (See Note 2). The piercing device shall not loosen, become damaged, or permit leakage after insertion.

Security of roller in clamp body. The roller clamp shall be removed from the set together with a piece of tubing approximately one inch longer that the clamp body. The tubing shall be present in the roller clamp in its normal position with a 1/2 inch length extending from each end of the roller clamp. The clamp shall be held verically, with the large end down, over a supporting gap so that the roller, if forced out of the body, is free to fall. A polished steel rod is applied to the top of the roller, so that the roller is able to rotate in the course of the test, and the vertical downward force required to separate the roller from the clamp body is measured. The roller shall withstand a 3.65 kilogram load without being separated from the clamp body. (See Figure 1)

Instructions. Each intravenous injection set shall be provided with complete instructions for use.

Workmanship. The intravenous injection set shall be leakproof and free from defects which detract from its appearance or impair its serviceability.

Unit. Package. (PG). One package containing 50 intravenous injection sets, each set individually packaged sterile, as specified, constitutes one unit. The individual package shall be a sealed, peel open type container capable of maintaining sterility of contents unless package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be Level S-3 and Acceptance Number zero, in accordance with MIL-STD-105. The unit of product for test purposes shall be one intravenous injection set.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE 1: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-153-5669	INTRAVENOUS INJECTION SET, Secondary, With Flexible Drip Chamber, Disposable, Non-Vented, 18 or 19 Gage Needle, 48 or 50s.

NOTE 2: The intravenous injection cassette set shall be compatible with a 250 ml container of Type I, size 2, strength 1, 5 percent USP, Dextrose Injection (NSN 6505-00-975-0604).

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51610

Project No. 6515-4045

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 6, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 8 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4866 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53469	INTRAVENOUS INJECTION SET

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*:
	☐ VENDOR
	☐ USER
3b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* - Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* - Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.