A-A-53475
23 May 1988

COMMERCIAL ITEM DESCRIPTION

PLATINUM FOIL (1 pennyweight)

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification QQ-P-00428C (DSA-DM).

Salient characteristics:

This commercial item description covers one type of platinum foil suitable for dental use.

The foil shall be a minimum of 99.9 percent platinum, when tested for purity as specified below. The foil shall be suitable in all respects for dental inlay or porcelain jacket matrices.

The foil shall be furnished in 1-pennyweight (0.050 troy ounce), wrinkle-free sheets, each 1 1/4 ± 1/8 inches wide and 3 5/8 ± 1/8 inches long. Average thickness of the foil shall be 0.0010 ± 0.0001 inch, when tested as specified below.

Individual sheets of foil shall weigh 0.050 troy ounce ± 0.001 troy ounce. The total weight of twenty (20) sheets of foil shall be at least 1.000 troy ounce, when weighed as specified below. The foil shall comply with the softness test specified below.

Workmanship. The platinum foil shall be free from defects which detract from its appearance or may impair its serviceability.

Unit. Each (EA). One sheet of platinum foil, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of supplies to which such records relate.

Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling.

For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 and 2.5 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one sheet of platinum foil.

For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 at inspection level S-1. AQL shall be 1.0 for thickness and softness tests, but any failure shall be cause for rejection in the purity and weight tests. The unit of product shall be 20 units of issue (20 sheets of platinum foil) for the average weight test, 10 units of issue for the thickness test, and 1 unit of issue for the other tests. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Examination. The platinum foil shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with the following classification of defects listed below. The foil shall be examined for defects including, but not limited to, those listed below.

Defects

Sheet of platinum foil has wrinkles.
Sheet of platinum foil has breaks or tears.
Length or width of sheet of platinum foil not as specified.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with the specification requirements.

Weight. Weights of individual sheets of platinum foil and the total weight of 20 sheets of foil (average weight) shall be determined using a suitable scale or balance accurate to 0.001 troy ounce.

Purity. The purity of the platinum foil shall be determined by a thermoelectric test against pure platinum, such as the method of Burgess and Sale in National Bureau of Standards Bulletin Number 254 (1915). The test shall be conducted with a difference of 1200 degrees Centigrade (^{o}C) between the hot and cold junctions.
The foil shall have a thermoelectric potential not greater than 0.15 millivolt.

Alternately, purity of the platinum foil may be determined by an optical emission spectroanalytical method, provided that approval is received from the contracting officer or procuring activity prior to award of the contract or purchase order.

Thickness. The thickness of the platinum foil shall be determined by measuring a layer of ten sheets of foil with a ratchet micrometer.

Softness. A bend test for softness shall be made on the platinum foil. The sheet of platinum foil shall be cut to approximately 63 millimeters in length and folded around a steel rod 20 millimeters in diameter so that the ends will meet. The end of the rod shall be placed against a smooth glass plate, and the foil permitted to spring open as it stands with the rod in a vertical position on the glass plate. The distance between the ends of the foil shall not be greater than 20 millimeters.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-00-580-2550	PLATINUM FOIL, 1 pennyweight

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2086

LOCATION: NCR DISK #5

48232

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53475	PLATIUM FOIL (1 pennyweight)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-0-26 10:5:202
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.