A-A-53476
23 May 1988

COMMERCIAL ITEM DESCRIPTION

PADDING, CAST, SYNTHETIC, Porous, Non-woven

The General Services Administration has authorized the use of this Commercial Item Description.

Salient characteristics:

This commercial item description covers a synthetic, non-woven bandage suitable for use as padding between an injured extremity and a resin impregnated fiberglass orthopedic cast. The padding shall be furnished in the following nominal sizes, as specified:

Size 1 - 2 inches by 4 yards.
Size 2 - 3 inches by 4 yards.
Size 3 - 4 inches by 4 yards.
Size 4 - 6 inches by 4 yards.

Material. The padding shall be a polyester non-woven bandage and shall be neatly rolled in one continuous length. The polyester fibers shall be entangled to maintain a suitable shape providing cushioning and protection for the skin and bony prominences.

The padding shall be self-adhering and capable of conforming to anatomical contours without slipping or rolling up. The padding shall be porous enough to allow the skin to breathe and reduce the potential for skin maceration or dermatitis. The padding shall have water shedding properties.

Dimensions. The padding shall have a minimum length of 3.9 yards. Size 1 shall have a minimum width of 1.9 inches. Size 2 shall have a minimum width of 2.9 inches. Size 3 shall have a minimum width of 3.9 inches. Size 4 shall have a minimum width of 5.9 inches. Thickness of the padding shall be from 0.060 to 0.070 inches.

Weight. The padding shall weigh from 2.85 to 3.15 ounces per square yard.

Breaking strength. Breaking strength of the padding shall be at least 0.50 lbs. per inch of padding width, when tested as specified below.

AMSC N/A FSC 6510
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Workmanship. Padding shall be clean and ready for use. The padding shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing twenty (20) individual rolls of padding in a moisture resistant sealed container, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0, 1.5 and 2.5 (percent defective) and an inspection level of II and S-2. The unit of product for examination purposes shall be one package of padding.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, at inspection level S-1, AQL 1.0. The unit of product for test purposes shall one package of padding. Where feasible, the same sample unit shall be used for the determination of two or more tests characteristics.

Examination. The padding shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with the following classification of defects listed below. The padding shall be examined for defects including, but not limited to, those listed below.

DEFECTS.

Padding material not as specified.
Roll of padding has breaks or tears.
Padding has other defects affecting function or serviceability.
Dimensions of padding not as specified.
Padding not clean.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Weight. Weight of padding shall be determined using a suitable scale or balance.

Breaking strength. Breaking strength shall be determined using a constant-rate-of-extension tester, such as Instron Model 1122. Crosshead speed shall be 12 inches per minute and gage length 3 inches. The jaws of the clamps used to grip the sample of padding shall be 1 inch (parallel to the application of load) by 3 inches (perpendicular to the application of load) and samples shall be 6 inches long. Before the clamps are tightened on the sample, a uniform 6 ounce tension shall be applied to the sample by means of a tensioning clamp or other means. The clamps shall be tightened and the test shall be run. The test shall be repeated at least 5 times and the average of the results reported.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document.

National Stock Number	Item Identification	Size
6510-01-196-8395	PADDING, CAST, SYNTHETIC, Porous, Non-woven, 2 inch (5.1 cm) by 4 yards (3.7 meters), 20s.	1
6510-01-196-8396	PADDING, CAST, SYNTHETIC, Porous, Non-woven, 3 inch (7.6 cm) by 4 yards (3.7 meters), 20s.	2
6510-01-196-8397	PADDING, CAST, SYNTHETIC, Porous, Non-woven, 4 inch (10.2 cm) by 4 yards (3.7 meters), 20s.	3
6510-01-196-8398	PADDING, CAST, SYNTHETIC, Porous, Non-woven, 6 inch (15.2 cm) by 4 yards (3.7 meters), 20s.	4

NOTE: The Padding is for use with the Bandages covered by the following National Stock Numbers:

6510-01-196-6202	BANDAGE, CASTING, RESIN IMPREGNATED FIBERGLASS, Extra-Fast Setting, 5 inch (12.7 cm) by 4 yards (3.7 meters), 10s.
6510-01-196-6203	BANDAGE, CASTING, RESIN IMPREGNATED FIBERGLASS, Extra-Fast Setting, 4 inch (10.2 cm) by 4 yards (3.7 meters), 10s.
6510-01-196-6204	BANDAGE, CASTING, RESIN IMPREGNATED FIBERGLASS, Extra-Fast Setting, 3 inch (7.6 cm by 4 yards (3.7 meters), 10s.
6510-01-196-6205	BANDAGE, CASTING, RESIN IMPREGNATED FIBERGLASS, Extra-Fast Setting, 2 inch (5.1 cm) by 4 yards (3.7 meters), 10s.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6510-0758

LOCATION: NCR DISK #5

06793

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

.UCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in mitting comments and suggestions for improvements. All users of military standardization documents are invited to provide gestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and ied. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the blems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an nowledgement will be mailed to you within 30 days to let you know that your comments were received and are being sidered.

)TE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of cification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

ENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

FICIAL BUSINESS
ALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53476	PADDING, CAST, SYNTHETIC, Porous, Non-woven

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T06:20Z
Check the source to verify that this is the current version before use.