A-A-53477
16 May 1988

COMMERCIAL ITEM DESCRIPTION

PAPER, ARTICULATING, DENTAL (25 Foot Roll)

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification GG-P-00106B.

This commercial item description covers articulating paper suitable for dental use.

Salient characteristics:

Articulating paper shall be suitable for use by the dentists to ascertain degree of occlusal contact between opposing teeth. It shall have a soft cushion, rough surface and shall be supersensitive. Paper shall give a clear mark on porcelain, plastic, natural teeth and metals, wet or dry, with the slightest pressure.

Material. Articulating paper shall be of suitable paper coated with a vegetable or paraffin wax substance mixed with Prussian blue (Iron blue) dye. Dying shall be uniform throughout. Prussian blue (Iron blue) dye shall conform to the requirements of the Federal Food, Drug and Cosmetic Act.

Irritation. Prior to delivery, the manufacturer shall furnish a statement directly to the contracting officer stating that the articulating paper does not normally irritate oral tissues and does not contain poisonous ingredients in sufficient concentration to be harmful to human beings when used as directed.

Dimensions. The articulating paper shall have the following dimensions:

Thickness:	
Coating (each side)	0.0002 to 0.00025 inch
Paper with coating	0.0027 to 0.0033 inch
Width	5/8 inch +0,- 3/64 inch
Length	25 ft (minimum)

Style and design. Articulating paper shall be furnished in rolls. Each roll shall be completely interleaved with tissue paper. The maximum outside diameter of the roll shall be 1.56 inches.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Workmanship. Articulating paper shall be free from pin holes, rips or any other defects which may detract from its appearance or impair its function or intended use.

Unit. Roll (RO). One roll containing 25 feet by 5/8 inches of paper, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the articulating paper, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Quality conformance inspection.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 and 2.5 (percent defective) and an inspection level of S-4. Unit of product for examination purposes shall be one roll.

Examination. Examination shall not be restricted to the defects listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Defects

Texture not as specified.
Does not give clear mark on porcelain, plastic, natural teeth and metals.
Roll not properly wound; does not unwind uniformly and evenly.
Dimensions not as specified.
Edges of paper ragged.
Packaging and marking not as specified.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as must be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufacturers using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-542-3150	PAPER, ARTICULATING, DENTAL, Blue, 25 ft. by 5/8 inch.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2090

LOCATION: NCR DISK #5

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 - 505-033/92358

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.