A-A-53478
16 May 1988

COMMERCIAL ITEM DESCRIPTION

SPATULA, DENTAL, (Plastic)

The General Services Administration has authorized the use of this Commerical Item Description in preference to Military Specification MIL-S-36697.

This commercial item description covers a double-ended spatula suitable for use in mixing cement powder and liquid for dental restorations.

Salient characteristics:

Material. The spatular shall be fabricated of methyl methacrylate resin conforming to Type I, Class 3, of L-P-380 (see material and color in the note section). In addition, raw material specimens shall be tested for a minimum Rockwell hardness of 90 (M Scale) in accordance with procedure specified in the Rockwell hardness in the test section.

Style, design and dimensions. The Spatula shall conform to Defense Personnel Support Center (DPSC) Drawing 22182 in style, design, and dimensions. The spatula shall be suitable for use in mixing cement powder and liquid for dental restorations. The methyl methacrylate resin, used in the spatula fabrication, shall not be attacked by the cement liquid and shall not affect the color or other physical properties of the dental restorations. The cement, when mixed, shall not permanently adhere to the spatula.

Odor and color. The spatula shall be odor-free and shall have a milky white, translucent color (see material and color in the notes section).

Finish. The spatula shall have a smooth and shiny finish on all surfaces.

Workmanship. The spatula shall be free from defects which detract from its appearance or impair its serviceability. The spatula shall be free from bubbles, foreign matter, and nicks.

Unit. Each (EA). One spatula, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the used of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one spatula. Unless otherwise specified, sampling for examination shall be conducted at inspection level III, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1. For tests acceptance number shall be zero (0) for all sample sizes.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the methyl methacrylate resin used in the spatula, may be furnished in lieu of actual performance of such testing by the contractor, provided that lot identity has been maintained and can be demonstrated to the Government.

Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered are classified as specified herein. (see Sampling for examination).

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and the acceptable quality levels (AWQL's) shall be as indicated herein. The unit of product for examination purposes shall be one spatula. The spatula shall be examined for defects including, but not limited to, those listed below:

Defects.

Dispenser not of two-piece construction.
Parts of dispenser cannot be easily taken apart or assembled.
Finish not as specified.
Style and design not as specified.
Manufacturer's name or registered trademark not present on dispenser.
Packaging, packing, and marking not as specified.

Sampling for tests. A sufficient quantity of each lot of methyl methacrylate resin used in the fabrication of the spatula shall be selected for the Rockwell hardness (see Material) and the physical property tests to be performed in accordance with the applicable requirements of the subsidiary specification.

Tests. Tests shall be conducted to determine compliance with the requirements of this specification.

Rockwell hardness. Rockwell hardness shall be determined in accordance with ASTM Method D-785.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and of regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-00-926-2052	SPATULA, DENTAL, PLASTIC, No. 142

Material and color. A material and color cubicle for fabricating the spatula is E.I. Du Pont De Nemours and Company, Lucite Acrylic Resin No. 140 or 147F; color WT-15, White No. WT-118, or as an alternate, color WT-29 (see material, odor and color).

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:20Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

48610

PROJECT NO. 6520-2073
LOCATION: NCR - DISK 8

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.