A-A-53485
17 June 1988

COMMERCIAL ITEM DESCRIPTION

BASIN, WASH

The General Services Administration has authorized the use of this Commercial Item Description in preference to Interim Federal Specification RR-B-00120b(DSA-DM).

This Commercial Item Description covers a shallow, corrosion-resisting steel surgeon's hand basin with a capacity of at least 3-7/8 quarts.

Salient characteristics:

The wash basin shall be a shallow, corrosion-resisting steel surgeon's hand basin with a capacity of at least 3-7/8 quarts. The basin dimensions shall approximate 12-3/8 inches (31.43 cm) in diameter across, the top, 5-3/8 inches (13.65 cm) in diameter at the bottom, and 3-1/2 inches (8.89 cm) deep. Acceptable dimensional tolerance shall be ± 1/4 inch (0.64 cm).

The basin shall be formed seamlessly from one piece of 0.031 inch (No. 22 U.S. Standard Gauge) thick corrosion-resisting steel, American Iron and Steel Institute (AISI) type 302 or 304, which shall not be reduced below 0.027 inch thick at any point in the finished basin. The basin shall have either a tight round rolled bead, or a tight flat double rolled bead, approximately 3/16 inch thick, at the upper rim.

The basin shall rest on a level surface without rocking or wobbling. The bottom of the basin may concave inward so that the highest point shall be no more than 1/16 inch above the surface upon which it rests.

The basin surface shall have a uniform #4 commercial finish or better and be readily cleanable. The basin shall be finished in such a manner as to provide a pleasing professional appearance. Grinding of surfaces or electropolishing of the inner surface shall be acceptable.

AMSC N/A FSC 7240

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The basin shall be free of sharp edges, burrs, pits, dents, cracks, ripples, and deformities.

The basin shall be free of grease, oil, dirt, chips, and foreign matter.

Workmanship. The wash basin shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One wash basin, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one wash basin as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Defects.

Basin not AISI type 302 or 304 corrosion-resisting steel.

Basin not formed seamlessly from one piece of 0.031 inch thick corrosion-resisting steel.

Upper rim bead not formed as specified, not free of burrs and sharpness.

Basin not free of sharp edges, burrs, pits, dents, cracks, ripples, and deformities.

Basin capacity not a minimum of 3-7/8 quarts.

Basin does not rest evenly on a flat surface without rock or wobble.

Basin finish not a uniform #4 commercial finish or better.

Basin finish not free of grease, oil, dirt, chips, and foreign matter.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for tests purposes shall be one wash basin.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

7240-00-771-0945

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 7240-0246

PGC 57340

Location: CID/BASIN/H13-23

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53485 .	BASIN, WASH

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.