A-A-53488
17 June 1988

COMMERCIAL ITEM DESCRIPTION

DUST COVER, PLASTIC, SURGICAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a heat or tape sealable plastic bag suitable for use as a dust cover for sterilized medical supplies and instruments.

Salient characteristics:

The surgical, plastic dust cover shall be a bag type, clear polyethylene cover suitable for protecting sterilized medical supplies from dust contamination when sealed by heat or tape. The dust cover shall be supplied in the following sizes:

Size I	10 inches wide by 15 inches long (25.4 by 38.1 cm)
Size II	16 inches wide by 20 inches long (40.6 by 50.8 cm).

Widths and lengths specified for indicated bag size shall be minimum dimensions. Dimensional tolerances within each lot of material shall be plus or minus 1/4 inch for width and length.

The polyethylene film used in the construction of the bag shall be 3 mils ± 1/2 mil thick.

The dust cover shall be sealed, or closed, on three ends. One sealed, or closed, end shall have a tear line suitable for rapid opening without use of a scissors, knife, or other instrument. A perforated tear line shall not be acceptable as this would compromise sterility. The open end of the dust cover shall be readily sealable by use of heat or tape. Cover seams shall be free of wrinkles or bunching.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The dust cover shall be distinctively imprinted on one side with the following, or equivalent, labeling:

a. Grasp either corner and snap smartly to open.
b. Dust cover to be used after sterilization.
c. Heat seal or tape.

Printing shall be legible and complete.

When tested as specified, the bag shall burst at the tear lines at a pressure of 3-1/2 PSI, or greater.

The dust cover shall be free of holes, tears, or cracks. It shall be free of dirt, grease, or foreign matter.

Workmanship. The surgical, plastic dust cover shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing two hundred and fifty dust covers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one surgical, plastic dust cover as specified in this description.

Defects.

Cover has no tear line.

Cover seams are not completely sealed or closed.

Cover not free of holes, tears, or cracks.

Cover printing missing, illegible, or incomplete.

Cover seams not free of wrinkles or bunching.

Cover not free of dirt, grease, or loose foreign matter.

Sampling for tests. Sampling for tests including dimensional tests shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one dust cover.

Tests

Bursting Point. The dust cover shall be placed over an air, or equivalent gas, outlet tube and clamped to prevent leakage. Gas pressure shall be gradually increased until the bag bursts. Bursting at a pressure of less than 3-1/2 PSI shall constitute a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-253-3806 (Size I)
6530-01-252-5566 (Size II)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1999

PGC 70065
70066

Location: CID/DUST-PLASTIC/H13-28

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.