A-A-53489
17 June 1988

# COMMERCIAL ITEM DESCRIPTION

## GOGGLES, DARKNESS ADAPTATION

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers fluoroscopy-adaptation goggles suitable for use by physicians in adapting the eyes to darkness prior to doing fluoroscopic examinations.

Salient characteristics:

The darkness adaptation goggles shall be suitable for use in adapting the eyes to darkness prior to doing fluoroscopic examinations. The goggles shall be capable of being worn over ordinary prescription spectacles.

The goggles shall consist of an opaque, flexible frame, a light filter, and an elastic braid headband for holding the frame in place. The frame shall be designed to conform to the contour of the normal average nose and face, when worn with and without spectacles. The frame shall provide a smooth contoured surface where it rests on the forehead and cheeks. The frame shall not contain plasticizers or additives that will migrate to the surface of the frame or cause toxicity or skin irritation. The frame shall be so shaped that light, except that entering through the filter, will be excluded from the eyes.

The goggles frame shall have breather vents to prevent fogging and steaming of the filtered viewing area. The breather vents shall be light-baffled to exclude unfiltered light.

The light filter shall be an integral part of the goggles, or a separate, replaceable sheet. If replaceable, it shall be easily inserted and removed. The filter shall be free from scratches, dirt, and other defects.

The elastic headband used for holding the frame in place shall be free from cuts, tears, or frays. The strap shall maintain its elasticity and not become slack during the normal useful life of the goggles.

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:20Z
Check the source to verify that this is the current version before use.

The component parts of the goggles shall be free from cracks, chipped out areas, and tears. The surface of the goggles shall be free from burrs, sharp edges, and indentation marks.

Workmanship. The darkness adaptation goggles shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One pair of goggles, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one set of darkness adaptation goggles as specified in this description.

Defects.

Goggles not complete with all component parts.

Frame surface not smooth where it rests on forehead and cheeks.

Ventilating areas not properly designed to prevent fogging.

Component parts not free from cracks, chipped out areas, and tears.

Filter not free from scratches, dirt, or other defects.

Surface of goggles not free from burrs, sharp edges, and indentation marks.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6532-00-603-0990

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6532-1178

PGC 50846

Location: CID/GOGGLES/H13-24



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.