A-A-53494
17 June 1988

COMMERCIAL ITEM DESCRIPTION

PROCESSING MACHINE, RADIOGRAPHIC FILM, AUTOMATIC

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a compact processor suitable for automatically developing, fixing, and washing all standard brands of intra-oral dental x-ray film.

Salient characteristics:

The processing machine shall be capable of automatically processing all brands and all sizes of intra-oral dental radiographic films (sizes 0,1,2,3, and 4). The processor shall be comprised of at least the following:

a. Light protecting hood (daylight loader) with sleeve-type film insertion ports and light filter.
b. Film transport roller (spindle type).
c. Film holders - six each.
d. Developer tank, fixer tank, and water tank.
e. Storage compartment.
f. Controls.

The film transport system shall consist of a rotating constant speed threaded spindle into which the hook-shaped base of film holders engage, and provide uniform straight line motion through the developing, washing, and fixing tanks. Suitable fixed cams shall be provided to automatically control dipping and raising action of film holders into and out of processing solution tanks.

The processing machine shall be provided with three solution tanks which shall be fabricated of high impact plastic and shall be removable for cleaning. The film holders shall be fabricated of corrosion-resisting steel and shall accept up to three intra-oral dental x-ray films, including large occlusal films.

The processing machine with daylight loader in place shall have approximate dimensions of 19 inches wide by 13 inches long by 14 inches high.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The processing machine shall have a complete developing cycle of approximately 6 minutes. When operating at capacity, the machine shall be capable of processing approximately 300 intra-oral dental films per hour.

The processing machine shall be fabricated of high quality materials throughout. All parts shall be manufactured from materials which have sufficient mechanical strength to maintain proper alignment during the entire life of the unit. All components which contain solutions, or are exposed to splash and fumes from processing solutions during use, shall be fabricated of suitably corrosion resistant materials.

The processing machine shall operate satisfactorily on either 115 volts, 60 Hz, AC, or 220 volts, 50 Hz, AC, depending on customer's requirements. The machine need not have the capability of operating on both of the above voltages.

The processing machine shall be furnished with a three conductor line cord (with ground lead) fitted with a three blade male connection conforming with ANSI Standard C73-73. The machine shall meet the electrical requirements of Underwriter's Laboratories, Inc. (UL).

The processing machine finish shall be free of pits, burrs, sharp edges, scratches, stains, or other appearance defects.

Each processing machine shall be supplied with one set of instructions for use.

Workmanship. The automatic radiographic film processing machine shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One processing machine, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for sampling purposes shall be one automatic radiographic film processing machine as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Defects.

Any component missing, broken, or not as specified.

Film holders not fabricated of corrosion-resisting steel.

Solution tanks not removable for cleaning.

Electrical line cord and connector not as specified.

Processing machine not free of pits, burrs, sharp edges, scratches, stains, or other appearance defects.

Instructions for use not furnished.

Sampling for tests. Sampling for tests, including dimensional testing, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one automatic radiographic film processing machine.

Tests

Light - Tightness Test

The hands of the operator and three unexposed film packets shall be introduced through the light-tight sleeves of the hood. The processor chamber shall then be checked for visible light leakage through the sleeves.

The film pack covers shall then be removed and the films shall be fastened in the clips of the film holders. The films shall be then automatically developed, rinsed, fixed, and washed. After drying, the fully developed films shall be checked for signs of light leakage during processing. Any evidence of light leakage onto the films shall constitute a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-00-420-9588

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

A-A-53494

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1234

PGC 49516 Location: CID/PROCESSING/H13-29

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53494	PROCESSING MACHINE, RADIOGRAPHIC FILM, AUTOMATIC

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.