A-A-53495
17 June 1988

COMMERCIAL ITEM DESCRIPTION

PUMP, INFLATING, MANUAL

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a manually-operated inflating pump of the single-cylinder plunger type suitable for inflating low pressure pneumatic pillows and mattresses.

Salient characteristics:

The inflating pump shall be a manually-operated, single-cylinder, plunger type suitable for inflating low pressure pneumatic pillows and mattresses.

The cylinder shall be black, paper-based phenolic laminate tubing, approximately 3 inches in outside diameter and approximately 10 inches in length. The cylinder shall be silicone-graphite lubricated. The air discharge outlet shall have an internal .302 - 32 thread and external 1/2 - 14 thread, 3/8 inch to 1/2 inch long. The threads shall not be stripped, torn, broken, or otherwise defective.

The overall length of the pump shall not exceed 14 inches. The total weight of the pump shall not exceed 26 ounces. No adapters or hose shall be furnished.

Carbon steel parts of the pump shall be plated or painted to resist corrosion. Painting of moveable parts shall not interfere with the operation of the unit.

The surface of the pump shall not be rough, misaligned, or contain cracks, nicks, burrs, sharp edges, or other flaws.

The pump surface shall be clean and shall not contain embedded foreign matter.

The pump shall successfully pass the operating force test, the leakage test, and the impact test, as specified.

AMSC N/A FSC 4320

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Each unit shall be supplied with two copies of service data. The service data shall contain complete instructions for operation, maintenance, and a list of component parts.

Workmanship. The manual inflating pump shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One pump, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for sampling purposes shall be one manual inflating pump as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Defects.

Any component missing, malformed, corroded, fractured, or otherwise damaged.

Any component loose, detached, not as specified, or otherwise not securely retained.

Threads not as specified.

Any threads missing, stripped, torn, broken, or otherwise defective.

Pump surface rough, misaligned, or containing cracks, nicks, burrs, sharp edges, or other flaws.

Surface unclean or containing embedded foreign matter.

Service data not supplied.

Sampling for tests. Sampling for tests, including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one manual inflating pump.

Tests

Operating force test. The pump shall be held in a vertical position with the handle up and fully extended with the air outlet open. A weight of 1 pound applied to the handle shall, within 5 seconds, depress the plunger assembly to the full "in" position. Failure of this test shall constitute a defect.

Pump leakage test. The plunger handle shall be fully extended and the air outlet capped. With the pump in a vertical position, a 15-pound force applied in line with the longitudinal axis of the pump shall not bring the plunger assembly to the fully returned position in less than 30 seconds. Failure of this test shall constitute a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Impact test. The pump shall be dropped once in each of the following positions from a height of 4 feet onto a concrete surface:

a. With the longitudinal axis of the pump parallel to the concrete surface and with the plunger assembly in the "out" position.

b. With the longitudinal axis of the pump 45 degrees to the concrete surface and with the plunger assembly in the "out" position and toward the concrete surface.

c. With the longitudinal axis of the pump 90 degrees to the concrete surface and with the plunger assembly in the "in" position and the handle toward the concrete surface.

Upon completion of these drops, the pump shall be subjected to and pass the following:

1. Examination of pump.
2. Operating force test.
3. Pump leakage test.

Failure of any of these tests shall constitute a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

4320-00-299-8519

NOTE: The pump shall be suitable for inflation of NSN 7210-00-299-8520 Pillow, Pneumatic and NSN 7210-00-299-8518 Mattress, Pneumatic

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 4320-0308

PGC 40231

Location: CID/PUMP/H13-30

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.