A-A-53498
17 June 1988

COMMERCIAL ITEM DESCRIPTION

TABLE, TREATMENT

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a hardwood treatment table with foam pad and vinyl cover.

Salient characteristics:

The Treatment Table shall be of rigid hardwood construction with the following approximate dimensions:

Width - 30 inches (76.2 cm)
Length - 78 inches (198.1 cm)
Height - 32 inches (81.3 cm)

The table shall have four legs and each leg shall be a minimum of 2-1/4 inches in diameter or 2-1/4 inches square.

The table shall have a full length lower shelf (H-brace), a minimum of 3/4 inches thick, located approximately 7 inches off the floor. The table shall have an upper, smooth gliding, pull-out drawer. All metal hardware shall be corrosion-resisting.

The table shall be provided with a 2 inch (5.1 cm) foam pad covered with vinyl upholstery, sized to fit the entire table. The following upholstery colors shall be available:

a. White
b. Brown
c. Tan
d. Olive Green
e. Light Blue
f. Red
g. Black
h. Gold
i. Orange
j. Persimmon

The foam pad shall have no large voids or objectionable odor. The vinyl cover shall be fire retardant and meet the requirements of California Technical Bulletin #117. The table shall have a raised rim around the top edge to prevent the pad from slipping.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The table shall sit stable on a flat floor surface. It shall sit without rock or wobble, with or without a patient on it.

The table finish shall be smooth and uniform and free of scratches, splinters, and discoloration.

The table shall be so constructed that all corners shall be rounded and edges smooth. There shall be no danger of tearing of clothing by inadvertently brushing against the table.

Workmanship. The treatment table shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One treatment table, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one treatment table as specified in this description.

Defects.

Table not the correct width, length, or height.

Table does not have a full length lower shelf or an upper smooth gliding, pull out drawer.

Table legs are not at least 2-1/4 inches in diameter or at least 2-1/4 inches square.

Metal hardware is not corrosion-resisting.

Table does not have a 2 inch thick, foam pad covered with fire-retardant, vinyl upholstery, sized to fit the entire table.

Table does not have a raised rim to prevent the pad from slipping.

Pad color is not the color ordered.

The foam pad has voids or objectional odor.

Table finish is not smooth and uniform or free of scratches, splinters, and discoloration.

Table corners are not rounded or edges smooth.

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one treatment table

Tests

Load Test

The fully assembled treatment table shall be tested with a distributed load of 300 pounds on top. Failure to be free from wobble or distribution shall constitute a defect.

The fully assembled treatment table shall be tested with an empty load. Failure to be free from wobble or any unsteadiness shall constitute a defect.

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-113-0135

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPQAS

Project No. 6530-1935

PGC 50702

Location: CID/TABLE/H13-25

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-[redacted] RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53498	TABLE, TREATMENT

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.