A-A-53499
17 June 1988

COMMERCIAL ITEM DESCRIPTION

SURGICAL PACK, DISPOSABLE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a sterile, disposable surgical pack which is used for Cesarean section.

Salient characteristics:

1. The surgical pack shall be a sterile, disposable, Cesarean section pack, for creating a sterile field specifically for Cesarean section (abdominal) delivery and care of the newborn infant. The surgical pack shall consist of at least the following components in the listed nominal dimensions:

   a. One (1) overwrap/tablecover, no less than 44 by 88 in (111.76 by 223.52 cm).
   b. One (1) Mayo stand cover, with pad, no less than 23 by 55 inches (58.4 by 139.7 cm).
   c. One or two large or extra large surgical gown, with hand towel.
   d. Two (2) extra protection , large or extra-large surgical gowns, with towels 19 by 20 inches (48.26 by 50.8 cm) or 18 by 24 inches (45.72 by 60.96 cm).
   e. Ten (10) pre-washed gauze laparotomy pads, no less than 12 by 12 inches (30.5 by 30.5 cm) (optional).
   f. One (1) Cesarean section drape sheet 113 inches (287.02cm) overall featuring:
      Cesarean section drape with reinforcement.
      Clipping tabs (optional).
      Adhesive film with triangular fenestration 7 by 7 by 8 inches (17.78 by 17.78 by 20.3 cm) (optional).
      Fluid collection bag with adhesive closure.
      Cord holders (optional).
      Armboard covers.
   g. Two (2) absorbent towels, 19 by 20 inches (48.3 by 50.8 cm) or 18 by 24 inches (45.72 by 60.96 cm).

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

h. One (1) baby blanket no less than 30 inches by 26 inches long (76.2 by 66.04 cm).
i. One (1) bulb syringe, 2 or 3 oz (59.2 or 88.7 ml).
j. One (1) umbilical cord clamp.
k. One (1) sanitary napkin (optional).
l. One (1) suture bag.
m. One (1) Yankauer suction tip (optional).
n. Ten feet (3 meters) of nonconductive suction tubing (optional).
o. One (1) sharps disposal system (optional).

2. The overwrap table cover shall be a nonwoven fabric of polyethylene laminate and shall be fan-folded for ease in opening on the operating room (OR) table. The overwrap table cover shall be impervious to fluids.

3. The Mayo stand cover with pad shall be polyethylene film with a nonwoven fabric pad. The Mayo stand cover shall be fan-folded to permit easy draping. The pad shall be 12 by 18 or 20 by 24 inches (30.48 by 45.72 or 50.8 by 60.96 cm). The Mayo stand cover with pad shall be impervious to fluids.

4. One (1) surgical gown and hand towel shall be separately over-wrapped for use by the scrub nurse.

5. The surgical pack shall contain two (2) surgical gowns with extra fluid barrier fronts, minimum barrier area from nipple line to mid-thigh. The gowns shall be spun laced nonwoven made of 60 percent wood pulp and 40 percent polyester fibers.

6. The surgical gowns shall completely cover the back of the individual gowned and shall be water repellent. The surgical gowns shall have stockinette cuffs that conform to wrists and snap or tie neck closure. The surgical gowns shall meet the AORN Recommended Practices for Aseptic Barrier Materials for Surgical Gowns.

7. The hand towels shall be a minimum 14 by 22 or 18.75 by 19.50 inches (35.5 by 55.8 cm or 47.63 by 49.53 cm). The towels shall be highly absorbent to sufficiently dry arms after a surgical scrub.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

8. The laparotomy pads shall be laundered gauze. The laparotomy pads shall be radiopaque with a 6 inch (15.2 cm) loop preattached. The laparotomy pads shall have bright colored tape attached for easy identification. The gauze shall meet the requirements as listed in the latest edition of the United States Pharmacopoeia (USP). The gauze shall be fabricated of cotton, polyester or of a cotton/polyester blend.

9. The Cesarean section drape sheet shall be a block T style. The top shall be approximately 100 by 36 inches (254 by 91.4 cm) to include armboard coverage. The length shall be approximately 113 inches (287.02 cm) long by 72 inches (183 cm) wide. The Cesarean section drape shall be fabricated of a water repellent, bacteriocidal nonwoven fabric with a reinforced abdominal area. The reinforced area shall be impervious to fluids yet highly absorbent, eliminating strike-through while preventing pooling of fluids. The Cesarean section drape shall have a triangular fenestration with adhesive backing to accomodate either a midline or low transverse incision for delivery of the infant. The Cesarean section drape shall have a fluid collection bag with adhesive closure for the ammiotic fluid. The Cesarean section drape sheet shall include clipping tabs and cord holders.

10. The surgical pack shall contain an infant blanket fabricated of a cotton flannel or a rayon-polyester or a polypropylene nonwoven fabric. The infant blanket shall measure no less than 30 inches wide by 30 inches long (76.2 by 76.2 cm). The edges shall be finished on all sides (optional).

11. The surgical pack shall contain a 2 or 3 oz (59.2 or 88.7 ml) ear and ulcer syringe for aspiration of fluid and mucous from the mouth of the newborn. The suction shall be 120/140 mm Hg. The bulb syringe shall be made from PVC.

12. The surgical pack shall contain a plastic umbilical cord clamp. The umbilical cord clamp shall be made from nylon resin.

13. The surgical pack shall contain a highly absorbent sanitary napkin and measure approximately 3 by 10-1/2 inches (7.6 by 26.7cm). The sanitary napkin shall be made from pulp, cellulose wadding and nonwoven fabric (optional).

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

14. The surgical pack shall contain a disposable suture bag. The suture bag should be printed white paper. The suture bag when folded flat shall measure no less than 11-3/8 inches long and 6-1/2 inches wide (28.89 or 30.48 by 16.51 cm).

15. The surgical pack shall contain a transparent plastic Yankauer suction tip and 10 feet (3 meters) of nonconductive tubing with a minimum I.D. 3/16 inches (.48 cm). The Yankauer suction tip shall be made from acrylic resin.

16. The sharps disposal system shall hold a minimum of ten (10) needles and two (2) knife blades. The sharps disposal system shall close to protect personnel against injuries. The sharps disposal system shall have adhesive and magnetic attachments.

17. The absorbent towels shall be cellulosic.

18. All nonwoven items shall be fabricated of a soft material with minimal memory for good drapeability. The nonwoven items shall resist tearing, pilling and linting.

19. The drape shall be a at a minimum 1.8 or 1.59 oz/yard spunlaced polypropylene nonwoven material made of 60 percent wood pulp and 40 percent polyester fibers.

20. The bacteriocidal nonwoven fabric with a reinforced abdominal area shall be an absorbent nonwoven fabric laminated to a polyethylene film with impregenated antimicrobial treatment.

21. The surgical pack shall be ETO sterilization or cobalt sterilized.

22. If the surgical pack is opened and not used, the surgical pack shall be capable of re-sterilization by steam or ethylene oxide (optional).

23. The pack shall be sterile and individually wrapped in a suitable tear resistant package which will allow aseptic handling and maintain sterility of the contents until opened or damaged.

24. Each surgical pack shall also be wrapped in a heat sealed exterior dust cover (optional).

25. All surgical pack components shall be clearly listed on the exterior wrapper.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sterility The surgical packs shall be sterilized in accordance with the requirements of the food and drug administration (FDA).

Workmanship. The surgical pack shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing three or four individually sterile surgical packs, disposable as specified, constitutes one unit. Each surgical pack shall be individually wrapped in a suitable tear resistant heat sealed package. The individual package shall be capable of maintaining sterility of contents unless the package is damaged or opened. Each pack shall also be wrapped in a heat sealed exterior dust cover. All pack components shall be clearly listed on the exterior wrapper.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one Cesarean surgical pack as specified in this description.

Defects.

Any component part missing, broken, torn or otherwise not useable.

Pack not free of tears, cuts, holes, or perforations.

Sterility-compromized or not sterilized as required.

Pack not free of sand, grease, lint, dust or foreign material.

Uneven or ragged edges.

Raised yarns on front or back.

Any objectional odor.

Sampling for tests. Sampling for tests (dimensional), shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-1. The unit of product for test purposes shall be one Cesarean surgical pack.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-156-3390

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:20Z
Check the source to verify that this is the current version before use.

A-A-53499

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1946

PGC 70028

Location: CID/SURGICAL/H18-7

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-[redacted] RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53499	SURGICAL PACK, DISPOSABLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.