METRIC

A-A-53504B
09 December 1994
SUPERSEDING
A-A-53504A
24 December 1991

COMMERCIAL ITEM DESCRIPTION

BAGS, URINE COLLECTION - MEASURING DEVICE

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-B-0037189A (DLA-DM), which is cancelled.

This Commercial Item Description covers a sterile, disposable plastic urine collection bag with urometer.

Salient characteristics:

The adult urine collection bag shall be supplied in the following two sizes:

Size 1 - 2000 cubic centimeters sterile bag, with urometer and adjustable hanger, to permit suspension from a bed, chair, or patient.

Size 2 - 2500 cubic centimeters sterile bag, with urometer and adjustable hanger, to permit suspension from a bed, chair, or patient.

The sterile disposable urine collection bag shall consist of a main drainage urine container, an integral or attached urometer, an inlet tube with catheter adapter, drain tube assemblies, and an adjustable hanger to permit suspension from a bed, chair, or patient.

The bag design shall be such that urine automatically overflows from the urometer into the main bag when the capacity of the urometer is exceeded. The design shall also permit the elective emptying of the urometer contents into the main bag. A closed system shall be maintained during overflow or emptying of urometer contents into the main drainage bag.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The main bag shall be constructed of two sheets of suitable flexible plastic, heat sealed on all edges and vented at the top or in the catheter adapter in a suitable manner. The bag shall have a capacity of at least 2000 or 2500 cubic centimeters (cc) according to the size supplied. The bag shall not leak, or burst at a pressure less than 150 mm Hg.

The bag urometer shall be fabricated from a clear plastic material to permit the viewing of flow and contents. The urometer shall be either integral or adequately attached to the main drainage urine container. A drain shall be mounted at the bottom of the urometer to facilitate aseptic withdrawal of urine and shall include on-off control mechanisms affixed to the drain tube which will not cut the tube in operation.

The inlet tube shall have a minimum wall thickness of 0.040 inch and shall be permanently attached to the catheter adapter and the inlet port. The catheter adapter shall be fitted with a suitable tapered connector for attachment to a catheter and have a tip protector. The tubing shall be of sufficient length and sufficiently clear to view flow and contents.

The drain tube assembly shall be permanently attached to the bottom of the main bag The drain tube shall be designed to be drip-proof and to facilitate aseptic withdrawal of urine and shall include on-off control mechanisms affixed to the drain tube which will not cut the tube in operation

The main bag graduations for all sizes shall be in 100 cc steps, beginning at 100 cc The size 1 main bag graduations shall end at not less than 2000 cc, but not more than 2500 cc. The size 2 main bag graduations shall end at not less than 2000 cc but not more than 2500 cc. At least every other main bag graduation, beginning at not greater than 200 cc, shall be identified Identification of additional main bag graduation shall be permissible. The volumetric accuracy of the main bag graduations shall be at least ± 10%.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

The urometer graduations shall be in steps no larger than 10 cc. Urometer graduations shall begin at 10 cc and end at not less than 200 cc. The following urometer graduations shall be identified: 10, 20, 30, 40, 50, 60, 70, 80, 100, 120, 140, 160, 180, and 200 cc. Identification of additional graduations shall be permissible. The volumetric accuracy of the urometer graduations shall be at least ± 5% for volumes less than 100 cc, and at least ± 10% for volumes greater than 100 cc.

The graduation markings shall be clear and indelible They shall not be attached or secured with adhesives. The numbers and letters shall be a minimum of 0 32 cm in height.

The adjustable hanger cord shall be nylon, cotton, or wax impregnated cotton, tightly woven, 0.5 cm to 1.0 cm long and 0.24 cm minimum diameter. The cord shall be furnished with a suitable corrosion resistant hook, capable of supporting the weight of a bag filled with 2000 cc of water for Size 1 and 2500 cc of water for Size 2. Plastic strap hangers or hook and loop hangers may be used in lieu of cord.

The urine collection bag shall be free of stains, holes, tears, cracks, or other imperfections, and shall not leak

Each urine collection bag shall be supplied with a comprehensive instruction sheet for use. As alternates, the instructions for use shall be permanently and legibly marked on each bag, or on each individual sterile bag wrapper.

Sterility. The urine collection bags shall be sterile in accordance with United States Pharmacopeia (USP) or Association for the Advancement of Medical Instrumentation (AAMI) requirements. Sterility assurance levels shall be 0.000001.

Workmanship. The urine collection bag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG).

For Size 1, each urine collection bag shall be individually packaged in a suitable sealed container capable of maintaining sterility of contents unless package is damaged or opened. One package containing 10 individually packaged bags, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

For Size 2, each urine collection bag shall be individually packaged in a suitable sealed container capable of maintaining sterility of contents unless package is damaged or opened. One package containing 20 individually packaged bags, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after the delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one urine bag collection as specified Sampling shall be inspection level II with ar Acceptable Quality Level (AQL) of 1.0.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one urine collection bag. Sampling shall be inspection level S-1 with an AQL of 2.5.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications and Sizes are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6530-01-038-0731	BAG, URINE COLLECTION, PLASTIC, DISPOSABLE, 10s	1
6530-00-105-8649	BAG, URINE COLLECTION, PLASTIC, DISPOSABLE, 20s	2

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No 6530-2442
Location: c:\finals\AA-53504B

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-53504B	2. DOCUMENT DATE (YYMMDD): 94/12/09

3. DOCUMENT TITLE: BAGS, URINE COLLECTION - MEASURING DEVICE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: DPSC-MAP	b. TELEPHONE (Include Area Code) (1) Commercial: (215) 737-8431	(2) AUTOVON: 444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19101-8419	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340, AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.