A-A-53505
25 July 1988

COMMERCIAL ITEM DESCRIPTION

PLUGGER, AMALGAM, DENTAL

(Modified Mortonson, No. 2; Double-ended)

The General Services Administration has authorized the use of this Commercial Item Description.

Salient characteristics:

This commercial item description covers a double-ended plugger suitable for use in compacting plastic silver amalgam around reinforcement pins in a prepared tooth.

Material. Pluggers of one-piece construction shall be fabricated of class 4, type 440A stainless steel. Pluggers of three-piece construction shall have operative ends fabricated of class 4, type 440A stainless steel with class 3, type 303 stainless steel handles. Stainless steel shall be in accordance with ASTM F899 for composition. Chemical analysis shall be tested as specified herein.

Hardness. The operative ends of the finished instrument shall have a Rockwell hardness of C52 to C56 when tested as specified herein.

Design and dimensions. The plugger shall be a double-ended, modified Mortonson, No. 2, suitable for use in compacting plastic silver amalgam around reinforcement pins in a prepared tooth, without dislodgement of the pins. The handle shall have serrated or knurled finger grip portions. The plugger shall be in accordance with the latest revision of Defense Personnel Support Center Drawing 23164. Pluggers of three-piece construction shall have handles in accordance with handle C of DPSC Drawing 24490.

Finish. The operative ends and shanks of the plugger shall have a mirror finish. The non-serrated portion of the handle shall have a mirror or bright finish.

Identification marking. The handle shall be permanently and legibly marked with the instrument identification. Pluggers of foreign manufacture shall also be marked "stainless".

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Corrosion resistance. The pluggers shall be properly passivated. Pluggers of one-piece construction and the operative ends of pluggers of three-piece construction shall pass the copper sulfate and boil tests as specified in ASTM F1089. The handle of pluggers of three-piece construction shall pass the boil test as specified in ASTM F1089.

Workmanship. The operative ends and shanks of the plugger shall be smoothly finished and free from grind marks, gouges, cracks, burrs or other irregularities. Serrated portions of the handles shall be free of barbs and shall provide adequate finger grip during use. The plugger shall be free from other defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One plugger, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and an inspection level of II. Examination shall include, but shall not be limited to the defects listed below. The unit of product for examination purposes shall be one plugger.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Defects

Style and design of instrument not as shown on applicable drawing.
Operative ends of instrument not free from nicks, feathers, pits, burrs, or corrosion.
Finish on operative ends not as specified.
Instrument not free from cracks.
Handle serrations (or knurling, when applicable) of incorrect type, or missing altogether.
Finish on shanks and handle not as specified.
Instrument not free from foreign matter, dirt, oil, or polishing compound.
Handle not free from coarse grind marks.
Identification marking not present or not permanent, legible, and complete.
Instrument handle not straight.
Number or depth of serrations not within range specified.
Instrument number(s) improperly located.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one plugger.

Test Criteria.

Characteristic	Inspection level	AQL
Chemical analysis	(See Sampling of raw material.)	*
Dimensions	S-2	2.5
Hardness	S-2	1.5
Boil test	S-2	1.0
Copper sulfate test	S-2	1.0

* Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Sampling of raw materials. A sufficient quantity of each lot of raw steel used in the fabrication of the instruments shall be selected for chemical analysis.

Sampling unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instruments shall be determined in accordance with method 111 or 112 of Fed-STD-151. In the event of a conflict, method 111 shall govern.

Dimensions test. Tests of dimensions shall be made with a toolmaker's microscope, or other instrument of equal accuracy. Manufacturer's gages may be used if properly calibrated checking standards are used in conjunction with them.

Hardness. Hardness shall be determined in accordance with ASTM Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6520-00-935-1093	PLUGGER, AMALGAM, DENTAL, (Modified Mortonson, No. 2: Double-ended.)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

A-A-53505

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2067

LOCATION: NCR DISK # 5

48635

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge (DO NOT STAPLE), and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4968 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-ASTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53505	PLUGGER, AMALGAM, DENTAL (Modifified Mortonson, No.2: Double-ended)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*
b ADDRESS *(Street City State ZIP Code)*	

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:20Z
Check the source to verify that this is the current version before use.