MIL-S-37434
19 August 1976

MILITARY SPECIFICATION

STOPCOCK, INTRAVENOUS THERAPY, 3 WAY, PLASTIC

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile three-way stopcock with attached plastic tubing suitable for use in intravenous therapy.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

MMM-A-250	-	Adhesive, Water-Resistant, (For sealing Fiberboard Boxes).
PPP-B-585	-	Boxes, Wood, Wirebound.
PPP-B-601	-	Boxes, Wood, Cleated-plywood.
PPP-B-621	-	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	-	Boxes, Fiberboard.
PPP-T-76	-	Tape, Pressure-Sensitive Adhesive Paper, Water-Resistant, (For Carton Sealing).

MILITARY

MIL-S-36586	-	Sterilization Test Strip Set, Bacterial Spore.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Hq., Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:42Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 - Marking for Shipment and Storage

DRAWING

Defense Personnel Support Center (DPSC)
20544 STOPCOCK, INTRAVENOUS THERAPY, 3-WAY

(Copies of specifications, standards,and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.
The Pharmacopeia of the United States of America (USP)
(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042).

AMERICAN NATIONAL STANDARDS INSTITUTE INC., (ANSI)
ANSI Z70.1 Glass and Metal Luer Tapers for Medical Applications, Dimensions of.
(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, NY 10018).

U. S. DEPARTMENT OF HEALTH, EDUCATION AND WELFARE
FOOD AND DRUG ADMINISTRATION
Federal Food, Drug, and Cosmetic Act and General Regulations for its Enforcement.
(Application for copies should be addressed to Food and Drug Administration, U. S. Department of Health, Education and Welfare, Washington, DC 20204).

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:42Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 Plastic and rubber components. All plastic components, including end protectors, and rubber components shall be fabricated from materials of the best quality and suitable in all respects for the purpose intended. All materials shall be sterile, safe (non-toxic) and pyrogen-free. In addition, the plastics shall meet the requirements of the USP for Class II plastics when tested as specified in 4.4.3. The total combined quantity of barium, cadmium and lead for each component (excluding end protectors) shall not exceed 5 parts per million (ppm) when tested as specified in 4.4.1.

3.1.2 Stopcock body, male Luer connector, rotating spigot and end protectors. The stopcock body, male Luer connector and rotating spigot shall be fabricated from a suitable rigid plastic. As an alternate, the rotating spigot shall be fabricated from flexible, natural rubber mounted to a suitable rigid plastic handle. The rotating spigot shall be of a color contrasting to the stopcock body. End protectors shall be fabricated from a suitable rigid or semi-rigid plastic.

3.1.3 Plastic tubing. The plastic tubing shall be fabricated from suitable transparent, pliant plastic. Tubing shall be flexible, but shall be rigid enough to prevent collapse during normal use.

3.1.4 Rubber injection site. The rubber injection site shall be fabricated from laboratory pure gum rubber and shall meet the requirements specified in 3.1.1.

3.2 Style, design and dimensions (general). The three-way stopcock, hereinafter referred to as "set", shall be an intravenous therapy unit suitable for administering and controlling the flow of blood and solutions for simultaneous and alternate intravenous therapy. The set shall consist of a stopcock proper with rotating spigot and secured length of tubing to the stopcock body. A male Luer connector (needle adapter) shall be secured to the distal end of the tubing. The injection site shall be assembled to the rotating spigot (Type I design) or secured in line with tubing above the male Luer connector (Type II design). The injection site shall enable piercing and immediate resealing characteristics when punctured with needles of 18 gage or smaller. The set shall have three end protectors to maintain sterility of fluid path; two assembled to the open female Luer cavities, and one assembled to the male Luer connector. The entire set shall be in accordance with DPSC Drawing 20544, Type I or Type II design, for general style, design and dimensions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

3.2.1 Stopcock body and rotating spigot (Type I design). The stopcock body shall have two sides ports, 180 degrees apart, each having internal female Luer taper in accordance with figure 2 (Luer Slip), female end, of ANSI Standard Z70.1, except that the taper shall be 0.060 ± .007 inch per inch, and the start of the taper shall have an inlet diameter of 0.169 ± .001 inch. The taper length shall be 0.315 ± .005 inch and have a diameter of 0.150 ± .001 inch at the distal end. The cavity of the rotating spigot that contains the rubber injection site shall also possess a female Luer taper as specified above. All female Luer tapers shall meet the requirements specified in 3.2.4. The rotating spigot shall have an arrow-type handle to indicate which female side Luer port will be part of the fluid path.

3.2.1.1 Stopcock body and rotating spigot (Type II design). The stopcock body shall have two side ports, 180 degrees apart, each having internal female Luer taper in accordance with figure 2, female end, of ANSI Standard Z70.1. The ports shall be of the Luer Lock type (reference figure 3 of ANSI Standard Z70.1). All female Luer tapers shall meet the requirements specified in 3.2.4. The rotating spigot shall have a handle to indicate which female Luer ports will be part of the fluid path. Type II design stopcocks shall have four fluid paths.

3.2.2 Tubing with male Luer connector. Tubing shall be secured to the stopcock outlet. A male Luer tapered connector shall be secured to the distal end of the tubing (Type I design) or the distal end of the rubber injection site (Type II design). All connections shall be made in a leakproof manner. When tested as specified in 4.4.7, there shall be no movement of the ring gage on the tip when the gage is pressed lightly from side to side in all directions. In addition, when the ring gage is fitted onto the male connector, the tip of the connector shall not protrude beyond the "NO GO" position (outermost surface of the gage) of the ring gage. The male connector shall be designed to form a leakproof connection with standard Luer hypodermic devices. The male connector shall be constructed with a finger-grip device for ease of needle insertion.

3.2.3 End protectors. The two stopcock open ports and male Luer connector shall be covered with readily removable plastic protectors. The end protectors shall maintain cleanliness and freedom from particle infiltration and shall maintain sterility of the fluid paths until manually removed.

3.2.4 Taper and depth of female Luer cavities. The two female Luer cavities of the stopcock body (all designs) and the female Luer cavity of the rotating spigot (Type I design) shall be tested for proper taper of Luer cavity as specified in 4.4.6. When the Luer cavities are fitted onto the plug gage, there shall be no movement of the Luer cavities when they are pressed lightly from side to side in all directions. In addition, the "GO" position (excessive depth position) of the feeler gage shall not be exceeded (i.e., the plug gage shall not bottom into the Luer cavity without allowing passage of the feeler gage between the hub cavity and the shoulder of the plug gage).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

3.2.5 Leakage of set. When tested as specified in 4.4.8, all female Luer cavities, male connector, injection site, tubing, all junctions and the connection of the rotating spigot and stopcock body shall not show leakage exceeding the formation of one clinging air bubble. As an alternate, all of the above specified components and junctions shall not show leakage exceeding the formation of one clinging drop when tested as specified in 4.4.8.1.

3.2.6 Clear passage of fluid path. All the fluid paths of the set shall be clean and free from particles, and shall be tested for clear passage of all fluid paths as specified in 4.4.9. Sets which are sealed or constricted shall be rejected.

3.2.7 Elevated temperature storage. When tested as specified in 4.4.10, all of the components of the set shall be free from tackiness, brittleness, or other evidence of deterioration.

3.2.8 Security of assembly. The male Luer connector and stopcock proper shall not separate or loosen from the plastic tubing when tested as specified in 4.4.11. In addition to the above, the rubber injection site shall not separate from the male Luer connector and tubing for Type II design sets.

3.3 Sterility. The sets shall be sterile when tested with a biological indicator as specified in 4.4.2.

3.3.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number (NSN).
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity or product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.2. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.4 Certificates of freedom from pyrogens, safety and biological tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested in accordance with the methods described in the USP and found to be free from pyrogens, safe, and have passed the biological tests. This certificate shall include the specific tests performed and the test results. These certificates shall be submitted by the contractor in the manner as specified in 3.3.1.

3.5 Manufacturer's record. The manufacturer shall place a code or lot number on the immediate container or unit package which can be verified against the manufacturer's records of all tests made on sterile, pyrogen free and nontoxic units. These records shall be made available for inspection by any duly qualified medical officer of the Armed Forces, National Institutes of Health, or any official of the Federal Food and Drug Administration. No shipment shall be made until all the tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.6 Compatibility. The materials of which sets are manufactured shall not react with, or be acted upon, by the liquids with which they will be in contact during normal use and operation, nor shall there by any materials incorporated in the set which will cause any deleterious effect when the set is used for the purpose intended.

3.7 Instructions. Each set shall be provided with a detailed instruction sheet for proper use of the item.

3.8 Workmanship. Workmanship shall be first class throughout. The sets shall be free from defects which detract from their appearance or impair their serviceability. The stopcock body shall be free from flash, cracks, holes, pits and breakage. The tubing shall show a smooth finished outer surface free from kinks, cracks, holes, and toolmarks; and inner surface free from scale, deep pits, deep furrows, and die marks.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:42Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in table I. The unit of product for examination purposes shall be one complete set.

TABLE I. Sampling criteria for examination

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	None defined	---

4.2.2 For tests.

4.2.2.1 For sterility test. Sampling for sterility test shall be in accordance with the USP for parenteral devices and as specified in 4.4.2.

4.2.2.2 For pyrogen and safety tests. Sampling for pyrogen and safety tests shall be as indicated in the USP for transfusion and infusion assemblies. Samples shall be selected which are to be representative of each lot of each plastic used in the manufacture of the sets. As an alternate, sampling shall be on a sterilizer lot basis when sterilizer lots are smaller than, or equal to, production lots.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:42Z
Check the source to verify that this is the currentversion before use.

4.2.2.3 For other tests. Sampling for other tests shall be conducted in accordance with MIL-STD-105, and as indicated in table II. The unit of product for testing purposes shall be one complete set. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics. Defective sets shall be rejected.

TABLE II. Sampling criteria for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Set	Material - Class II	3.1.1	4.4.3	1/	2/
	Extractable metals	3.1.1	4.4.1	1/	2/
	Dimensions	3.2	-----	S-1	2.5
	Taper of male Luer connector	3.2.2	4.4.7	S-1	1.0
	Taper and depth of female Luer cavities	3.2.4	4.4.6	S-1	1.0
	Leakage test	3.2.5	4.4.8 or 4.4.8.1	S-1	1.0
	Clear passage of fluid paths	3.2.6	4.4.9	S-1	1.0
	Elevated temperature test	3.2.7	4.4.10	S-1	2.5
	Security of assembly	3.2.8	4.4.11	S-1	1.0

1/ Sufficient samples to perform the biological tests and extractable metal tests shall be selected from each lot and shall be representative of each batch of plastic and rubber used in the manufacture of the sets.

2/ No defects permitted.

4.3 Examination.

4.3.1 Examination of sets. The sets shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any sets in the sample containing one or more defects shall be rejected and if the number of defective sets in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects.

Categories	Defects
Critical	None defined.
Major	
101	Stopcock not of design specified.
102	Stopcock not free from cracks, holes, pits, burrs, slivers, flash and foreign matter.
103	Rotating spigot does not exhibit smooth operation without sticking and binding when turned in all positions.
104	Fluid paths not protected by end protectors.
105	Fluid paths obstructed and not free from foreign matter.
106	Silicone or other lubricants present in Luer cavities and on male Luer connector.
107	Tubing not free from kinks, pits, and toolmarks.
108	Rubber injection site not assembled to unit.
109	End protectors not hand removable.
110	Spigot and body of same color.
111	Instructions for use missing or incomplete.

4.3.2 Packaging inspection, The inspection of the unit, intermediate package, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Extractable metals. The set components (excluding end protectors) shall be tested for extractable metals. The specimens shall be cut such that no thickness exceeds 3 millimeters (mm) and eluted for either 72 hours at 50 degrees Centigrade (^{o}C), 24 hours at 70°C, or 60 minutes at 121°C in 0.9 percent saline. Each gram of each specimen shall be covered with 5 milliliters (ml) of the menstruum (recommended amount is 10 grams of the specimen plus 50 ml of solvent). The menstruum shall be collected after agitation, at the end of the selected elution period and submitted to spectographic analysis for barium, cadmium and lead. The metals shall be reported in ppm in the eluate. A value of 1.0 ppm in the eluate shall be equivalent to extracting 1 milligram from 200 grams of test material in the elution period into 1000 ml elution menstruum.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:42Z
Check the source to verify that this is the current version before use.

4.4.2 Sterility test. All fluid paths and end protectors of the set shall be tested for sterility in accordance with the method specified in the USP for parenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.2.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I sterilization and 30 for Type II Sterilization including carrier and product test units.

4.4.2.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.2.2 Inoculated carrier. Inoculated carriers (paper spore strips) used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.2 Sterility test and 4.4.2.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.2.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.2.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method.

4.4.2.4 Product testing. Testing for sterility of the product after receipt by the Government shall be without a biological indicator and in accordance with the USP.

Source: https://assist.dla.mil -- Downloaded: 20260-21T10:42Z
Check the source to verify that this is the current version before use.

4.4.3 Biological tests for Class II plastics. All plastic components used in fabrication of the stopcocks shall be tested in accordance with the requirements of the USP for Class II plastics.

4.4.4 Safety (toxicity test). Samples of the finished set, including end protectors, shall be tested for safety in accordance with the USP for transfusion and infusion assemblies.

4.4.5 Freedom from pyrogens. Samples of the finished set, including end protectors, shall be tested for freedom from pyrogens in accordance with the USP for transfusion and infusion assemblies.

4.4.6 Taper and depth of female Luer cavities. All female Luer cavities of the stopcock proper shall be tested with the male Luer taper plug gage in accordance with figure 5 of ANSI Standard Z70.1 (this gage shall be as shown in the ANSI Standard, except that it shall be constructed with a suitable pneumatic path to perform the leakage test as specified in 4,4,8). The female Luer cavities shall be slipped onto this gage with a straight motion, without turning or twisting the plug gage into the cavity. Once the female Luer cavity is slipped onto the gage, no additional force shall be exerted to force the plug gage into the Luer cavity. The cavities shall be observed for movement when they are pressed lightly from side to side on the plug gage. The feeler gage, in accordance with figures 5 and 6 of ANSI Standard Z70.1, shall then be inserted between the Luer cavities and the shoulder of the plug gage to insure that the "GO" position is not violated.

4.4.7 Taper of male Luer connector. The taper of the male connector shall be tested with the female Luer ring gage of figure 7 of ANSI Standard Z70.1. The male connector shall be slipped into this gage with a straight motion, without turning or twisting the male connector into the gage. Once the male connector is slipped into the gage, no additional force shall be exerted to force the male connector into the gage. The connector shall be tested for movement (wobble) and protrusion beyond the "NO GO" position of the gage.

4.4.8 Leakage test (pneumatic method) - Type I design. Samples of the set shall be subjected to an internal air pressure of 8 pounds per square inch (psig). The end protectors of the set shall be removed and the rubber injection site shall remain assembled to the stopcock proper. The injection site shall be punctured with an 18 gage hypodermic needle. The position of the arrow of the spigot shall be such that flow is created through one of the two female Luer cavities and through the tubing. The modified plug gage specified in 4.4.6, shall be inserted into the female Luer cavity that is opened to flow as indicated by the position of the spigot arrow. The plug gage shall be inserted into the female Luer cavity with a straight motion, without turning or twisting the plug into the cavity with a force not to exceed 5 pounds. The feeler gage shall then be placed onto the plug gage to assure that the "GO" position is not violated. The other female Luer cavity shall not be plugged. A hypodermic

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

needle, with the lumen sealed, shall be used to seal the male Luer connector in the following manner: The hub of the hypodermic needle shall be tested to comply with acceptable Luer cavity as specified in 4.4.6, except that in addition, the Luer cavity shall also pass the "NO GO" position of the feeler gage. A needle possessing a hub cavity that passes the above requirements for diameter and depth of Luer cavity shall then be used to seal the male Luer connector of the set. Air shall then be internally admitted to the testing arrangement in a leakproof manner with internal pressure of 8 psig and the entire testing assembly shall be completely immersed into a suitable container of water, with the air pressure applied for at least 3 minutes. The remaining female Luer cavity shall then be tested by repeating the above procedure while insuring that the spigot arrow position is open to flow to allow air entry through the modified plug gage. The entire set shall be observed for leakage as specified in 3.2.5.

4.4.8.1 Leakage test (hydrostatic method) - Type II design. The end protectors of the set shall be removed and the rubber injection site punctured with an 18 gage hypodermic needle. The test shall then be performed as specified in 4.4.8, except that an internal water pressure of 20 psig shall be applied for at least 3 minutes. The entire set shall be observed for leakage as specified in 3.2.5.

4.4.9 Clear passage of fluid path. The sets shall be tested with an appropriate testing fluid applied internally to the stopcock body with the stopcock body placed in a horizontal position. Each fluid path shall be tested, in turn, using testing assembly fixtures as specified in 4.4.8, except that the male connector shall not be plugged and that the female Luer cavity of the rotating spigot shall be tested by removal of the rubber injection site (for Type I design only). The set shall allow free passage of the testing fluid within 3 seconds after an internal pressure of 3.5 psig maximum has been applied to each female Luer cavity.

4.4.10 Elevated temperature storage test. Samples of the packaged set (immediate container of each set) shall be placed into an oven for seven days at 145° ± 2°F. Upon completion of the aging period the set shall be unpackaged and observed for characteristics specified in 3.2.7.

4.4.11 Security of assembly test. The sample sets shall be subjected to the security of assembly test after completion of the elevated temperature test specified in 4.4.10. The tubing of the set shall be severed at the approximate midpoint of the tubing length. The security of assembly of the tubing junction with the stopcock body, and the tubing junction with the male connector, shall be tested by gradually applying a five pound load for 2 minutes (dead weight) to the severed end of the tubing while each junction is held in a clamping device with the tubing hanging vertically downward. As an alternate for sets of Type II design, all junctions shall be subjected to an axial force of 8 pounds utilizing an appropriate tension testing device.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:42Z
Check the source to verify that this is the currentversion before use.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One package containing fifty (50) stopcock sets, as specified, constitutes one unit of issue.

5.1.1.1 Level A.

5.1.1.1.1 Individual package. Each stopcock set shall be packaged in a sealed peel open type bag of appropriate size and design and be capable of maintaining sterility of contents unless damaged or opened.

5.1.1.1.2 Unit package. Fifty (50) individually packaged stopcock sets (one unit of issue) shall be packaged in a snug fitting box constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Four (4) units of issue shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.2 Level B. Four (4) units of issue shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers at the lowest rate to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Individual package. In addition to any commercial markings, each individual package shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. The sterility lot (control) number and the legend "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement shall be shown.

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Sterility lot (control) number, contract or purchase order number and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A -5/74 - G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Sterility lot (control) number and method of sterilization as specified in 5.3.2 shall be shown.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) National Stock Number (NSN).

6.2 This specification covers the following item listed in the Federal Supply Catalog: NSN 6515-00-864-8864 STOPCOCK, INTRAVENOUS THERAPY, 3 Way, Plastic, Disposable, 50s.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1993

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-S-37434 STOPCOCK, INTRAVENOUS THERAPY, 3 WAY, PLASTIC

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974—603-680/6126 2 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

FOLD

d determined valid for use in acquisition

have been reviewed an

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:42Z
Check the source to verify that this is the current version before use.