INCH-POUND

MIL-DTL-25666/5 (USAF)
AMENDMENT 1
24 July 2000

DETAIL SPECIFICATION SHEET

CONVERTER, LIQUID OXYGEN GCU-17/A

This document remains inactive for new design.

This amendment forms a part of MIL-DTL-25666/5(USAF), dated 15 September 1995, and is approved for use by the Department of the Air Force and is available for use by all Departments and Agencies of the Department of Defense.

PAGE 3

3.3.7.1: Delete and substitute:

"3.3.7.1 Constant flow. The converter shall deliver a flow of gaseous oxygen up to 150 liters per minute for not less than 2 hours with the operating pressure remaining between 280 and 320 psig after the first and before the last 15 minutes of flow. An intermittent increase or decrease of pressure by an additional 10 psig is allowable provided these variances do not exceed a 5 minute interval for any one pressure cycle. The flow at the specified rate shall be initiated within 10 minutes after the converter has reached the operating pressure. The full test procedure is specified in MIL-DTL-25666."

PAGE 4

3.3.9: Delete and substitute:

"3.3.9 Operating temperature. The operating pressure shall remain between 280 and 325 psig after the first and before the last 15 minutes of flow at the low temperature and between 280 and 320 psig after the first and before the last 15 minutes of flow at the high temperature. An intermittent increase or decrease of pressure by an additional 10 psig is allowable provided these variances do not exceed a 5-minute interval for any one pressure cycle."

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T06:55Z
Check the source to verify that this is the current version before use.

Custodian:
Air Force – 99

Reviewer:
Air Force - 11

Preparing activity:
Air Force – 71

Agent:
Air Force – 99

Project 1660-F0815

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T06:55Z
Check the source to verify that this is the current version before use.