U-W-00149a(DSA-DM)
AMENDMENT - 1
13 April 1967

AMENDMENT
TO
INTERIM FEDERAL SPECIFICATION

WAX, DENTAL, STICKY

This amendment was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this amendment as a valid exception to Interim Federal Specification U-W-00149a (DSA-DM), dated 9 September 1966.

PAGE 3 Paragraph 3.4.3, line 1. Delete "0.5" and substitute "0.65".

PAGE 6 Paragraph 4.4.1.2. Delete period (.) at end of paragraph and add the following: ", after which the specimen shall be removed and cooled in air to 23.0 ± 2.0° C. (73.4 ± 3.6° F). The cellophane shall be stripped off and the final length determined in the same manner as the original length. The flow, as evidenced by the change in length, shall be reported as percentage of the initial length. The value for flow at any temperature shall be the average value for three specimens and shall be reported to the nearest 0.1 percent."

PAGE 7 After paragraph 5.1.2.1 insert new paragraph: "5.1.3 Level A."

PAGE 7 Paragraph 5.1.3 Unit package. Change paragraph number to 5.1.3.1.

*U.S. GOVERNMENT PRINTING OFFICE: 1967-251-519/6003

FSC 6520

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:20Z
Check the source to verify that this is the current version before use.