COMMERCIAL ITEM DESCRIPTION

INTRAUTERINE DEVICE, CONTRACEPTIVE

The General Administration has authorized the use of this commercial item description.

This commercial item description covers an intrauterine device suitable of use in intrauterine contraception.

Salient characteristics:

General. The system shall be a sterile, T-shaped, copper and polyethylene intrauterine device with an inserter and plunger.

Materials. The T-shaped body shall be fabricated of Alathon 20 polyethylene. Copper wire shall be Grade 1 copper wire. Copper collars shall be Grade 1 copper 99.99% pure. The polyethylene "T" shall contain 22% barium sulfate added to render it radiopaque. Strings shall be fabricated of polyethylene containing up to 1% titanium dioxide for visibility.

Style and design. The system shall be a T-shaped polyethylene body with approximately 176 mg of copper wire wound around the vertical portion of the "T". Copper collars each weighing 66.5 mg shall be provided on each arm of the horizontal sections of the "T". The bottom of the T shall be bulb shaped with a monofilament polythylene thread tied through the bulb. The thread shall allow protrusion through the os for conformation checks that the intrauterine device is insitu. An insertion tube and plunger shall be furnished with each device. The insertion tube shall be equipped with a moveable flange to aid in gauging the depth to which the insertion tube is inserted through the cervical canal and uterine cavity. The inserter shall be 8-1/8 $\pm$ 1/16 inches long with an O.D. of 0.0173 $\pm$ 0.004 inches and an I.D. of 0.146 $\pm$ 0.003 inches. IUD vertical leg and plunger distal end shall easily fit in the inserter. The inserter I.D. containing the IUD vertical leg shall allow the bent ends of both cross legs to also fit in the inserter. The cross arms shall sustain no damage and shall return to the original "T" configuration when bent into the inserter with the vertical position of the "T" and then released.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Plunger shall be 7.5 ± 0.1 inches long from handle brace to tip with a body diameter of 0.10 inches and a distal end diameter of 0.11 inches. Plunger shall have a finger ring on the proximal end for use in manipulating the plunger. The IUD shall be 1.245 to 1.270 inches wide by 1.405 to 1.425 inches high with at least 4-1/2 inches of polyethylene string attached to the bulb. The plastic body shall be 0.064 inches in diameter. The copper bands shall be 0.198 inches wide each located 7/32 inches from the vertical portion of the "T".

Instructions. Instructions for use shall be provided with each IUD which shall also contain a minimum of one patient information brochure and one patient wallet card. Shall also contain additional pertinent information on the item for physicians use.

Biocompatibility. The IUD shall meet the requirements of the USP for Class II plastics.

Sterility. The set components and inside surfaces of the containing envelope shall be sterile and have a minimum quality assurance level of 0.000001 in accordance with AAMI guidelines.

Shelf life. A minimum of 40 months shelf life shall remain on shipment of the IUD to the Government.

Workmanship. The IUD shall be free from defects which detract from its appearance or impair its serviceability. Shall be free from pits, cracks, distortion, corrosion, burrs, sharp edges or foreign material.

Unit. Package (PG). One package containing five individually packaged sterile intrauterine devices, with instruction for use, as specified, constitutes one unit. Each device shall be packaged in a peel back container for aseptic removal. The individual package shall be a sealed peel-open container capable of maintaining sterility of contents unless the package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein. Unit of product for examination purposes shall be one intrauterine device. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one intrauterine device. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the products is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-283-5192	INTRAUTERINE DEVICE, CONTRACEPTIVE, Sterile, T-shaped, copper and Plastic

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-5015

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3 and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

1. RECOMMENDED A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54403	5-22-91

3. DOCUMENT TITLE: Intrauterine Device, Contraceptive

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) / e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial / (2) AUTOVON
Defense Personnel Support Center	(215)737-2870 / 444-2870
c. ADDRESS (Include Zip Code) ATTN. DPSC-MSTE 2800 S 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426 OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.