INCH-POUND

A-A-54404A
29 May 1992
SUPERSEDING
A-A-54404
17 May 1991

# COMMERCIAL ITEM DESCRIPTION

## LUMBAR PUNCTURE KIT, STERILE, DISPOSABLE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable lumbar puncture kit containing all equipment for a single lumbar puncture.

Salient Characteristics:

General. The lumbar puncture kit shall be of the following types:

Type I - Adult
Type II - Infant
Type III - Pediatric
Type IV - Adult

The lumbar puncture kit shall consist of a single/double tier, compartmented plastic tray containing at least the following components:

| Component | Quantity | | | |
|---|---|---|---|---|
| | Type I | Type II | Type III | Type IV |
| Cup, prep, removable | 1 | 1 | 1 | 1 |
| Sponges, prep with long handle, 5 inches (13 cm) minimum length | 2 | 2 | 2 | |
| Prep swabs on handles | | | | 3 |

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

| Component | Quantity | | | |
|---|---|---|---|---|
| | Type I | Type II | Type III | Type IV |
| Towel, absorbent, with fluid repellant laminations, 18 inches by 13 inches (46 cm x 33 cm) minimum size | 1 | 1 | 1 | 1 |
| Bandage, adhesive, 3 inches by 3/4 inch (7.6 cm x 1.9 cm) | 1 | 1 | 1 | 1 |
| Drape, 26 inches by 18 inches (66 cm x 46 cm) minimum size with fenestration not exceeding 15.6 square inches in area | 1 | | | |
| Drape, 13.5 inches by 13.5 inches (34 cm x 34 cm) minimum size, with fenestration not exceeding 15.6 square inches in area | | 1 | 1 | |
| Drape, minimum 17 inches X 28 inches (43 cm X 71 cm) with fenestration opening minimum 3 inches X 3 inches (8 cm X 8 cm) centrally located, minimum 2 adhesive areas on fenestrated drape to hold drape in place | | | | 1 |
| Syringe, 2-1/2 cc capacity minimum | 1 | 1 | 1 | 1 |
| Needle, 25 gauge by 5/8 inch | 1 | 1 | 1 | 1 |
| Needle, infiltration, 21 gauge by 1-1/2 inch or 22 gauge by 1-1/2 inch | 1 | | | 1 |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

| Component | Quantity | | | |
|---|---|---|---|---|
| | Type I | Type II | Type III | Type IV |
| Needle, spinal, 20 Gauge by 3-1/2 inches | 1 | | | 1 |
| Needle, spinal, 20 gauge, 3-1/2 inches long (8.9 cm) with stylet | | | | 1 |
| Needle, spinal, 22 gauge by 1-1/2 inches | | 1 | | |
| Needle, spinal, 21 gauge by 2-1/2 inches, 22 gauge by 2-1/2 inches or 22 gauge by 2 inches | | | 1 | |
| Anesthetic, local, Lidocaine, Hcl injection, USP, 1% (2 cc scored ampule with directions for use. Type I, II, III) | 1 | 1 | 1 | 1 |
| Stopcock, 3-way or 4 way (*1) | 1 | 1 | 1 | 1 |
| Manometer, either one-piece or two-piece plastic, accurately calibrated to 55 cm minimum range | 1 | | | 1 (two piece) |
| Manometer, one piece plastic, accurately calibrated to 36 cm minimum range (*2) | | | 1 | |
| Tube, extension having a minimum length of 5 inches (12.7 cm) (*3) | 1 | 1 | 1 | 1 (one end male other female) |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

| Component | Quantity | | | |
|---|---|---|---|---|
| | Type I | Type II | Type III | Type IV |
| Containers (tubes), specimen with a capacity of at least 7 cc. numbered with attached screw-on caps and label or space for writing. Tubes to stand upright in depressions provided | 3 | 3 | 3 | 3 |
| Sterilization indicator | | | | 1 |
| (*1) A suitable valve which performs the functions of the stopcock may be substituted | | | | |
| (*2) May be two-piece with range of 55 cm minimum | | | | |
| (*3) Not required with valve | | | | |

Tray shall be recessed to receive and retain components in position, and in the sequence of use. Needle shall be furnished with suitable protectors.

The lumbar puncture kit shall be packaged so as to maintain sterility of the tray and contents until manually opened.

Material. The needle shall be fabricated from austenitic corrosion resistant steel in accordance with ASTM Designation F899.

Sterility. Components shall be sterile in accordance with the USP or AAMI guidelines. Sterility assurance level shall be 0.000001.

Biocompatibility. Plastic component material in contact with bodily tissue or fluids shall meet the requirements of the USP for Class II plastic.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Expiration dating. Expiration dating shall be not less than 36 months. Not less than 32 months of the expiration dating shall remain at the time of delivery to the Government. Type IV kit shall be furnished with no less than 24 months to the expiration date at the time of delivery to the government.

Instructions. Instructions for use and a complete list of contents shall be provided with each kit.

Workmanship. The lumbar puncture kit shall be free from defects which detract from its appearance or impair its serviceability. Components shall be free from tears, holes, flash, burrs, sharp edges, corrosion or foreign material.

Unit. Package (PG). One package, containing twenty type I Kits, ten type II or type III, each kit packaged sterile, as specified, constitutes one unit. The individual kit container shall be capable of maintaining sterility of contents unless package is damaged or opened. Unit for type IV item shall appear in solicitation.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one lumbar kit. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one lumbar puncture kit. Sampling shall be inspection level S-2 with an AQL of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction , the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers, Item Identifications, and their Types are covered by this document:

| NATIONAL STOCK NUMBER | ITEM IDENTIFICATION | TYPE |
|---|---|---|
| 6515-00-082-8264 | LUMBAR PUNCTURE KIT, Adult, Sterile, Disposable, 20's | I |
| 6515-00-082-8265 | LUMBAR PUNCTURE KIT, Infant, Sterile, Disposable, 10's | II |
| 6515-00-116-0654 | LUMBAR PUNCTURE KIT, Pediatric, Sterile, Disposable, 10's | III |
| 6515-01-056-6034 | LUMBAR PUNCTURE KIT, Sterile, Adult | IV |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

A-A-54404A

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5390
Location: c:\word\nish\lumbar



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

## INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

| I RECOMMEND A CHANGE: | 1 DOCUMENT NUMBER<br>A-A-54404A | 2. DOCUMENT DATE (YYMMDD)<br>29 May 1992 |
|---|---|---|

3 DOCUMENT TITLE
LUMBAR PUNCTURE KIT, STERILE, DISPOSABLE

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6. SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION | |
|---|---|---|
| c. ADDRESS (Include Zip Code) | d TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON (If applicable) | 7. DATE SUBMITTED (YYMMDD) |

8 PREPARING ACTIVITY

| a NAME<br>Defense Personnel Support Center<br>ATTN: DPSC-MSE | b TELEPHONE (Include Area Code)<br>(1) Commercial<br>(215)737-2870 | (2) AUTOVON<br>444-2870 |
|---|---|---|
| c ADDRESS (Include Zip Code)<br>2800 South 20th Street<br>Philadelphia, PA 19145 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756 2340 AUTOVON 289-2340 | |

DD Form 1426, OCT 89 Previous editions are obsolete. 198.29

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:33Z
Check the source to verify that this is the current version before use.